

COLDPROOF

COLDPROOF EXCURSION REPORT

Fridge ID Domestic/Bar Fridge (Non-Compliant)
Date Range Analysed Not specified

GEYER Classification	GEYER RED
Outcome Status	CRITICAL

Clinical Summary

Fridge Name	Domestic/Bar Fridge (Non-Compliant)
Model	Model not specified
Date Range	Not specified
Total Data Points	31146
Number of Excursions	42
Longest Excursion	499.6 min
Time Above 8°C	818.6 min
Time Below 2°C	0.0 min
Classification Triggered	REPORTABLE BREACH

GEYER Classification

REPORTABLE BREACH

Reason: Critical breach

Pharmacist Review

A risk-based review is recommended to determine whether the recorded temperature behaviour aligns with Strive-for-5 cold chain requirements.

Fridge Signature

Mean temperature	—
Min / Max	— / —
Drift	—
Spike frequency	—
Cooling return slope	—
Power-cycle events	—
Sensor noise index	—

Sensor Data Characteristics (Technical Note)

- Duplicate timestamps may appear due to logger time resolution.
- Clock jitter or micro-drift can occur naturally in embedded clocks.
- Minor noise around stable readings is normal for medical sensors.
- Flat-line periods may indicate internal averaging or buffered writes.
- Grouped samples occur in high-frequency logging modes.

Excursions (Audit Record)

Start Time	End Time	Duration (min)	Peak Temp	GEYER	Notes
09-09-2024 16:13	09-09-2024 21:20	307.7	13.3	Reportable	
09-09-2024 21:21	09-09-2024 21:21	0.1	8.1	Non-reportable	
09-09-2024 21:21	09-09-2024 21:21	0.0	8.1	Non-reportable	
09-09-2024 21:21	09-09-2024 21:21	0.0	8.1	Non-reportable	
09-09-2024 21:22	09-09-2024 21:22	0.0	8.1	Non-reportable	
09-09-2024 21:23	09-09-2024 21:23	0.0	8.1	Non-reportable	
09-09-2024 21:25	09-09-2024 21:25	0.0	8.1	Non-reportable	
09-09-2024 21:25	09-09-2024 21:25	0.0	8.1	Non-reportable	
09-09-2024 21:25	09-09-2024 21:26	0.8	8.1	Non-reportable	
09-09-2024 21:26	09-09-2024 21:39	13.2	8.2	Non-reportable	
09-09-2024 21:39	09-09-2024 21:39	0.0	8.1	Non-reportable	
09-09-2024 21:39	09-09-2024 21:39	0.0	8.1	Non-reportable	
09-09-2024 21:40	09-09-2024 21:40	0.0	8.1	Non-reportable	
09-09-2024 21:43	09-09-2024 21:43	0.0	8.1	Non-reportable	
09-09-2024 21:45	09-09-2024 21:45	0.0	8.1	Non-reportable	
10-09-2024 00:41	10-09-2024 01:06	24.8	8.2	Non-reportable	
10-09-2024 01:06	10-09-2024 01:06	0.2	8.1	Non-reportable	
10-09-2024 01:07	10-09-2024 01:07	0.0	8.1	Non-reportable	
10-09-2024 01:09	10-09-2024 01:09	0.0	8.1	Non-reportable	
10-09-2024 01:10	10-09-2024 01:10	0.0	8.1	Non-reportable	
10-09-2024 08:12	10-09-2024 08:12	0.2	8.1	Non-reportable	
10-09-2024 08:13	10-09-2024 16:32	499.6	10.9	Reportable	
10-09-2024 16:32	10-09-2024 16:32	0.1	8.1	Non-reportable	
10-09-2024 16:33	10-09-2024 16:33	0.1	8.1	Non-reportable	
10-09-2024 16:33	10-09-2024 16:33	0.0	8.1	Non-reportable	
10-09-2024 16:33	10-09-2024 16:34	1.0	8.1	Non-reportable	
10-09-2024 16:34	10-09-2024 16:35	1.2	8.1	Non-reportable	
10-09-2024 16:36	10-09-2024 16:36	0.8	8.1	Non-reportable	
10-09-2024 16:37	10-09-2024 16:37	0.0	8.1	Non-reportable	
10-09-2024 19:07	10-09-2024 19:07	0.0	8.1	Non-reportable	
10-09-2024 19:07	10-09-2024 19:09	1.8	8.1	Non-reportable	
10-09-2024 19:09	10-09-2024 19:09	0.0	8.1	Non-reportable	
11-09-2024 08:54	11-09-2024 09:30	35.8	8.7	Reportable	
11-09-2024 09:30	11-09-2024 09:31	0.2	8.1	Non-reportable	

Start Time	End Time	Duration (min)	Peak Temp	GEYER	Notes
11-09-2024 09:31	11-09-2024 09:31	0.0	8.1	Non-reportable	