

COLDPROOF

COLDPROOF EXCURSION REPORT

Fridge ID

Domestic/Bar Fridge (Non-Compliant)

Date Range Analysed

Not specified

GEYER Classification	GEYER RED
Outcome Status	CRITICAL

Clinical Summary

Fridge Name	Domestic/Bar Fridge (Non-Compliant)
Model	Model not specified
Date Range	Not specified
Total Data Points	156887
Number of Excursions	138
Longest Excursion	1214.8 min
Time Above 8°C	5139.0 min
Time Below 2°C	0.0 min
Classification Triggered	REPORTABLE BREACH

GEYER Classification

REPORTABLE BREACH

Reason: Critical breach

Pharmacist Review

A risk-based review is recommended to determine whether the recorded temperature behaviour aligns with Strive-for-5 cold chain requirements.

Fridge Signature

Mean temperature	—
Min / Max	— / —
Drift	—
Spike frequency	—
Cooling return slope	—
Power-cycle events	—
Sensor noise index	—

Sensor Data Characteristics (Technical Note)

- Duplicate timestamps may appear due to logger time resolution.
- Clock jitter or micro-drift can occur naturally in embedded clocks.
- Minor noise around stable readings is normal for medical sensors.
- Flat-line periods may indicate internal averaging or buffered writes.
- Grouped samples occur in high-frequency logging modes.

Excursions (Audit Record)

Start Time	End Time	Duration (min)	Peak Temp	GEYER	Notes
02-09-2024 09:31	02-09-2024 21:17	705.4	22.5	Reportable	
02-09-2024 21:17	02-09-2024 21:17	0.1	8.1	Non-reportable	
02-09-2024 21:18	02-09-2024 21:18	0.0	8.1	Non-reportable	
02-09-2024 21:19	02-09-2024 21:19	0.0	8.1	Non-reportable	
02-09-2024 21:21	02-09-2024 21:21	0.0	8.1	Non-reportable	
02-09-2024 21:21	02-09-2024 21:21	0.0	8.1	Non-reportable	
02-09-2024 21:21	02-09-2024 21:22	0.4	8.1	Non-reportable	
02-09-2024 21:22	02-09-2024 21:35	12.8	8.3	Non-reportable	
02-09-2024 21:35	02-09-2024 21:35	0.1	8.1	Non-reportable	
02-09-2024 21:35	02-09-2024 21:35	0.1	8.1	Non-reportable	
02-09-2024 21:35	02-09-2024 21:35	0.0	8.1	Non-reportable	
02-09-2024 21:36	02-09-2024 21:36	0.2	8.1	Non-reportable	
02-09-2024 21:36	02-09-2024 21:36	0.2	8.1	Non-reportable	
02-09-2024 21:36	02-09-2024 23:18	101.2	9.2	Reportable	
02-09-2024 23:18	02-09-2024 23:18	0.0	8.1	Non-reportable	
02-09-2024 23:24	02-09-2024 23:24	0.0	8.1	Non-reportable	
02-09-2024 23:26	02-09-2024 23:26	0.0	8.1	Non-reportable	
02-09-2024 23:50	02-09-2024 23:50	0.0	8.1	Non-reportable	
02-09-2024 23:50	03-09-2024 20:05	1214.8	12.9	Reportable	
03-09-2024 20:05	03-09-2024 20:06	0.2	8.1	Non-reportable	
03-09-2024 20:10	03-09-2024 20:10	0.0	8.1	Non-reportable	
03-09-2024 20:10	03-09-2024 20:10	0.0	8.1	Non-reportable	
03-09-2024 20:10	03-09-2024 20:10	0.0	8.1	Non-reportable	
03-09-2024 20:11	04-09-2024 00:10	239.2	9.5	Reportable	
04-09-2024 00:10	04-09-2024 00:10	0.0	8.1	Non-reportable	
04-09-2024 00:10	04-09-2024 00:10	0.0	8.1	Non-reportable	
04-09-2024 00:11	04-09-2024 00:11	0.0	8.1	Non-reportable	
04-09-2024 00:12	04-09-2024 00:12	0.3	8.1	Non-reportable	
04-09-2024 00:12	04-09-2024 07:42	449.5	10.9	Reportable	
04-09-2024 07:42	04-09-2024 07:42	0.1	8.1	Non-reportable	
04-09-2024 07:42	04-09-2024 07:42	0.0	8.1	Non-reportable	
04-09-2024 07:43	04-09-2024 07:43	0.0	8.1	Non-reportable	
04-09-2024 07:43	04-09-2024 07:44	0.8	8.1	Non-reportable	
04-09-2024 07:44	04-09-2024 18:45	660.4	14.6	Reportable	

Start Time	End Time	Duration (min)	Peak Temp	GEYER	Notes
04-09-2024 18:45	04-09-2024 18:52	6.9	8.4	Non-reportable	