

COLDPROOF

COLDPROOF EXCURSION REPORT

Fridge ID

Domestic/Bar Fridge (Non-Compliant)

Date Range Analysed

Not specified

GEYER Classification	GEYER RED
Outcome Status	CRITICAL

Clinical Summary

Fridge Name	Domestic/Bar Fridge (Non-Compliant)
Model	Model not specified
Date Range	Not specified
Total Data Points	766
Number of Excursions	1
Longest Excursion	148.0 min
Time Above 8°C	135.5 min
Time Below 2°C	0.0 min
Classification Triggered	REPORTABLE BREACH

GEYER Classification

REPORTABLE BREACH

Reason: Critical breach

Pharmacist Review

A risk-based review is recommended to determine whether the recorded temperature behaviour aligns with Strive-for-5 cold chain requirements.

Fridge Signature

Mean temperature	—
Min / Max	— / —
Drift	—
Spike frequency	—
Cooling return slope	—
Power-cycle events	—
Sensor noise index	—

Sensor Data Characteristics (Technical Note)

- Duplicate timestamps may appear due to logger time resolution.
- Clock jitter or micro-drift can occur naturally in embedded clocks.
- Minor noise around stable readings is normal for medical sensors.
- Flat-line periods may indicate internal averaging or buffered writes.
- Grouped samples occur in high-frequency logging modes.

Excursions (Audit Record)

Start Time	End Time	Duration (min)	Peak Temp	GEYER	Notes
02-09-2024 12:21	02-09-2024 14:49	148.0	18.2	Reportable	