

COLDPROOF

COLDPROOF EXCURSION REPORT

Fridge ID	Purpose-Built Vaccine Refrigerator (PBVR)
Date Range Analysed	Not specified
Timezone	Australia/Perth

GEYER Classification	GEYER GREEN
Outcome Status	PASS

Clinical Summary

Fridge Name	Purpose-Built Vaccine Refrigerator (PBVR)
Model	Model not specified
Date Range	Not specified
Total Data Points	725
Number of Excursions	0
Longest Excursion	0.0 min
Time Above 8°C	0.0 min
Time Below 2°C	0.0 min
Classification Triggered	COMPLIANT – NO REPORTABLE BREACH

GEYER Classification

COMPLIANT – NO REPORTABLE BREACH

Reason: No reportable breach

Pharmacist Review

A risk-based review is recommended to determine whether the recorded temperature behaviour aligns with Strive-for-5 cold chain requirements.

Fridge Signature

Mean temperature	—
Min / Max	— / —
Drift	—
Spike frequency	—
Cooling return slope	—
Power-cycle events	—
Sensor noise index	—

Sensor Data Characteristics (Technical Note)

- Duplicate timestamps may appear due to logger time resolution.
- Clock jitter or micro-drift can occur naturally in embedded clocks.
- Minor noise around stable readings is normal for medical sensors.
- Flat-line periods may indicate internal averaging or buffered writes.
- Grouped samples occur in high-frequency logging modes.

Excursions (Audit Record)

Start Time	End Time	Duration (min)	Peak Temp	GEYER	Notes
------------	----------	----------------	-----------	-------	-------