

COLDPROOF

Digital Cold Chain Integrity Report

Fridge ID	Uncertain
Date Range Analysed	Not specified
Timezone	Australia/Perth
Generated	12 December 2025 at 13:04 AWST

Clinical Summary

Fridge Name	Uncertain
Model	Model not specified
Date Range	Not specified
Total Data Points	750
Number of Excursions	7
Longest Excursion	0.0 min
Time Above 8°C	0.0 min
Time Below 2°C	0.0 min
Classification Triggered	COMPLIANT – NO REPORTABLE BREACH

AM/PM Logging Required
Record MIN/MAX twice daily as per Strive for 5.

GEYER Classification

COMPLIANT – NO REPORTABLE BREACH
Reason: No reportable breach

Pharmacist Review

A risk-based review is recommended to determine whether the recorded temperature behaviour aligns with Strive-for-5 cold chain requirements.

Fridge Signature

Mean temperature	—
Min / Max	— / —
Drift	—
Spike frequency	—
Cooling return slope	—
Power-cycle events	—
Sensor noise index	—

Sensor Data Characteristics (Technical Note)

- Duplicate timestamps may appear due to logger time resolution.
- Clock jitter or micro-drift can occur naturally in embedded clocks.
- Minor noise around stable readings is normal for medical sensors.
- Flat-line periods may indicate internal averaging or buffered writes.
- Grouped samples occur in high-frequency logging modes.

Excursions (Audit Record)

Start Time	End Time	Duration (min)	Peak Temp	GEYER	Notes
02-09-2024 09:32	02-09-2024 09:32	0.0	9.3	Non-reportable	
02-09-2024 11:17	02-09-2024 11:17	0.0	8.6	Non-reportable	
02-09-2024 13:24	02-09-2024 13:24	0.0	9.7	Non-reportable	
02-09-2024 13:39	02-09-2024 13:39	0.0	8.8	Non-reportable	
02-09-2024 14:21	02-09-2024 14:21	0.0	8.5	Non-reportable	
02-09-2024 15:24	02-09-2024 15:24	0.0	9.4	Non-reportable	
02-09-2024 18:13	02-09-2024 18:13	0.0	8.7	Non-reportable	