

FLEXHAND · DESIGN DECISION RECORD

FlexHand TRL 2 review decisions

Design Molecule Lab · designmolecule.com

DOCUMENT CONTROL

Document ID	FXH-DDR-001	Version	0.2
Status	Draft	Date	2026-09-25
Author	Amish Chadha	License	CERN-OHL-S-2.0
Project TRL	TRL 3 Proof of concept		
Repository	github.com/BoujeeEnjinia1701/flexhand		
Source	docs/decisions/0001-trl2-review-decisions.md @ uncommitted		

REVISION HISTORY

Version	Date	Author	Change
0.2	2026-09-25	Amish Chadha	Recommendations accepted by Amish (DDR-002)
0.1	2026-09-25	Amish Chadha	Record Amish's 2026-09-25 decisions on the TRL 2 review points; list open items and new items from FXH-CAL-001

0001: TRL 2 review decisions

- **Date:** 2026-09-25
- **Status:** accepted (items D1 to D8; items N1 to N5 accepted on 2026-09-25, see FXH-DDR-002); item O1 remains proposed

■ Context

The TRL 2 review note (docs/REVIEW.md , session of 2026-09-25, /populate) listed nine design choices marked "Proposed, awaiting Amish". On 2026-09-25 Amish reviewed the review points for every repo and wrote: "proceed with all of your recommendations across all batches. Make sure we don't proceed to TRL 4 on any of them." Every item that carried a recommendation is therefore decided as recommended. The item without a recommendation stays open. TRL 4 is on hold by the same instruction.

FlexHand does not use the portfolio's SwapCell pack, so the cross-cutting SwapCell interface decisions (wake method, charge-while-discharging mode, latch vibration rating, shared pack pricing) do not apply to this design.

■ Options considered

Table 1. Items with a recommendation in the TRL 2 review.

#	Item	Options	Recommendation
D1	Motor route (drives R2 and R7)	(a) 25 mm 227:1 gearmotors on the forearm; (b) micro (N20) gearmotors with a 20 N force target; (c) bench-top motor unit with longer sheaths	(a) for the first build; revisit the 450 g pack target with a therapist, since the forearm rests on a support during sessions
D2	Motor count	One motor for four fingers; two motors (one per finger pair); four motors	Two
D3	Passive only or plan for active assist (pitch-level)	Passive motion only; passive plus an active-assist mode triggered by the user's effort	Keep the pitch as passive motion for TRL 3; record active assist as a later option
D4	Cuff pressure fix (R11)	Wider padded cuffs; a distal anchor; a lower force limit	20 mm cuffs with a soft liner, checked for PIP clearance at TRL 3
D5	Design load	30 N per finger for mild to moderate flexor tone; higher or lower	Keep 30 N and check against published stiffness data
D6	Thumb	Passive with an abduction spacer; actuated	Passive for TRL 3
D7	Actuation	Tendons through Bowden sheaths; pneumatic glove	Tendons
D8	Pack voltage	2S (7.4 V nominal) with 12 V rated motors; 3S	Keep 2S

■ Decision

- **D1.** Decided by Amish, 2026-09-25: go with recommendation. Two Pololu 4869 class 25D 227:1 gearmotors on the forearm. The 450 g pack target in R7 stays in FXH-REQ-001 until a therapist has been consulted; FXH-CAL-001 shows about 675 g, so R7 is not met.
- **D2.** Decided by Amish, 2026-09-25: go with recommendation. Two motors: index and middle on motor 1, ring and little on motor 2.
- **D3.** Decided by Amish, 2026-09-25: go with recommendation. Passive motion only for TRL 3. An active-assist mode (user effort sensed by motor current or surface EMG) is recorded as a later option. The pitch and problem lines in `project.yaml` are unchanged.
- **D4.** Decided by Amish, 2026-09-25: go with recommendation. Anchor cuffs up to 20 mm wide with a 2 mm soft liner, checked for PIP clearance. The check in FXH-CAL-001 (section H) shows that a 20 mm cuff does not fit on any middle phalanx with 3 mm clearance to the joint creases. The model therefore uses the widest cuff that fits (12.2 to 18.6 mm), and R11 is still not met. See N2.
- **D5.** Decided by Amish, 2026-09-25: go with recommendation. The design load stays 30 N per finger. The check against published stiffness data (FXH-CAL-001, section B) shows it covers mild tone (MAS 1+) with a margin of about 1.7 but not MAS 3, which needs about 148 N. See N1.
- **D6.** Decided by Amish, 2026-09-25: go with recommendation. Thumb held passively in abduction by a spacer.
- **D7.** Decided by Amish, 2026-09-25: go with recommendation. UHMWPE tendons in PTFE-lined Bowden sheaths.
- **D8.** Decided by Amish, 2026-09-25: go with recommendation. 2S Li-ion pack (7.2 V under load) driving 12 V rated motors.

Budget and pitch: the TRL 2 review made no recommendation to change either, so `budget_usd` stays at \$500 and the pitch and problem lines are unchanged.

Items that remain open (no recommendation was made, so they stay "Proposed, awaiting Amish"):

- **O1.** First clinical co-design partner: a stroke unit occupational therapy team, a community rehabilitation service or a university rehabilitation lab. No recommendation; portfolio guidance is that community designs pick co-design partners per area later. Proposed, awaiting Amish.

New items raised by FXH-CAL-001 at TRL 3. Amish accepted all recommendations later on 2026-09-25 ("i accept all your recommendations, go with them across all repos"); the changes are recorded in FXH-DDR-002:

- **N1.** Severity band and R2 wording. Options: (a) reword R2 to "mild flexor tone (MAS 1 to 1+)" and keep 30 N; (b) raise the design load toward moderate tone (MAS 2 to 3 needs about 80 to 150 N), which needs larger motors and makes R7 worse. Recommendation: (a) for the first build. Decided by Amish, 2026-09-25: go with recommendation (FXH-DDR-002).
- **N2.** Cuff pressure (R11). Options: (a) add a fingertip thimble that shares the extensor load with the middle-phalanx cuff (about 23 to 40 kPa); (b) a lower force limit on the little finger; (c) review the 50 kPa limit with a therapist. Recommendation: (a). Decided by Amish, 2026-09-25: go with recommendation (FXH-DDR-002).
- **N3.** Per-finger force limit (R3). Current sensing sees the sum of a finger pair, so one finger could carry up to 80 N before a 40 N per finger limit trips. Options: (a) a balance pulley in the anchor block so both tendons of a pair carry equal tension; (b) set the pair limit to 40 N total; (c) rely on the 60 N breakaway. Recommendation: (a). Decided by Amish, 2026-09-25: go with recommendation (FXH-DDR-002).

- **N4.** Pack mass (R7), about 675 g against 450 g. Options: (a) ball-detent breakaways instead of magnets (about 40 g lighter); (b) a lighter, perforated cuff shell; (c) a relaxed target agreed with a therapist (see D1); (d) move the cells to a pouch on the forearm support. Recommendation: (a) and (b), then (c). Decided by Amish, 2026-09-25: go with recommendation (FXH-DDR-002); (c) waits on O1.
- **N5.** Stroke-time range (R5). The fastest extension at design load is about 3.3 s against the 3 s lower limit. Recommendation: relax the lower limit to 4 s at design load, which is also gentler; R4 still holds for strokes up to 5 s. Decided by Amish, 2026-09-25: go with recommendation (FXH-DDR-002).

■ Consequences

- FXH-PRB-001, FXH-PRC-001 and FXH-REQ-001 move to v0.3: the decided choices are no longer "proposed", and the requirement status comes from FXH-CAL-001.
- The TRL 3 layout places the gearmotors along the forearm with the spools at the front and idlers turning the tendons forward; this narrows the pack to 76 mm (FXH-DWG-001).
- R7 and R11 are not met; R2, R3 and R5 are at risk. Items N1 to N5 were applied to the design after Amish accepted them (FXH-DDR-002).
- TRL 4 work (test articles, bench tests, build procedures) is on hold by Amish's instruction.