

STEP CUE · DESIGN DECISIONS REGISTER

StepCue design decisions register

Design Molecule Lab · designmolecule.com

DOCUMENT CONTROL

Document ID	STC-DEC-001	Version	0.2
Status	Draft	Date	2026-10-02
Author	Amish Chadha	License	CERN-OHL-S-2.0
Project TRL	TRL 3 Proof of concept		
Repository	github.com/BoujeeEnjinia1701/stepcue		
Source	docs/06-design-decisions.md @ uncommitted		

REVISION HISTORY

Version	Date	Author	Change
0.2	2026-10-02	Amish Chadha	Amish approved the recommendations for open decisions 1 to 10 (STC-DDR-003 accepted); moved to decisions made
0.1	2026-10-01	Amish Chadha	Register opened with the build plan; open items from STC-DDR-001 to STC-DDR-003 and the review note; budget treated as a value-engineering target

StepCue design decisions register

Every design decision still to be made, and every decision made, in one place. Each decision is argued in full in its decision record under `docs/decisions/`; this register is the index Amish works from. The build plan (`docs/05-build-plan.md`) describes the design as it stands and does not list open decisions.

StepCue is a research and educational prototype, not a medical device.

■ Open decisions

None. All open decisions were decided on 2026-10-02.

■ To confirm when parts are bought

#	What to confirm	Why it matters	Source
1	The pressure sensor's tail reaches about 12 mm past the disc and its two tabs are 2.54 mm apart	The tab pads on the carrier are placed for this; a longer tail moves them	STC-DDR-003, P2
2	How far the controller module's USB-C receptacle stands past its board end (1.4 mm assumed)	Sets whether a plug seats fully in the bottom-wall notch	STC-DDR-003, P8
3	The 8-way flat flex cable has same-side contacts and its 9 mm width matches the ZIF connector	The tail must go into the connector with its contacts toward the board	STC-DDR-003, P5 and P7
4	The cell is 502535 class (about 35 x 25 x 5 mm) and protected	The pod's inside is sized round it, and R15 needs a protected cell	STC-CAL-001, R10 and R15
5	The heel counter of the test shoe is about 3.5 mm thick and curved at about 32 mm radius inside	The clip's 1 mm grip and its curve are set for this counter	STC-CAL-001, R12
6	The polyimide film takes solder at about 300 °C without lifting its adhesive	The sensors, tail strips and motor leads are soldered to pads on it	STC-DDR-003, P1

■ Value engineering

Value-engineering target: USD 200 (a hypothetical control target, not a limit). Estimated cost of the constructable design: USD 89.94 for one instrumented insole and one heel pod (USD 110.06 under the target). Main cost drivers and savings worth trying:

- The largest lines are the controller module (USD 15.99), the sensor laminate materials (USD 15.00), the five pressure sensors (USD 14.95), the cell (USD 7.00), and the interface board parts and hardware (USD 6.00 each).
- Making the design constructable added the tail connector (USD 2.50) and the spacer foam (USD 3.00) and repriced the hardware line (USD 5.00 to USD 6.00); the estimate rose from USD 83.44 to USD 89.94.
- Savings worth trying: buy laminate materials (polyimide film, copper tape) in quantities shared across several insoles, since most of the USD 15.00 is a minimum pack; a second insole for the other foot adds about USD 47 for the insole parts alone if it shares one pod design.

■ Decisions made

Date	Decision	Decided by	Record
2026-09-25	TRL 2 review items D1 to D8: one instrumented insole; nRF52840 module with built-in IMU; pressure plus IMU; electronics in a clip-on heel pod; haptic cue by default with audio through a phone or earbuds; rule-based detector first; commercial FSRs; requirement targets R1 to R15 adopted	Amish: "proceed with all of your recommendations across all batches. Make sure we don't proceed to TRL 4 on any of them."	STC-DDR-001
2026-09-25	O2: separate 0.3 s cue-start target for the phone or earbud audio route (0.1 s kept for haptic)	Amish: "i accept all your recommendations, go with them across all repos."	STC-DDR-002
2026-09-25	O3: 2.5 mm EVA insole base, 4.50 mm stack (the motor floor and laminate relief of that decision are since replaced by a through-hole, STC-DDR-003, accepted on 2026-10-02)	Amish, same instruction	STC-DDR-002
2026-09-26	This repo chosen for the first batch of product renders	Amish	Review note, 2026-09-26
2026-09-30	Make the design physically buildable while drawing the build plan	Amish: "If you are realising that the design cannot be built as per concept - fix the design assumptions to match and be physically feasible as you draw the illustrations."	STC-DDR-003 (changes accepted on 2026-10-02, below)
2026-10-01	Budget treated as a value-engineering target, not a limit	Amish: "the budgets are a hypothetical control target to ensure we are thinking along a value engineering lens."	This register
2026-10-02	Design for construction accepted: the changes P1 to P12 as modelled (option a)	Amish: "i approve your recommendations for all 555 open decisions."	STC-DDR-003, Table 1
2026-10-02	Charging: insole and pod come out of the shoe together (option a) for the supervised prototype sessions; an outside connector so the pod unplugs from the tail (option b) is to be decided before any take-home use	Amish: "i approve your recommendations for all 555 open decisions."	STC-DDR-003, A1
2026-10-02	Pod 37 mm wide with four M2 lid screws (option a); a snap lid can follow once the pod is printed	Amish: "i approve your recommendations for all 555 open decisions."	STC-DDR-003, A2
2026-10-02	Traces across the arch and ball of the foot: build as modelled (option a), check every trace for continuity before each wear session, and inspect after the first wear trials at TRL 4	Amish: "i approve your recommendations for all 555 open decisions."	STC-DDR-003, A3
2026-10-02	First co-design partner: a Parkinson's patient group, recruited through a local support group affiliated with the Parkinson's Foundation (the first candidate to approach), with a physiotherapy	Amish: "i approve your recommendations for all 555 open decisions."	STC-DDR-001 and STC-DDR-002, O1

Date	Decision	Decided by	Record
	practice brought in for the first supervised user session		
2026-10-02	Hero render: the 8 degree toe-up insole tilt for the hero render only; the flat fitted state stays the reference	Amish: "i approve your recommendations for all 555 open decisions."	Review note, 2026-09-26, item 1
2026-10-02	Smooth spline insole outline adopted at the next model revision; the traces keep their 1.5 mm edge margin	Amish: "i approve your recommendations for all 555 open decisions."	Review note, 2026-09-26, item 2
2026-10-02	Printed sensor rings, forefoot perforations, grip ribs, lid bezel recess and cadence mark are appearance only	Amish: "i approve your recommendations for all 555 open decisions."	Review note, 2026-09-26, item 3
2026-10-02	Rounded pod envelope with a 5 mm corner radius and a parting-line groove adopted at the next model revision	Amish: "i approve your recommendations for all 555 open decisions."	Review note, 2026-09-26, item 4
2026-10-02	Clay shoe shell and foot kept for the renders only	Amish: "i approve your recommendations for all 555 open decisions."	Review note, 2026-09-26, item 5