

TREMORTRACE · DESIGN DECISION RECORD

TremorTrace recommendations accepted

Design Molecule Lab · designmolecule.com

DOCUMENT CONTROL

Document ID	TRT-DDR-002	Version	0.1
Status	Draft	Date	2026-09-25
Author	Amish Chadha	License	CERN-OHL-S-2.0
Project TRL	TRL 3 Proof of concept		
Repository	github.com/BoujeeEnjinia1701/tremortrace		
Source	docs/decisions/0002-recommendations-accepted.md @ uncommitted		

REVISION HISTORY

Version	Date	Author	Change
0.1	2026-09-25	Amish Chadha	Record Amish's acceptance of all open recommendations (2026-09-25)

0002: Recommendations accepted

- **Date:** 2026-09-25
- **Status:** accepted (items D6 and D7); items O1 to O3 remain proposed

■ Context

After the TRL 3 session, two recommendations were still waiting for Amish: the charge current (TRT-DDR-001, item O4) and a lighter strap to restore mass margin under R6 (docs/REVIEW.md , TRL 3 session, recommended next step). On 2026-09-25 Amish wrote: "i accept all your recommendations, go with them across all repos." Every item that carried a recommendation is therefore decided as recommended. Items without a recommendation stay open. TRL 4 remains on hold.

■ Decision

Table 1. Items decided by this record.

#	Item	Options	Decision
D6	Charge current (was TRT-DDR-001 O4)	50 mA (0.33 C, about 3.6 h, about 65 mW charger heat) or 100 mA (0.67 C, about 1.8 h, about 130 mW)	Decided by Amish, 2026-09-25: go with recommendation. 50 mA; the firmware selects the 50 mA setting
D7	Strap	12 g silicone strap (TRL 3 baseline) or a woven textile strap of about 8 g	Decided by Amish, 2026-09-25: go with recommendation. Woven textile two-piece quick-release strap, 22 mm, about 8 g

Budget and pitch: no recommendation proposed a change to either, so `budget_usd` stays at \$150 and the pitch and problem lines are unchanged. The parts total rose from \$41.19 to \$43.19, well inside the budget.

■ What changed in the repo

- **D6.** TRT-CAL-001 v0.2, section 5: the 50 mA setting is recorded as decided and as a firmware rule. TRT-REQ-001 v0.4, R4 names the 50 mA recharge setting. TRT-PRC-001 v0.5 records the decision. Drawing TRT-DWG-002 Rev P2 notes "Charge at 50 mA". docs/04-calcs/sizing.py marks the setting as decided.
- **D7.** bom/bom.csv item 1 is now a woven textile strap, about 8 g, \$10.00 (was silicone, about 12 g, \$8.00); parts total \$43.19 (was \$41.19). cad/src/model.py strap thickness 1.4 mm (was 2.5 mm); STEP and STL re-exported. docs/04-calcs/sizing.py strap mass 8 g (was 12 g). TRT-CAL-001 v0.2: mass 25.4 g (was 29.4 g), margin 4.6 g (was 0.6 g), R6 met (was at risk). TRT-REQ-001 v0.4: R6 met. TRT-PRC-001 v0.5: component table, key numbers and design choices updated. TRT-DWG-002 Rev P2 (was P1): parts list names the textile strap. Concept media regenerated with the new key figures.

■ Items still open

These carried no recommendation, so they stay "Proposed, awaiting Amish":

- **O1.** First co-design and validation partner: a movement disorders clinic, a patient association, or the OpenRatio network. Proposed, awaiting Amish.

- **02.** Confirmation of "OpenRatio" as a possible partner (part of O1). Proposed, awaiting Amish.
- **03.** Approval of kit 1.3.1 (forearm-for-scale rule for small objects in the concept media). Proposed, awaiting Amish.

■ Consequences

- Requirement status: nine met, two not verifiable at TRL 3 (R7, R8), none at risk, none not met.
- No cross-repo action arises from these decisions.
- TRL 4 work (building a band, weighing it, measuring charge temperature, writing firmware) is on hold by Amish's instruction. The design stays at TRL 3.