

TREMORTRACE · DESIGN DECISION RECORD

TremorTrace recommendations accepted

Design Molecule Lab · designmolecule.com

DOCUMENT CONTROL

Document ID	TRT-DDR-002	Version	0.2
Status	Draft	Date	2026-10-02
Author	Amish Chadha	License	CERN-OHL-S-2.0
Project TRL	TRL 3 Proof of concept		
Repository	github.com/BoujeeEnjinia1701/tremortrace		
Source	docs/decisions/0002-recommendations-accepted.md @ uncommitted		

REVISION HISTORY

Version	Date	Author	Change
0.2	2026-10-02	Amish Chadha	01 to 03 recorded as decided by Amish on 2026-10-02
0.1	2026-09-25	Amish Chadha	Record Amish's acceptance of all open recommendations (2026-09-25)

0002: Recommendations accepted

- **Date:** 2026-09-25
- **Status:** accepted (items D6 and D7); items O1 to O3 decided by Amish on 2026-10-02 (TRT-DEC-001)

■ Context

After the TRL 3 session, two recommendations were still waiting for Amish: the charge current (TRT-DDR-001, item O4) and a lighter strap to restore mass margin under R6 (docs/REVIEW.md , TRL 3 session, recommended next step). On 2026-09-25 Amish wrote: "i accept all your recommendations, go with them across all repos." Every item that carried a recommendation is therefore decided as recommended. Items without a recommendation stay open. TRL 4 remains on hold.

■ Decision

Table 1. Items decided by this record.

#	Item	Options	Decision
D6	Charge current (was TRT-DDR-001 O4)	50 mA (0.33 C, about 3.6 h, about 65 mW charger heat) or 100 mA (0.67 C, about 1.8 h, about 130 mW)	Decided by Amish, 2026-09-25: go with recommendation. 50 mA; the firmware selects the 50 mA setting
D7	Strap	12 g silicone strap (TRL 3 baseline) or a woven textile strap of about 8 g	Decided by Amish, 2026-09-25: go with recommendation. Woven textile two-piece quick-release strap, 22 mm, about 8 g

Budget and pitch: no recommendation proposed a change to either, so `budget_usd` stays at \$150 and the pitch and problem lines are unchanged. The parts total rose from \$41.19 to \$43.19, well inside the budget.

■ What changed in the repo

- **D6.** TRT-CAL-001 v0.2, section 5: the 50 mA setting is recorded as decided and as a firmware rule. TRT-REQ-001 v0.4, R4 names the 50 mA recharge setting. TRT-PRC-001 v0.5 records the decision. Drawing TRT-DWG-002 Rev P2 notes "Charge at 50 mA". docs/04-calcs/sizing.py marks the setting as decided.
- **D7.** bom/bom.csv item 1 is now a woven textile strap, about 8 g, \$10.00 (was silicone, about 12 g, \$8.00); parts total \$43.19 (was \$41.19). cad/src/model.py strap thickness 1.4 mm (was 2.5 mm); STEP and STL re-exported. docs/04-calcs/sizing.py strap mass 8 g (was 12 g). TRT-CAL-001 v0.2: mass 25.4 g (was 29.4 g), margin 4.6 g (was 0.6 g), R6 met (was at risk). TRT-REQ-001 v0.4: R6 met. TRT-PRC-001 v0.5: component table, key numbers and design choices updated. TRT-DWG-002 Rev P2 (was P1): parts list names the textile strap. Concept media regenerated with the new key figures.

■ Items left open on 2026-09-25 (decided 2026-10-02)

These carried no recommendation on 2026-09-25; Amish decided all three on 2026-10-02 (TRT-DEC-001):

- **O1.** First co-design and validation partner: a movement disorders clinic, a patient association, or the OpenRatio network. **Decided 2026-10-02:** an academic movement disorders clinic with its own ethics board, introduced

through the OpenRatio network, with the International Essential Tremor Foundation as the route to participants, shared with StillBand and SteadySleeve, as the first candidate to approach.

- **02.** Confirmation of "OpenRatio" as a possible partner (part of O1). **Decided 2026-10-02:** confirmed, in soft, non-clinical wording.
- **03.** Approval of kit 1.3.1 (forearm-for-scale rule for small objects in the concept media). **Decided 2026-10-02:** approved.

■ Consequences

- Requirement status: nine met, two not verifiable at TRL 3 (R7, R8), none at risk, none not met.
- No cross-repo action arises from these decisions.
- TRL 4 work (building a band, weighing it, measuring charge temperature, writing firmware) is on hold by Amish's instruction. The design stays at TRL 3.