VOIDSCOPE · REGULATORY AND STANDARDS REGISTER

VoidScope regulatory and standards register

Design Molecule Lab · designmolecule.com

DOCUMENT CONTROL

Document ID	VDS-REG-001	Version	0.1
Status	Draft	Date	2026-10-24
Author	Amish Chadha	License	CERN-OHL-S-2.0
Project TRL	TRL 3 Proof of concept		
Repository	github.com/BoujeeEnjinia1701/voidscope		
Source	docs/07-regulatory.md @ 463eaf8		

REVISION HISTORY

Version	Date	Author	Change
0.1	2026-10-24	Amish Chadha	First issue (kit 1.9.1 upgrade): nine entries, cited with the date checked; classification and certification routes are to assess

VoidScope regulatory and standards register

Which standards and regulations might apply to VoidScope, where, and what is needed. VoidScope is published as an open engineering reference for a paper concept; nothing here is a claim of compliance, and anything marked "to assess" needs a qualified reviewer before any product claim. Not legal advice. Method: STANDARDS section 23.

> **Safety:** VoidScope is rescue equipment that may supply water and air to a trapped person and carries a lithium battery. This document is an open engineering reference, not certified rescue equipment, not a medical device and not legal advice.

Table 1. Regulatory and standards register.

#	Standard or regulation	Jurisdiction	What it covers	Applicability to this design	Owner (role)	Status
G1	IEC 60529, degrees of protection provided by enclosures (IP code)	International	Ingress of dust and water into enclosures	The camera head is bought as IP68 and the surface unit case has a lid; the IP rating of the tip housing, sheath and case is a design aim (PR-004)	Design lead	To assess (the bought camera's own rating is a supplier claim)
G2	UN Manual of Tests and Criteria, section 38.3 (lithium battery tests for transport)	International (UN, adopted into air, sea and road rules)	Tests a lithium cell or pack must pass before it is shipped	The two LiFePO4 packs are bought; the supplier should hold the 38.3 report	Design lead	Applies to the bought packs; supplier evidence to request
G3	Cell and pack safety standard for the LiFePO4 pack and its charger	International and market-specific	Safe design of the cells, protection circuit and charger	Packs have built-in protection and a matching charger (build plan stops S2, S3, S8)	Electrical lead	To assess (no standard number chosen)
G4	Regulation (EC) No 1935/2004, materials and articles intended to come into contact with food	European Union	Safety of materials touching food and drink	The bottle, tube, capillary and bite valve touch drinking water; food-grade materials are required by the problem statement constraints	Design lead	Applies as a guide for material choice; to assess for a product
G5	Regulation (EU) 2017/745, medical devices	European Union	Definition and classification of medical devices	The probe delivers water and air to a person; whether that makes it a device is not settled. It is published as a research and educational prototype, not a medical device	Regulatory consultant	To assess (classification with a qualified reviewer)
G6	INSARAG Guidelines	International (UN advisory group)	Methodology for classified urban search and rescue teams and their equipment	A search camera is a technical search tool in a classified team's kit; the guidelines do not certify products	Co-design partner	Does not certify; to assess for team use
G7	NFPA 1670, operations and training for technical search and rescue incidents	United States	Training and operational levels for technical rescue (including structural collapse)	Sets the context for the drill card and procedure, not the design of the probe	Co-design partner	To assess for the procedure and drills
G8		Market-specific	Interference and electrical safety of			To assess

#	Standard or regulation	Jurisdiction	What it covers	Applicability to this design	Owner (role)	Status
	EMC and electrical safety rules for the market of sale		the monitor, blower and modules	Low-voltage (12.8 V) equipment with a bought monitor and blower	Electrical lead	
G9	Export and import rules for batteries and cameras	Market-specific	Shipping of lithium packs and image sensors	Relevant only if kits are shipped; the design is published, not sold	Project lead	To assess

Status: to assess, applies, does not apply, met by design, certification needed later. No standard has been identified for the 15 kg carried-load target (PR-016) or the 55 C touch limit (PR-014 uses the burn-threshold figures in VDS-STM-001 Sources, not a product standard).

Sources

All checked 2026-10-24.

- IEC. IEC 60529:1989+AMD1:1999+AMD2:2013 Degrees of protection provided by enclosures (IP Code). https://webstore.iec.ch/en/publication/2447
- TUV SUD. Lithium battery testing under UN/DOT 38.3 (white paper on the UN Manual of Tests and Criteria section 38.3). https://www.tuvsud.com/en-us/resource-centre/white-papers/lithium-battery-testing
- European Parliament and Council. Regulation (EC) No 1935/2004 on materials and articles intended to come into contact with food. https://eur-lex.europa.eu/eli/reg/2004/1935/oj/eng
- European Parliament and Council. Regulation (EU) 2017/745 on medical devices. https://eur-lex.europa.eu/eli/reg/2017/745/oj (address by the standard ELI pattern; the page was not opened this session)
- INSARAG (International Search and Rescue Advisory Group). INSARAG Guidelines. https://insarag.org
- NFPA. NFPA 1670, Standard on Operations and Training for Technical Search and Rescue Incidents; overview at https://lifesaving.com/issues-safety-rescue/nfpa-1670-standard-for-technical-rescue/