WATERWATCH · REGULATORY AND STANDARDS REGISTER

WaterWatch regulatory and standards register

Design Molecule Lab · designmolecule.com

DOCUMENT CONTROL

Document ID	WWT-REG-001	Version	0.1
Status	Draft	Date	2026-10-25
Author	Amish Chadha	License	CERN-OHL-S-2.0
Project TRL	TRL 3 Proof of concept		
Repository	github.com/BoujeeEnjinia1701/waterwatch		
Source	docs/07-regulatory.md @ 96e344a		

REVISION HISTORY

Version	Date	Author	Change
0.1	2026-10-25	Amish Chadha	Kit 1.9.1 upgrade: first register for the TRL 3 design (option A), with notes for options B and C

WaterWatch regulatory and standards register

Which standards and regulations would apply to WaterWatch as a product, where, and what is needed. Not legal advice; anything marked "to assess" needs a qualified reviewer before any product claim. WaterWatch is a monitoring aid that reports measurements; it never certifies water as safe (PR-013), and no certification is pursued while the portfolio is capped at TRL 3. The first region is not yet chosen (East Africa, through the first candidate partner, WWT-DEC-001), so national rows use Kenya as the worked example and are "to assess" for the region chosen. Rows W1 to W3 reuse the water-security research shared across the portfolio (LMF-REG-001, LumaFlow), each checked on the date in its source. Method: STANDARDS section 23.

Table 1. Regulatory and standards register.

#	Standard or regulation	Jurisdiction	What it covers	Applicability to this design	Owner (role)	Status
R1	WHO *Guidelines for Drinking-water Quality*, 4th edition with the first and second addenda	International (guidance adopted into national standards)	Chlorine residual and turbidity guidance; the basis of the field values in WWT-PRB-001	Sets the meaning of the alert thresholds (0.2 mg/L free chlorine, 5 NTU; WWT-DDR-001 D6) and the rule that monitoring does not replace laboratory testing	Project lead	Applies (guidance)
R2	ISO 7027-1:2016, water quality, determination of turbidity, quantitative methods	International	Nephelometric turbidity at 860 nm with detection at 90 degrees	The turbidity optics follow its wavelength and geometry (WWT-STM-001 S10); a claim of conformance would need its calibration and test procedure	Design lead	To assess (method followed, conformance not claimed)
R3	ISO 7393-2, water quality, determination of free and total chlorine, colorimetric method with DPD	International	The DPD method used in the field reference (caretaker comparator) and in option B	Calibration reference for PR-001; the working principle of option B	Design lead	Applies (reference method)
R4	EN 1717, protection against pollution of potable water in installations and devices to prevent pollution by backflow (current edition EN 1717:2025+A1:2026)	European Union and countries that adopt it	Fluid categories and backflow devices, including air gaps	The model for PR-014: an isolation valve and check valve on the supply, and an open air gap at the cell outlet so drain water can never reach the cell or the supply; the local water code at the site governs	Design lead with the operator	To assess at the first site (design follows the air-gap principle)
W1	NSF/ANSI 61, drinking water system components, health effects	United States and Canada (referenced elsewhere)	Contaminants leached from materials in contact with drinking water	Wetted parts upstream of the check valve (saddle tee, isolation valve, strainer); downstream	Design lead with a certification body	To assess (choose certified grades at part selection)

#	Standard or regulation	Jurisdiction	What it covers	Applicability to this design	Owner (role)	Status
				parts drain to waste		
W2	EU Drinking Water Directive (EU) 2020/2184, Article 11 (materials in contact with water)	European Union	Hygiene of materials and products in contact with drinking water	Wetted materials if sold in the EU	Design lead	To assess
R5	National drinking water standard: for East African Community members the potable water specification EAS 12 (adopted nationally, for example as US EAS 12:2014 in Uganda; a 2022 revision was in draft)	East African Community	Limits for treated drinking water, including residual chlorine and turbidity	Context for site thresholds; WaterWatch measures, the operator complies	Scheme operator	To assess for the first region
R6	IEC 60529, degrees of protection provided by enclosures (IP code)	International	Dust and water ingress ratings	The enclosure is a bought IP66 box; glands IP68; PR-009 asks for IP65 or better	Design lead	Met by design (bought ratings), to verify at TRL 4 after drilling
R7	Kenya Information and Communications Act and the Importation, Type Approval and Distribution of Communications Equipment Regulations 2010 (Communications Authority of Kenya)	Kenya (worked example)	Type approval of radio and telecommunications equipment before import, sale or use	The cellular modem (BOM line 7) and the LoRaWAN variant; choose a modem module already type-approved in the country	Electronics lead	To assess; certification needed before any field unit
R8	EU Radio Equipment Directive 2014/53/EU; US FCC Part 15	European Union; United States	Radio and EMC conformity	Only if units are sold there	Electronics lead	Does not apply now; to assess at product stage
R9	UN Manual of Tests and Criteria, section 38.3, and the dangerous goods rules for lithium cells (UN 3480, or UN 3481 when packed with or in equipment)	International (transport by air, sea and road)	Tests a lithium cell or battery type must pass before it is shipped	The one 32700 LiFePO4 cell (19.2 Wh): buy cells with a UN 38.3 test summary from the maker	Procurement	Applies (buy certified cells)
R10	IEC 62133-2, safety of secondary lithium cells	International	Cell and pack safety	The cell and its protection	Electronics lead	

#	Standard or regulation	Jurisdiction	What it covers	Applicability to this design	Owner (role)	Status
	and batteries for portable applications			board; a recognised test is usual for a product		To assess (product stage)
W3	EU RoHS Directive 2011/65/EU and WEEE Directive 2012/19/EU; national e-waste rules	European Union; national	Hazardous substances and take-back of electrical equipment	Boards, modem, battery; end-of-life return through the operator	Electronics lead	To assess (product stage)
R11	Data protection law in the region (for Kenya, the Data Protection Act 2019)	National	Personal data processing	Readings are not personal data, but SMS recipients' phone numbers are; the operator owns the data (WWT-DEC-001)	Scheme operator	To assess

Status: to assess, applies, does not apply, met by design, certification needed later.

Option B (reagent colorimetric analyser) adds the safety data sheets for its DPD and buffer reagents, their transport and storage, and the discharge of reacted sample, which the operator would need to clear locally (to assess). Option C (probes in the riser) puts glass pH and ORP electrodes and their reference electrolyte inside the drinking water path, so W1 and W2 would apply to the probes themselves (to assess).

Sources

- World Health Organization, *Guidelines for Drinking-Water Quality: Fourth Edition Incorporating the First and Second Addenda* (Geneva: WHO, 2022), https://www.who.int/publications/i/item/9789240045064, accessed 2026-10-25.
- ISO 7027-1:2016 listing, International Organization for Standardization, https://www.iso.org/standard/62801.html, accessed 2026-10-25.
- EN 1717:2025+A1:2026 listing, Genorma, https://genorma.com/en/standards/en-1717-2025-a1-2026, and BSI Knowledge, https://knowledge.bsigroup.com/products/protection-against-pollution-of-water-intended-for-human-consumption-in-potable-water-installations-and-general-requirements-for-devices-to-prevent-pollution-by-backflow, accessed 2026-10-25.
- US EAS 12:2014, *Potable Water: Specification*, Uganda (Trade Barriers listing), https://tradebarriers.org/ntm/regulations/view/890, and the draft DEAS 12:2022, Tanzania Bureau of Standards, https://www.tbs.go.tz/uploads/publications/en-1664261414-DEAS%2012%20-%202022%20Potable%20water.pdf, accessed 2026-10-25.
- Communications Authority of Kenya, *Importation, Type Approval and Distribution of Communications Equipment Regulations, 2010*, https://www.ca.go.ke/sites/default/files/2023-06/Importation-Type-Approval-And-Distribution-Of-Communications-Equipment-Regulations-2010-1.pdf, and the type approval application form

(revised 2023), https://ca.go.ke/sites/default/files/2024-04/Type-Approval-Application-Form-Revised-2023.pdf, accessed 2026-10-25.

- UN 38.3 overview: Transport Canada, "Research Summary: Evaluation and Analysis of Substandard Lithium-Ion Batteries by UN 38.3 Testing," https://tc.canada.ca/en/dangerous-goods/publications/lithium-battery-transport-research/research-summary-evaluation-analysis-substandard-lithium-ion-batteries-383-testing, accessed 2026-10-25.
- Rows W1 to W3: as cited in LMF-REG-001 (LumaFlow, water-security focus area), checked 2026-10-25.
- Rows R3, R6, R8, R10 and R11 are listed from the standards' and laws' own titles; their applicability is "to assess" and needs a test laboratory or regulatory consultant.