

Supplementary Appendix S1:

Machine perfusate consisted of 500 mL of Gelofusine (B Braun, Melsungen, Germany), and 3 units of type O-packed red blood cells. Bolus priming additives included calcium gluconate 10 mL 10% solution (Fresenius Kabi, Richmond Hill, ON, Canada), cefuroxime 750mg (SteriMax Inc., Oakville, ON, Canada), and sodium heparin 10,000 IU (Pharmaceutical Partners Canada, Richmond Hill, ON, Canada). Sodium bicarbonate 30 mL 8.4% (Hospira, Montreal, QC, Canada) was added as needed to maintain pH between 7.35 – 7.45. Additional additives, delivered by continuous infusion for up to 24 hours, included sodium taurocholate (supplied by OrganOx, UK), insulin (Eli Lilly Canada Inc., Toronto, ON, Canada), heparin, prostacyclin (GlaxoSmithKline Inc., Mississauga, ON, Canada), and variable rate infusion Nutriflex (B Braun, Melsungen, Germany) carbohydrate/protein, non-lipid parenteral nutrition. NMP was performed at 37°C, with physiological pH, portal venous and hepatic arterial vascular pressures. After machine priming and liver cannulation, all pumps, perfusion parameters and arterial blood gas adjustments were made automatically by proprietary algorithms. An onboard O₂ generator, CO₂ clearance respirator and in-line blood gas analyzer (Terumo CD500, Terumo Corp., Hatagaya, Japan) was fully integrated into the *metra*[®] design, and did not require maintenance during perfusion.