IMPERIAL

MedTechONE

Knowledge Base

What are the key factors that affect medical device usability?

1 User Interface Design

2 Task Complexity and Workflow integration

3 Training and User Experience

4 Use Environment

5 User Feedback and Alerts

6 Accessibility

7 Error Recovery and Resilience

8 Cognitive Load

1. User Interface Design

The user interface (UI) is the point of interaction between the user and the device, and its design plays a critical role in usability. Poorly designed interfaces can lead to user confusion, errors, and delays in operation.

Key Elements:

- Clarity and Simplicity: The UI should be easy to navigate, with clear labels, intuitive controls, and well-organized information. Complex or cluttered interfaces increase cognitive load and make errors more likely.
- Feedback Mechanisms: Users need clear feedback on their actions, such as visual or auditory signals confirming that an action was successful or alerting them to errors.
- Error Prevention: The UI should be designed to prevent user errors through safeguards, such as requiring confirmation for critical actions (e.g., administering medication) or limiting available options to reduce the risk of incorrect inputs.

- Example: A ventilator with a complex display that requires multiple button presses to access key settings can confuse users, increasing the risk of misconfigurations. Simplifying the interface by providing preset modes and clear visual feedback can reduce this risk

2. Task Complexity and Workflow integration

Explanation: The complexity of tasks required to operate a device and how well the device integrates into the clinical workflow are important determinants of usability. Devices should streamline tasks and fit naturally into existing workflows to avoid interrupting user processes.

Key Elements:

- Task Complexity: Devices should simplify complex tasks to reduce the cognitive burden on users. For example, automating repetitive tasks or providing guided instructions can improve usability.

- Workflow Compatibility: Devices that align with users' typical workflows and don't require significant adjustments or retraining will be easier to adopt and use efficiently.

- Example: A portable ultrasound machine that integrates with existing patient records and can be operated with minimal steps reduces the time needed for setup and scanning, thereby improving usability

3. Training and User Experience

The amount of training required to use a device and the prior experience of the users significantly affect usability. Devices should be designed with the user's knowledge and skills in mind.

Key Elements:

- Training Requirements: Devices that require extensive training or frequent retraining may be prone to misuse, especially in emergency situations where users may not have time to review instructions.
- User Expertise: The usability of a device must consider the different experience levels of users. For example, devices intended for patients or caregivers at home should be far simpler than those used by trained healthcare professionals.

- Example: A home glucose meter should have an easy-to-understand user interface and minimal setup steps to accommodate users who may not have medical training

4. Use Environment

The environment in which a device is used—whether it's a hospital, clinic, or home—greatly impacts its usability. The design must account for the specific challenges posed by different settings.

Key Elements:

- Lighting Conditions: Devices used in low-light environments (e.g., emergency rooms) should have displays that are visible in different lighting conditions.

- Space Constraints: Devices used in small or crowded spaces should have a compact, ergonomic design that is easy to handle without requiring too much room.
- Environmental Factors: Noise, temperature, and other environmental factors can also influence usability. For example, alarms should be audible in noisy hospital settings, and devices must be resilient in harsh environments.

- Example: A defibrillator designed for public use should have large, clearly labeled buttons and audible prompts that guide untrained users through the process, ensuring that it can be operated correctly even in chaotic environments

5. User Feedback and Alerts

Feedback mechanisms and alerts are crucial for helping users navigate the device, avoid errors, and respond appropriately in critical situations. These should be clear, timely, and contextually appropriate.

Key Elements:

- Visual, Auditory, and Tactile Feedback: Devices should provide multiple forms of feedback to ensure users are aware of device status or errors. For example, beeps, lights, and vibrations can signal different statuses.
- Alarm Fatigue: In environments like hospitals where multiple devices may issue alarms, there is a risk of users becoming desensitized. Therefore, alarm design should differentiate between critical and non-critical issues.

- Example: An infusion pump that gives both visual and auditory alerts when a medication dose is complete or if there is an error in the infusion flow ensures that users can quickly address any issues

6. Accessibility

Accessibility refers to how well a device accommodates users with physical, sensory, or cognitive limitations. A medical device must be designed so that a broad range of users, including those with disabilities, can use it effectively.

Key Elements:

- Physical Accessibility: Devices should have controls that are easy to reach, buttons that are large enough for users with limited dexterity, and displays that are easy to read.
- Sensory Accessibility: Considerations should include auditory and visual feedback for users who may have hearing or vision impairments.
- Cognitive Load: Devices should simplify interactions to minimize the mental effort required, especially for users with cognitive impairments.

- Example: A hearing aid should provide clear auditory feedback with simple volume controls that can be easily adjusted by users with limited dexterity

7. Error Recovery and Resilience

No device is entirely immune to errors, but well-designed devices make it easy for users to recover from mistakes without causing harm. Devices should be resilient in the face of errors and guide users back to correct operation.

Key Elements:

- Undo Functionality: Devices should allow users to reverse actions without causing harm. This is especially important for high-risk tasks such as administering medication.
- Error Identification: Devices must clearly identify errors and suggest corrective actions, ensuring that users can recover from mistakes easily.

- Example: An automated medication dispenser might prompt users to confirm the dose and alert them if they make an error, allowing them to correct the mistake before proceeding

8. Cognitive Load

The mental effort required to use the device impacts its usability, especially in high-stress or fast-paced environments. Devices that are easy to learn and operate reduce cognitive load, allowing users to focus on critical tasks.

Key Elements:

- Simplicity: Devices with clear instructions, fewer steps, and intuitive interfaces reduce cognitive load.
- Error Prevention: Design features like predictive text, autocomplete, or guided workflows reduce the mental effort required from users.

- Example: A surgical navigation system that provides real-time, easy-to-understand guidance during complex procedures reduces the cognitive load on surgeons, enabling them to focus on critical decisions

Concluding Remarks:

The usability of a medical device is influenced by multiple factors, including its user interface design, task complexity, training requirements, use environment, feedback systems, accessibility, error recovery, and cognitive load. By addressing these factors early in the design and development process, manufacturers can ensure that their devices are not only compliant with regulatory standards like IEC 62366 but also safe, efficient, and easy to use in real-world settings.