

# AudiCalPro Calibration Certificate

**Kelompok 212**

Generated: 2026-05-12 21:51:31

| Frequency (Hz) | Measured dB | Reference dB | Correction dB | Status |
|----------------|-------------|--------------|---------------|--------|
| 125            | 69.50       | 70.00        | 0.50          | PASS   |
| 250            | 68.00       | 70.00        | 2.00          | PASS   |
| 500            | 72.50       | 70.00        | -2.50         | PASS   |

***Total Frequencies: 3***

***PASS: 3***

***FAIL: 0***

***Overall Result: PASS***

Operator Signature: \_\_\_\_\_