BioMedical Engineering OnLine

RESEARCH   **Open Access**

A photoacoustic image reconstruction method using total variation and nonconvex optimization

Chen Zhang, Yan Zhang and Yuanyuan Wang[*]

* Correspondence:
yywang@fudan.edu.cn
Department of Electronic
Engineering, Fudan University,
Shanghai 200433, China

Abstract

Background: In photoacoustic imaging (PAI), the reduction of scanning time is a major concern for PAI in practice. A popular strategy is to reconstruct the image from the sparse-view sampling data. However, the insufficient data leads to reconstruction quality deteriorating. Therefore, it is very important to enhance the quality of the sparse-view reconstructed images.

Method: In this paper, we proposed a joint total variation and L_p-norm (TV-L_p) based image reconstruction algorithm for PAI. In this algorithm, the reconstructed image is updated by calculating its total variation value and L_p-norm value. Along with the iteration, an operator-splitting framework is utilized to reduce the computational cost and the Barzilai-Borwein step size selection method is adopted to obtain the faster convergence.

Results and conclusion: Through the numerical simulation, the proposed algorithm is validated and compared with other widely used PAI reconstruction algorithms. It is revealed in the simulation result that the proposed algorithm may be more accurate than the other algorithms. Moreover, the computational cost, the convergence, the robustness to noises and the tunable parameters of the algorithm are all discussed respectively. We also implement the TV-L_p algorithm in the *in-vitro* experiments to verify its performance in practice. Through the numerical simulations and *in-vitro* experiments, it is demonstrated that the proposed algorithm enhances the quality of the reconstructed images with faster calculation speed and convergence.

Keywords: Photoacoustic imaging, Image reconstruction, Total variation, L_p-norm, Nonconvex optimization

Introduction

Photoacoustic imaging (PAI), also known as optoacoustic tomography (OAT) or thermoacoustic tomography (TAT), is a novel hybrid biomedical imaging modality which combines the strengths of both optical and ultrasound imaging [1-6]. Due to its non-ionizing nature, it has been considered as a promising imaging technique and developed rapidly during the past decade. PAI reveals physiologically specific optical absorption contrast of the biological tissues, which has great potential in clinic applications such as early tumor detection [7,8], vessel imaging [9,10] and brain imaging [11].

PAI is developed based on the photoacoustic effect [1,2], which is a process describing that the imaging tissues absorb the laser energy and convert it into acoustic waves. In this paper, we focus on the computed-tomographic PAI. In this kind of imaging mode, a laser pulse is used to illuminate the imaging tissues from the top. Some of the

laser energy is absorbed and converted into heat, leading to thermoelastic expansion and thus wideband ultrasonic wave emission. The generated photoacoustic signals are then detected by a scanning ultrasound transducer or a transducer array to form images. Based on these detected signals, the optical absorption deposition of the imaging tissues can be calculated by using an image reconstruction algorithm.

In PAI, the reconstruction algorithms have become the vital factor of imaging quality. The PAI reconstruction result benefits a lot from a stable, accurate and efficient algorithm. A variety of analytical image reconstruction algorithms have been developed. Reconstruction algorithms based on the inverse spherical radon transform have been proposed in both the time-domain [12,13] and the frequency-domain [14,15]. The filtered back-projection (FBP) algorithm proposed by Xu *et al.* is the most popular algorithm due to its accuracy and convenience [16,17]. The deconvolution reconstruction algorithm proposed by Zhang *et al.* has specific advantage under the circumstance of limited-angle sampling and heterogeneous acoustic medium [18,19]. Several investigations have been made to propose the algorithms in plane geometries for imaging with the linear array of transducer [20,21]. The analytical reconstruction methods have advantage in the computational cost and implementation convenience. However, the analytical algorithms fail to keep effective when the sampling points are sparse. The ignorance of measurements noises leads to the severely quality decline in the noisy situation. Those drawbacks above limit the applications of the analytical algorithms and impair their performance. Then the iterative image reconstruction methods are proposed to overcome these shortcomings and enhance the image quality of PAI.

The iterative image reconstruction methods usually build up a model to describe the relationship between the detected photoacoustic signals and the optical absorption deposition. So they are also called the model-based algorithms. Most of them calculate the optical absorption deposition iteratively to get the final reconstructed image. With proper optimization condition setup, the model-based methods can provide a more accurate and robust image reconstruction compared to the analytical ones [22,23]. Many methods that proved to be useful in other aspects have been adopted in PAI reconstruction as an optimization condition of the model-based methods. Some algorithms focus on the compensation of the acoustic inhomogeneous phenomenon [24,25]. Jose *et al.* proposes an iterative approach that takes the speed-of-sound of subject into account. They acquire the 2D speed-of-sound distributions and use this speed-of-sound map in their reconstruction algorithm [24]. Huang *et al.* develop and establish a full-wave iterative reconstruction approach in PAI to deal with the acoustic inhomogeneous and acoustic attenuation problem [25]. The compressed sensing has been involved in PAI reconstruction aiming to reduce the measurements and accelerate the data acquisition [26,27]. The model-based algorithm proposed by Rosenthal *et al.* recovers the image in the wavelet domain with a different strategy [28]. Meng *et al.* develop a compressed sensing framework by using partially known support [29,30]. The reported results show some improvement of image qualities. The total variation (TV) coefficient is always used to de-noise the image. Some algorithm are proposed by using the total variation minization to PAI image reconstruction [31-34]. Yao *et al.* propose the total variation minization (TVM) with the TV coefficient involved in the finite elment method to enhance the image quality and overcome the limit-angle problem [31,32]. An adaptive steepest-descent-projection onto convex sets (ASD-POCS) is proposed by

Wang *et al.* with the TV utilized in the iteration [33]. They investigate and employ the TV-based iterative image reconstruction algorithms in three-dimensional PAI. Zhang *et al.* utilizes the TV coefficient along with the gradient descent method in PAI reconstruction to propose the Total variation based gradient descent (TV-GD) algorithm [34]. The TV-GD method is reported to be stable and efficient under the sparse-view circumstance for PAI reconstruction. From the discussion noted above, it can be deduced that the iterative algorithms for PAI reconstruction have advantage in reconstruction qualities and robustness to noise. Now the reduction of scanning time is the main concern of PAI. A popular strategy is to reconstruct the image from sparse-view sampling data. Also there exists photoacoustic imaging system which can image the whole area with one laser exposure. These systems usually have large amount of transducers around the imaging area. With the help of sparse-view photoacoustic imaging reconstruction method, the transducer amount can be reduced. This reduction benefits the system from two main aspects. First, it helps to manage the system complexity to a lower level. The lower complexity system is more stable and easier to maintain. Second, this reduction also means the reduction of data scale. The data scale reduction can make the acquisition process more simple and flexible. Besides these two aspects, it is also worth to mention that it reduces the cost of the whole system. All those mentioned above is very important for further clinical applications. It is very important to develop a sparse-view imaging system. In this situation, the qualities of the iterative reconstructed images have room for improvement. Take the TV-GD method for example, it is reported to be an efficient and high-quality algorithm in sparse-view situation. But the paintinglike artifacts emerge and some detail information is lost in the extremely sparse-view reconstruction.

The compressed sensing theories have been adopted in PAI reconstruction, in which the L_1-norm of the signal is minimized to obtain the reconstructed image. Recently, Chartrand reported that by replacing the L_1-norm with the L_p-norm ($0 < p \leq 1$), accurate reconstruction is possible with substantially fewer measurements [35]. This nonconvex optimization setting has been successfully applied to Magnetic Resonance Imaging (MRI) image reconstruction [36,37]. The results show that the algorithm with L_p-norm can provide accurate reconstruction image with fewer measurements comparing to the L_1-norm based algorithms. To another dimension of this optimization problem, several algorithms have been proposed to get a better performance in image reconstruction through jointly minimizes the TV value and L_1-norm value [38,39] in MRI image reconstruction.

In this paper, we present a novel algorithm to the problem of reconstructing the image from sparse-view data in PAI. The algorithm is based on the jointly minimization of total variation and nonconvex L_p-norm (TV-L_p). The reconstructed image is updated by calculating its joint total variation value and L_p-norm value. The operator-splitting framework is used to reduce the computational cost, and the Barzilai-Borwein step size selection method is adopted to obtain the faster convergence. Through the numerical simulation, the image reconstruction in the case of insufficient sampling data was accomplished. The reconstruction result is compared with several other algorithms including the FBP [16], the L_1-norm [27] and the TV-GD method [34]. The computational cost and the convergence of the proposed algorithm is also discussed and compared with other algorithms. The numerical simulations also cover the robustness to the noise and the tunable

parameter investigation. Through the numerical simulations and *in-vitro* experiments, it is demonstrated that the proposed algorithm enhances the quality of the reconstructed images with the faster calculation speed and convergence. It's worthwhile to mention that like Ref. [27] and other iteration method, we also used a projection matrix to connect the acoustic pressure measurements with the reconstructed image. But there are some implementation differences between our method and that one. We both use an intermediate variable to simplify our equations. Ref. [27] used the velocity potential as the intermediate variable and we used a linear integration of the initial pressure along an arc whose center is the position of the ultrasound sensor and with a certain radius *ct*. The Ref. [27] used a sparsifying matrix and minimized the L_1-norm in sparsifying domain to get the reconstruction. We used the information from sparsifying domain and piecewise continuous behavior to reconstruct the image. Also, we adapted the *p*-norm minimization into the algorithm, so it can be a more accurate algorithm in sparse-view PAI.

The main contribution of this paper is to develop a novel algorithm for solving the problem of reconstructing the image from sparse-view data in PAI. Our contributions are threefold. First, we include the nonconvex optimization into the PAI reconstruction. This nonconvex optimization setting can provide more stable and accurate result under the sparse-view situation. Second, we combine the nonconvex optimization with TV minimization. The combined method is able to reconstruct more detailed image with sharp edge. Finally, we implement the Barzilai-Borwein method accelerates the reconstruction speed and improves the convergence considerably.

This paper is organized as follows. 'Theory and method' describes the theory of the proposed algorithm. The numerical simulation is introduced in 'Simulation'. The *in-vitro* experimental results are shown in '*In-vitro* experiments'. The conclusions of this work are drawn in 'Conclusion'.

Theory and method
Photoacoustic theory

In this paper, the two-dimensional PAI is concerned in the simulations and experiments. In 2D PAI, a laser pulse is used to illuminate the imaging tissues from the top. Due to the photoacoustic, the illumination creates an initial acoustical pressure field. The initial acoustical pressure field propagates as ultrasound waves, which can be detected by ultrasound transducers. Based on the physical principle of the photoacoustic effect, assuming that the illumination is spatially uniform, the relationship between the acoustical pressure measurements and the initial pressure rise distribution can be derived as :

$$\nabla^2 p\left(\vec{r},t\right) - \frac{1}{c^2}\frac{\partial^2 p\left(\vec{r},t\right)}{\partial t^2} = -\frac{\mu}{C_p}u\left(\vec{r}\right)\cdot\frac{\partial I(t)}{\partial t} \tag{1}$$

where $p\left(\vec{r},t\right)$ is the acoustic pressure measurements at the position r and the time *t*, *c* is the sound speed, C_p is the specific heat, μ is the isobaric expansion coefficient, *I(t)* is the temporal profile of the laser pulse and $u\left(\vec{r}\right)$ is the initial pressure rise distribution. In our study and many photoacoustic tomography studies, we employ a laser pulse with a very short duration. Its duration is nano seconds. So here we made an approximation to treat the *I(t)* as a Dirac-delta function.

In order to recover initial data for the wave equation, several inversion formulas have been established to solve this as a filtered back-projection problem [12,40]. By using the Green's function [12] to solve equation (1), the acoustic pressure measurements can be deduced as:

$$p\left(\vec{r}_0, t\right) = \frac{\mu}{4\pi C_p} \frac{\partial}{\partial t} \oiint_{\left|\vec{r}-\vec{r}_0\right|=ct} \frac{u\left(\vec{r}\right)}{t} \mathrm{d}^2 \vec{r} \qquad (2)$$

where $\vec{r}_0$ is the position of the ultrasound transducer.

In PAI experiments, an ultrasound transducer is used to receive the acoustic pressure measurements at different positions, and the image reconstruction is regarded as an inverse problem to obtain the initial pressure rise distribution. In the iteration of the image reconstruction, a projection matrix A is typically established to connect the acoustic pressure measurements with the reconstructed image. The measurements can be calculated based on the reconstructed image, and then the reconstructed image can be repeatedly corrected by minimizing the difference between the calculated measurements and the real ones. In this way, the optimization method can be used for collaboration and then the iteration reconstruction algorithm can be developed.

Compressed sensing for PAI

If the sampling data is insufficient, the projection matrix A is ill-conditioned. Thus, the matrix A does not have an exact inversion. As a result, it leads to streaking artifacts in the reconstructed image. This problem can be treated by incorporating the compressed sensing theory into PAI.

We define a new variable f as:

$$f\left(\vec{r}_0, t\right) = \frac{4\pi C_p}{\mu} \int_0^t p\left(\vec{r}_0, t\right) \mathrm{d}t \cdot t \qquad (3)$$

Then the equation (2) can be converted as follows:

$$f\left(\vec{r}_0, t\right) = \oiint_{|\mathbf{r}'-\mathbf{r}_0|=ct} u\left(\vec{r}\right) d^2 \vec{r} \qquad (4)$$

In practical imaging, the reconstructed image and the measurements are processed discretely, and the image is reshaped into vectors for convenience. If the size of the reconstructed image $u\left(\vec{r}\right)$ is X pixels $\times$ Y pixels, then the total pixel number of the reconstructed image $u\left(\vec{r}\right)$ is N ($N = XY$). After vectorization, the reconstructed image $u\left(\vec{r}\right)$ becomes a vector u with the length of N. If the total number of the detection points is Q, the length of measurement in each detection point is M, the equation (4) can be expressed as:

$$f_i = A_i^{\mathrm{T}} \cdot \mathrm{u} \quad i = 1, 2, \cdots, Q \qquad (5)$$

where f_i is the integration of the $u\left(\vec{r}\right)$ along the arc that is centered in ith detection point and with a radius of ct, A_i is the projection matrix of the ith detection point,

T is the transpose operation of a matrix. The calculation of the projection matrix is as follows:

(a) Calculate an matrix $A_i(j)$ as:

$$A_i(j) = \max\left\{1 - \left|\frac{d \cdot \mathrm{d}x}{c \cdot \mathrm{d}t} - j\right|, 0\right\} \quad (1 \leq j \leq M) \tag{6}$$

where $d = \sqrt{(m - m_i)^2 + (n - n_i)^2}$, (m, n) is the position of the jth point in the reconstructed image, (m_i, n_i) is the position of the ith detection point, $\mathrm{d}x$ is the actual length between the two pixels in the reconstructed image, $\mathrm{d}t$ is the discretized time step and M is the total sampling points at one detection point.

(b) Vectorize the matrix $A_i(j)$ as the jth column vector in projection matrix A_i.
(c) Repeat the calculation M times to get the projection matrix A_i.
(d) Repeat step(a) to step(c) Q times to get the projection matrix in the different sampling positions $(A_1, A_2, \ldots, A_Q)$. Then write the projection matrixes in the forms as follows:

$$A = \begin{pmatrix} A_1^{\mathrm{T}} \\ A_2^{\mathrm{T}} \\ \vdots \\ A_Q^{\mathrm{T}} \end{pmatrix} \tag{7}$$

The equation (6) can be expressed as:

$$f = A \cdot \mathrm{u} \tag{8}$$

where the sizes of f, A and u are MQ pixels $\times$ 1 pixel, MQ pixels $\times N$ pixels and N pixels $\times$ 1 pixel respectively.

To reconstruct the photoacoustic image from incomplete measurements by using the compressed sensing theory, we can solve an optimization problem as follows:

$$\min_u \left\{ \left| \Psi^{\mathrm{T}} \mathrm{u} \right|_1 + \frac{1}{2} |A\mathrm{u} - f|_2^2 \right\} \tag{9}$$

where Ψ is a sparse transform matrix, $|\ |_1$ and $|\ |_2$ are the L_1-norm and L_2-norm respectively. By projecting the image onto an appropriate basis set, we can get a sparse representation of the original image. In this domain, most coefficients of the image are small, and a few large coefficients capture most information of the signal. In this way, we can recover a much more accurate image from those undersampled measurements.

In practical applications of PAI, the reconstructed images often show piecewise continuous behavior. The images like this always have small total variation (TV) values, which is defined as follows:

$$TV(\mathrm{u}) = \int |D\mathrm{u}| = \sum_{i=1}^{N} |D_i \mathrm{u}|_2 \quad i = 1, 2 \ldots N \tag{10}$$

where D_i is a matrix with the size of 2 pixels $\times N$ pixels that has two nonzero entries in each row to calculate the finite difference of u at the ith pixel. D is a matrix with the

size of $2N$ pixels $\times N$ pixels, and $D = (D^X;D^Y)$,D^X and D^Y are the horizontal and vertical global finite difference matrixes respectively.

It is reported that the TV based reconstruction algorithm can recover the image accurately from sparse sampling data [34]. Using TV values to reconstruct the image can be expressed mathematically as:

$$\min_u \left\{ TV(u) + \frac{1}{2}|Au - f|_2^2 \right\} \tag{11}$$

However, the TV minimization still has some limitations that impair its performance. The optimization of the TV value encourages the recovery of images with sparse gradients, thus resulting in the paintinglike staircase artifacts in the reconstructed images.

Recently, some research find out that the nonconvex optimization can reconstruct an accurate image with fewer measurements by replacing the L_1-norm with the L_p-norm ($0 < p \le 1$). Aiming to enhance the reconstruction quality and overcome the problem of TV based algorithm, we joint the L_p-norm with TV values to establish a new optimization which can be defined by:

$$\min_u \left\{ \alpha \cdot TV(u) + \beta \left| \Psi^T u \right|_p^p + \frac{1}{2}|Au - f|_2^2 \right\} \quad 0 < p \le 1 \tag{12}$$

where α and β are parameters corresponding to the weights of the TV value and L_p-norm value , $| \;|_p$ is the L_p –norm in this optimization problem respectively.

Therefore, we can obtain the reconstructed image by solving this new optimization problem in equation (12).

PAI reconstruction algorithm

In this part, we solve the optimization problem in equation (12) to establish a novel photoacoustic image reconstruction algorithm by using the total variation and nonconvex optimization.

We define the finite difference approximations to partial derivatives of u at the ith pixel along the coordinate as variable $\omega_i = D_i u$, the ith pixel's sparse coefficient as variable $z_i = \Psi_i^T u$, where Ψ_i is the sparse transform matrix of the ith pixel. The equation (12) can be deduced as:

$$\min_{u,z,\omega} \left\{ \alpha \sum_i |\omega_i|_2 + \beta \sum_i |z_i|_p^p + \frac{\rho}{2}|A \cdot u - f|_2^2 \right\} \quad i = 1, 2...N \tag{13}$$

where ρ is the parameter corresponding to the weight of the constraint condition in this optimization problem.

We form the augmented Lagrangian defined by

$$L\left(\omega, z, u; b^k, c^k\right) = \alpha \sum_i \left(|\omega_i|_2 + \frac{\rho}{2}|\omega_i - D_i u^k - b_i^k|_2^2 \right) + \beta \sum_i \left(|z_i|_p^p + \frac{\rho}{2}|z_i - \Psi_i^T u^k - c_i^k|_1^2 \right) + \frac{1}{2}|A \cdot u - f|_2^2 \tag{14}$$

where b_i^k is the TV step parameter in kth iteration, c_i^k is the L_p-norm step parameter in kth iteration, u^k is the vectorized image reconstruction in kth iteration.

This problem can be solved by

$$
\begin{cases}
\left(\omega^{k+1}, z^{k+1}, \mathbf{u}^{k+1}\right) = \min_{\omega,z,\mathbf{u}} L\left(\omega, z, \mathbf{u}; b^k, c^k\right), \\
b_i^{k+1} = b_i^k - \left(\omega_i^{k+1} - D_i \mathbf{u}^{k+1}\right), \\
c_i^{k+1} = c_i^k - \left(z_i^{k+1} - \Psi_i^{\mathrm{T}} \mathbf{u}^k\right),
\end{cases}
\tag{15}
$$

where ω^{k+1} is the finite difference approximations to partial derivatives of u in $(k + 1)$th iteration, z^{k+1} is the sparse coefficient in $(k + 1)$th iteration, $\mathbf{u}^{k+1}$ is the vectorized image reconstruction in $(k + 1)$th iteration, z_i^{k+1} is the sparse coefficient of the ith pixel in $(k + 1)$th iteration, ω_i^{k+1} is the finite difference approximations to partial derivatives of u at the ith pixel along the coordinate in $(k + 1)$th iteration; b_i^{k+1} is the TV step parameter in $(k + 1)$th iteration, c_i^{k+1} is the L_p-norm step parameter in $(k + 1)$th iteration.

By using the standard augmented Lagrangian method, the optimization problem in (15) can be deduced as

$$
\begin{aligned}
\left(\omega^{k+1}, z^{k+1}, \mathbf{u}^{k+1}\right) = \min_{\omega,z,\mathbf{u}} \Bigg\{ & \alpha \sum_i \left(|\omega_i|_2 + \frac{\rho}{2} \left| \omega_i - D_i \mathbf{u}^k - b_i^k \right|_2^2 \right) + \beta \sum_i \left(|z_i|_p^p + \frac{\rho}{2} \left| z_i - \Psi_i^{\mathrm{T}} \mathbf{u}^k - c_i^k \right|_1^2 \right) \\
& + \frac{\delta_k}{2} \left(\left| \omega^{k+1} - \omega^k \right|_2^2 + \left| z^{k+1} - z^k \right|_2^2 + \left| \mathbf{u} - \mathbf{u}^k + \delta_k^{-1} A^{\mathrm{T}} \left(A\mathbf{u}^k - f \right) \right|_2^2 \right) \Bigg\}
\end{aligned}
\tag{16}
$$

where ω^k is the finite difference approximations to partial derivatives of u in kth iteration, z^k is the sparse coefficient in kth iteration, $\mathbf{u}^k$ is the vectorized image reconstruction in kth iteration; δ_k is the Barzilai-Borwein step parameter in kth iteration.

After using the Barzilai-Borwein method to determine the step size δ, the optimization problem in equation (13) can be transformed into three sub-problem as follows:

$$
\begin{cases}
\omega_i^{k+1} = \min_{\omega_i} \left\{ |\omega_i|_2 + \frac{\rho}{2} \left| \omega_i - D_i \mathbf{u}^k - b_i^k \right|_2^2 + \frac{\delta_k}{2\alpha} \left| \omega_i - \omega_i^k \right|_2^2 \right\}, \\
z_i^{k+1} = \min_{z_i} \left\{ |z_i|_p^p + \frac{\rho}{2} \left| z_i - \Psi_i^{\mathrm{T}} \mathbf{u}^k - c_i^k \right|_1^2 + \frac{\delta_k}{2\beta} \left| z_i - z_i^k \right|_2^2 \right\}, \\
\mathbf{u}^{k+1} = \min_{\mathbf{u}} \left\{ \alpha\rho \left| D\mathbf{u} - \omega^{k+1} \right|_2^2 + \beta\rho \left| \Psi^{\mathrm{T}} \mathbf{u} - z^{k+1} \right|_2^2 + \delta_k \left| \mathbf{u} - \left(\mathbf{u}^k - \delta_k^{-1} A^{\mathrm{T}} \left(A\mathbf{u}^k - f \right) \right) \right|_2^2 \right\}, \\
b_i^{k+1} = b_i^k - \left(\omega_i^{k+1} - D_i \mathbf{u}^{k+1} \right), \\
c_i^{k+1} = c_i^k - \left(z_i^{k+1} - \Psi_i^{\mathrm{T}} \mathbf{u}^k \right), \\
\delta_{k+1} = \left| A\left(\mathbf{u}^{k+1} - \mathbf{u}^k \right) \right|_2^2 \Big/ \left(\left| \omega^{k+1} - \omega^k \right|_2^2 + \left| z^{k+1} - z^k \right|_2^2 + \left| \mathbf{u}^{k+1} - \mathbf{u}^k \right|_2^2 \right)
\end{cases}
\tag{17}
$$

where ω_i^k is the finite difference approximations to partial derivatives of u at the ith pixel along the coordinate in kth iteration respectively, z_i^k is the sparse coefficient of the ith pixel in kth iteration respectively; δ_{k+1} is the Barzilai-Borwein step parameter in $(k + 1)$th iteration.

We use the soft shrinkage operator to obtain the solution to ω-subproblem in equation (17), the operation is as follows:

$$
\begin{cases}
\omega_i{}^{k+1} = \max\left\{ \left\| \dfrac{a_1 t_1 + a_2 t_2}{a_1 + a_2} \right\| - \dfrac{1}{a_1 + a_2},\, 0 \right\} \dfrac{1/(a_1 + a_2)}{\|1/(a_1 + a_2)\|} \\
a_1 = D_i \mathrm{u}^k + b_i{}^k \\
a_2 = \omega_i{}^k \\
t_1 = \rho \\
t_2 = \delta_k/\alpha
\end{cases}
\qquad (i = 1,2...N) \qquad (18)
$$

where a_1, a_2, t_1 and t_2 are the variables used for a succinct expression.

As for the z-subproblem in equation (17), we use the soft p-shrinkage operator to solve it. The operator is defined by:

$$
\begin{cases}
z_i{}^{k+1} = \max\left\{ \left\| \dfrac{a_3 t_3 + a_4 t_4}{a_3 + a_4} \right\| - \dfrac{1}{a_3 + a_4} \left\| \dfrac{a_3 t_3 + a_4 t_4}{a_3 + a_4} \right\|^{1-p},\, 0 \right\} \dfrac{1/(a_3 + a_4)}{\|1/(a_3 + a_4)\|} \\
a_3 = \Psi_i{}^{\mathrm{T}} \mathrm{u}^k + c_i{}^k \\
a_4 = z_i{}^k \\
t_3 = \rho \\
t_4 = \delta_k/\beta
\end{cases}
\qquad (i = 1,2...N)
$$

$$(19)$$

where a_3, a_4, t_3 and t_4 are the variables used for a succinct expression.

The u-subproblem in equation (17) is a typical least squares problem. The solution can be easily obtained by:

$$
\mathrm{u}^{k+1} = F^{\mathrm{T}} \left\{ \frac{F\left[\alpha\rho D^{\mathrm{T}}\omega^{k+1} + \beta\rho\,\Psi^{\mathrm{T}} z^{k+1} + \delta_k \mathrm{u}^k - A^{-1}\left(A\mathrm{u}^k - \mathrm{f}\right) \right]}{\alpha\rho F^{\mathrm{T}} D^{\mathrm{T}} D F + \beta\rho\,\mathrm{I} + \delta_k \mathrm{I}} \right\}
\qquad (20)
$$

where F is the Fourier transform matrix.

As a result, the **TV- L_p** algorithm is summarized as follows:

(1) Initialization: input f, α, β, ε ,p and ρ. Set the reconstructed image $\mathrm{u}^0 = 0$, $b = c = 0$, $\delta_0 = 1$, $k = 0$.

(2) Apply equation (18) and (19) to update the value of ω and z.

(3) Apply equation (20) to update the value of u.

(4) Apply equation (17) to update the value of b, c and δ.

(5) If the exiting condition is met, end the iterations and output the result. Otherwise repeat the step from (2) to (4). The exiting condition is as follows:

$$
\frac{\left\| \mathrm{u}^k - \mathrm{u}^{k-1} \right\|}{\left\| \mathrm{u}^k \right\|} < \varepsilon
\qquad (21)
$$

Simulation

To verify the effectiveness of the proposed TV- L_p algorithm on PAI reconstructions, the simulations are designed. All the simulations are performed in Matlab v7.14 on a PC with a 3.07 GHz Intel Xeon processor (only 1 core is used in computation) and 32 GB memory. The sparsisfying operator Ψ is set to Haar wavelet transform using Rice wavelet toolbox. The sound speed is set to be consistent in the simulation as 1500 m/s.

Sparse-view reconstruction

In the simulation, we choose the Shepp-Logan phantom to be the initial pressure rise distribution. The forward simulation and inverse reconstruction are all performed in 2D. The phantom is shown in Figure 1. The measurements from the phantom are generated by using equation (2). The size of the phantom is 89.6 mm × 89.6 mm, the radius of the scanning circle is 42 mm and the size of the reconstructed image is 128 pixels × 128 pixels. During the simulation, the scanning circle covers 360°around the imaging phantom. Four different measurements are collected. The scanning step of tomographic angels is set to 2.25°, 4°, 12° and 20° respectively. So the sampling points are 160 views, 90 views, 30 views and 18 views correspondingly.

The parameters α, β, ε and ρ are set to be 1×10^{-2}, 1×10^{-2}, 1×10^{-5} and 1 respectively. The influence of these parameters will be discussed later. And the parameter p are set to be two different values as 0.5 and 0.8.

We choose the FBP [16], the L_1-norm [27] and the TV-GD [34] algorithms to be the comparison besides our proposed TV- L_p algorithm. The simulation results by using these different algorithms are shown in Figure 2. It's worthwhile to note that the weight used in the TV-GD algorithm is an adaptive parameter, as same as it is reported in [27]. The negative values in FBP reconstructed image are set to be zero.

It is shown in the first column of Figure 2 that, all three iterative algorithms have comparable reconstruction results when the sampling data is sufficient. Moreover, it is shown in Figure 2(a) that the contrast of FBP reconstructed image is not as high as the other three. But its resolution is comparable with the others visually. When the number of sampling points reduces, the qualities of the reconstructed images are strongly affected in the FBP reconstruction. When the sampling point gets sparse in the FBP reconstruction, the arc-like artifacts appears due to the back-projection arcs

Figure 1 The shepp-logan phantom.

Figure 2 The shepp-logan phantom reconstruction results by different algorithms. The first to fifth
rows refer to the FBP **(a-d)**, L1-norm **(e-h)**, TV-GD **(i-l)**, TV-Lp (p = 0.5) **(m-p)** and TV-Lp (p = 0.8) **(q-t)**
reconstructed images respectively and the first to fourth columns refer to the results from 160-view,
90-view, 30-view and 18-view respectively.

cannot be canceled out with each other. The iterative algorithms can provide better
qualities of the reconstructed images than the FBP method in sparse-view reconstruc-
tions. Among them, the L_1-norm method struggles to depress the noise. Meanwhile,
the TV-GD algorithm and the TV-L_p algorithm provides high-resolution images and
have no visually distinguishable decline in qualities of the reconstructed images as the
number of sampling points decreases.

As for the extreme sparse sampling points situation (18-view and 30-view), the image
reconstructed by the FBP algorithm, shown in Figure 2(d), has extremely severe artifacts.
The L_1-norm reconstruction and the TV-GD algorithm have a decline in image qualities.
The noise in the reconstructed images by the L_1-norm reconstruction, as shown in
Figure 2(j), cannot be depressed effectively. As for the TV-GD algorithm, the reconstruc-
tion produces piecewise artifacts which also make the qualities of the reconstructed

images decrease. In the TV-L_p image, the noise is depressed more effectively than the above three algorithm. The quality of the reconstructed image is not affected substantially by the insufficient sampling data.

We calculate the peak signal-to-noise ratios (PSNR) of the reconstructed images with the original phantom as a gold standard to provide a numeric quantification of the results. The bigger the value of PSNR is, the better quality of the image is. The PSNR is defined as:

$$PSNR = 10\log_{10}\left(\frac{XY \cdot MAX}{\sum\limits_{(i,j)=1}^{X,Y}\left(u(i,j) - t(i,j)\right)^2}\right) \tag{22}$$

where $t(i,j)$ means the gray-value of the original image, MAX the maximum possible pixel value of the image which in our simulation is 1.

We calculate the value of the PSNR of all images in Figure 2. The quantitative results are shown in Table 1.

From Table 1, it is shown that the PSNR of the FBP algorithm is always in a very low level due to its unsuitability for sparse-view sampling condition. As for those three compressed sensing based algorithms, the PSNR value of images reconstructed by the TV-L_p algorithm are the highest. The L_p-norm optimization constraint can provide the better performance in the extremely sparse sampling. With this improvement, the TV-L_p algorithm is more accurate than other algorithms in the sparse-view sampling condition shown in the quantitative results. Between the two different value of p, the parameter p that is set to be 0.5 has a slightly advantage against the other one. Also it is revealed from Table.1 that the 90-view shows higher PSNR than 160-view. When the sampling points are sufficient, it is possible that the fewer-view projection can produce better reconstruction results. But their PSNR is very close with same algorithm. It is fair to say that the results are on the same level of image quality.

From Figure 2, it is shown that the TV-GD image and the TV-L_p image are very close in the image quality. Here we choose the FORBILD phantom [41], a more complicated and more challenging phantom, to further compare the proposed algorithm with the TV-GD algorithms. The phantom is shown in Figure 3. The scanning step of tomographic angels is set to 2.25°, 4°, 6° and 12°. So the sampling points are 160 views, 90 views, 60 views and 30 views correspondingly. The other numerical implementation conditions remain the same with the shep-logan simulation. The simulation results by using the TV-GD algorithms and the proposed TV-L_p algorithm are shown in Figure 4. It is shown in Figure 4 that, when the sampling data is sufficient, both algorithms can

Table 1 PSNRs (dB) of reconstructed images of shep-logan phantom

	160-view	90-view	30-view	18-view
FBP	20.35	18.36	15.68	13.14
L_1-norm	33.83	34.98	34.21	31.19
TV-GD	38.01	38.23	36.68	34.68
TV-L_p($p = 0.8$)	38.45	39.05	36.91	36.72
TV-L_p($p = 0.5$)	38.85	39.27	37.01	36.81

Figure 3 The FORBILD phantom.

Figure 4 The FORBILD phantom reconstruction results by different algorithms. The first to the third rows show images reconstructed by the TV-GD **(a-d)** and TV-Lp algorithm with p = 0.8 **(e-h)** and p = 0.5 **(i-l)**, respectively. The first to fourth columns refer to the results from 160-view, 90-view, 60-view, 30-view respectively.

reconstruct the accurate image. When the sampling angles get sparse during the simulation, it is seen that the TV-GD reconstruction results in paintinglike staircase artifacts in the smooth regions. Also it fails to give the accurate image in the low contrast regions in the top and left of the phantom. The proposed algorithm provides reasonably good reconstructions in these regions. The PSNR of the reconstructed images are shown in Table 2. From this table, we observe that the TV-L_p algorithm provides the better PSNR for all the cases. In the case of more complicated phantom, the TV-L_p algorithm shows significant improvement than the TV-GD algorithm.

Also, we include a line-plots image of the reconstruction result by the TV-GD algorithm and the TV-L_p($p = 0.8$) algorithm from 30-view data. The location of the pixel profile in the image is displayed in Figure 5(a). The comparisons of pixel profiles are displayed in the Figure 5(b).

In Figure 5(b), the solid line and the dotted line represent the pixel profiles of the TV-L_p and the TV-GD image respectively. It is shown in Figure 5(b) that the TV-L_p can reconstruct the image more precisely than the TV-GD one. The edges from the TV-L_p are sharper than that from the TV-GD. The pixel number from 90 to 100 is the high resolution area, the TV-L_p image shows the high-speed change of the pixel value while the TV-GD fails to do so. In the continuous area, the TV-L_p image is smoother.

We continuously decrease the number of the detect points try to find the limit density of the sampling points. During the simulation, we set the criterion of acceptability that the PSNR of the reconstructed image reaches 30 dB. It is found out that the total number of the sampling points is able to be reduced to 15 for TV-L_p algorithm in the reconstruction of shep-logan phantom and to 18 for the forbild phantom.

In this part, the TV-L_p algorithm is proved to be more accurate and stable than the other algorithms for PAI image reconstruction in the sparse sampling condition.

Convergence and calculation

In this part, we discuss the theoretical calculation complexity and study the convergence of the proposed algorithm. As mentioned above in 'Theory and method', in step (2) the update of ω and z is using the soft shrinkages and the computational costs are both $O(N)$. The update of z also includes a wavelet transforms which computational costs are $O(N\log N)$. In step (3), the update of u involves two fast Fourier transforms which computational costs are $O(N\log N)$ and two operations of A with the computational costs of $O(NMQ)$. The update of the parameters b and c in step (4) are all simple calculations with the computational costs of $O(N)$. As for the paremeter δ, although it involves an operation of A, it can use the result computed in the step (3). So its computational cost is also $O(N)$.

In a nutshell, the calculation complexity of the proposed algorithm in one iteration is $5O(N) + 4O(N\log N) + 2O(NMQ)$. The first two terms is much smaller than the last term in the practical use of photoacoustic imaging and most iterative algorithm is with

Table 2 PSNRs (dB) of reconstructed images of FORBILD phantom

	160-view	90-view	60-view	30-view
TV-GD	35.36	31.75	29.68	26.63
TV-L_p($p = 0.8$)	39.55	40.67	38.06	36.72
TV-L_p($p = 0.5$)	39.13	41.12	38.91	37.41

Figure 5 The gray-scale profiles. (a) The location of the pixel profile in the image. **(b)**The gray-scale profiles of the reconstructed images from 30-view data.

the operation. In each iteration, we just use the projection matrix twice. So the proposed algorithm has a cheap per iteration computation.

The TV-GD algorithm is reported as an efficient and stable iterative algorithm in photoacoustic imaging. In the 'Sparse-view reconstruction', its reconstruction result is closest to the proposed algorithm. So here we select it to be a comparison with the TV-L_p algorithm. We calculate the time cost of those two algorithms in a simulation. The simulation condition is same as in 'Sparse-view reconstruction'. But the iteration

ends when the PSNR values reach 30 dB. The result is shown in Table 3. From Table 3, it is shown that the proposed algorithm is faster than TV-GD algorithm in the computational time. Based on this result, it could be inferred that the TV-L_p algorithm is a more efficient image reconstruction algorithm comparing to the TV-GD algorithm. The value of p has also some influence on the time cost. The smaller the p is, the more iteration times are needed to reach the reconstruction result.

Thanks to the use of Barzilai-Borwein step size selection method, the convergence speed can also be significantly improved. For the quantitative analysis, we use a parameter that represents the distance between the reconstructed image and the original phantom image. The parameter d is defined as:

$$d = \left(\frac{\sum_{i=1}^{X} \sum_{j=1}^{Y} (u(i,j) - t(i,j))^2}{\sum_{i=1}^{X} \sum_{j=1}^{Y} t(i,j)^2} \right)^{1/2} \tag{23}$$

where u is the reconstructed image and t is the original image. The size of the image is $X \times Y$. The smaller the parameter d is, the closer the reconstructed image is with the original phantom. In the TV-L_p algorithm, there is a small rate of chance that the optimization will lead to the wrong solution due to its non-convex nature. So we use the original image to calculate the parameter d to show the image quality and use the parameter d as a reference. We want to show the improvement of the image quality in every iteration step.

The simulation condition is set to be the same as in 'Sparse-view reconstruction'. The sampling view is 60. The parameter p is set to be 0.8. The defined distance d is calculated after each iteration step. If the distance is smaller than 0.05, the iteration will stop. The simulation result is shown in Figure 6. The x-axis is the value of distance and the y-axis is the iteration times. The line '--' refers to the TV-GD algorithm and the line '*-' represent the TV-L_p ($p = 0.8$)algorithm. The result is shown in Figure 6. The images reconstructed by TV-L_p algorithm in each iteration have smaller value of d than the TV-GD ones and the TV-L_p iteration only takes 9 times as the distance is met the request.

For discussions noted above, it can be surmised that the convergence of TV-L_p algorithm is faster and the TV-L_p has a cheaper computational cost.

Robustness to the noise

In the practical applications of the photoacoustic tomography, the measurements are usually polluted by those white measuring noises from the ultrasound transducer and the system electronics. Hence, it is very important for an algorithm to maintain stable performance under the noises polluted circumstances. To analyze the robustness of the

Table 3 Time cost(second) of reconstructed images

	160-view	90-view	30-view	18-view
TV-GD	41.79	37.19	24.63	14.61
TV-L_p ($p = 0.8$)	18.45	12.05	6.29	4.41
TV-L_p ($p = 0.5$)	20.76	14.01	8.69	6.16

Figure 6 The distance between the reconstructed images and the original phantom image versus the iteration number.

TV-L_p algorithm, we choose 30-view simulated photoacoustic signals that we used in 'Sparse-view reconstruction'. The signals are added with white noises of different noise power levels. We use TV-L_p algorithm with two different settings of parameter p ($p = 0.5$ and $p = 0.8$) and TV-GD algorithm to reconstruct images from these white noise polluted measurements.

The reconstruction results are shown in Figure 7. In the first row to the last row of the Figure 7, the signal to noise ratio (SNR) of the polluted measurements is 10 dB, 5 dB, 3 dB and 0 dB, respectively. As shown in the image, when the power level of the noises is not very strong (10 dB and 5 dB), the reconstructed images by using the noisy measurements have basically no obviously difference with the ones reconstructed with the noiseless signals. As the noise becomes stronger, the quality of the reconstructed images decreases.

We also plot the profiles of a pixel line in order to show the detail qualities of recon-structed images clearly. In Figure 7, the dotted line and the solid line are the pixel pro-files of the reconstructed image and the original image, respectively. From the line plots, it is revealed that the proposed algorithm has better performance in the edge preservation and more accurate in the smooth area. We calculated the PSNR of the re-constructed image. The result is shown in Table 4. Our algorithm outperformed the TV-GD algorithm in any noise power level. Giving the credit to the optimal conditions, the reconstructed image is intended to be continuous and sharp. During the iteration, the photoacoustic signals get enhanced and the noise is suppressed.

As we can see from the table that the TV-L_p algorithm reconstructed images have a slightly better performance than the TV-GD algorithm when it comes to the image qualities. When the noise is extremely strong (0dB), the TV-L_p algorithm has a huge advantage than the TV-GD one in image quality. As for the different settings of para-meter p in the TV-L_p algorithm, there is no major difference that can be observed be-tween the two images. The PSNRs of the $p = 0.5$ setting is about 0.3 dB bigger than the

Figure 7 The reconstruction results from noisy signal. The first to the third columns show images reconstructed by the TV-GD **(a-d)** and TV-Lp algorithm with p = 0.8 **(e-h)** and p = 0.5 **(i-l)**, respectively, from signals with SNR of 10 dB, 5 dB, 3 dB and 0 dB.

$p = 0.8$ setting in the first three noise power level. But when the noise getting extremely strong (0 dB), the $p = 0.8$ setting is 0.1 dB bigger than the $p = 0.5$ setting in PSNR value.

From this part of simulation we can conclude that our TV-L_p algorithm is robust to noise and has a better performance than the TV-GD algorithm in the noisy measurement circumstance.

Parameter investigation

As the original optimization problem of the image reconstruction is described in Eq. (14), the TV-L_p algorithm contains some parameters that are tunable, which are α, β, ε, p and ρ. In those parameters, the choice of ρ does not affect the performance of the

Table 4 PSNRs (dB) of the noise simulation image

	10 dB	5 dB	3 dB	0 dB
TV-GD	32.24	28.01	22.44	16.96
TV-L_p ($p = 0.8$)	35.14	30.13	27.95	25.21
TV-L_p ($p = 0.5$)	35.63	30.40	28.10	25.06

TV-L_p algorithm theoretically. The result of the simulation also shows that the image quality is not sensitive to the parameter ρ for a large range. Here we set the parameter ρ to a steady value 1. The ε is the exiting condition parameter. It can be easily deduced that smaller ε will leads to slightly more accurate reconstructed image at the cost of more iteration times. In this part we focus on analyzing the parameter settings of *p*, α and β.

Parameter setting of p

In the TV-L_p algorithm, we replace the L_1 norm with the L_p norm $(0 < p \le 1)$. It is reported in Ref. [35] that theoretically fewer measurements are required for accurate reconstruction in the L_p norm situation. But it also leads to failure in solving the optimization problem. It's kind of a dilemma for the setting of *p*. So here we take different values of *p* to see its influence to the image reconstruction. The parameter α and β are both set to be 1×10^{-2}.

We choose the 90-view and 18-view simulated photoacoustic signals that we used in 'Sparse-view reconstruction'. We set the *p* value as 0.3, 0.5, 0.8 and 1. We calculate the reconstructed images' PSNR value. It is shown in Table 5. When the *p* is set as 0.5, it has advantage in the quality of the reconstructed image. However, when the value of *p* continues to reduce to 0.3, there is no obvious improvement in image quality. But in the same time, the smaller the *p* is, the higher probability of the solving failure is during the simulation. The reduction of *p* leads to the increasing of the iteration times in our simulation. So taking these two factors into account, we set the *p* as 0.8 so that it can provide a great reconstruction performance and stability with a fast convergence.

Parameter settings of α and β

As we describe above in Eq. (14), the parameter α and β are parameters corresponding to the weights of TV value and L_p norm value in this optimization problem respectively. We use these two parameters to balance the terms of the objective function. With different kinds of the objective image, the settings of those two parameters are different. Here we select three different images as the given optical energy deposition to test the universality of our algorithm and present further investigation of the parameter settings. We select a phantom that stand for the vessels and a phantom of dots with different energy degree in the simulation. We also choose a real brain MRI as the original optical energy deposition to demonstrate the performance of the proposed algorithm in reconstructing extremely detailed and complex structured imaging object. Here the TV-GD algorithm is used as a comparison. There are four groups of parameter settings, which are $(\alpha = 1 \times 10^{-2}, \beta = 5 \times 10^{-3})$, $(\alpha = 1 \times 10^{-2}, \beta = 1 \times 10^{-2})$, $(\alpha = 5 \times 10^{-2}, \beta = 5 \times 10^{-3})$ and $(\alpha = 5 \times 10^{-3}, \beta = 5 \times 10^{-3})$. The reconstructed images are shown in Figure 8. From first row of Figure 6, we can see in the reconstruction of gradient sparse phantom.

Table 5 PSNRs (dB) of reconstructed images with different *p*

	90-view	18-view
p = 0.3	39.26	36.62
p = 0.5	39.27	36.81
p = 0.8	39.05	36.72
p = 1	37.65	34.79

Figure 8 The different image objectives reconstructed by TV-GD and TV-Lp algorithms. The first to sixth rows refer to the original phantom **(a, g, m)**, the TV-GD reconstructed images **(b, h, n)**, the TV-Lp reconstructed images with parameter settings as ($\alpha = 1 \times 10^{-2}$, $\beta = 5 \times 10^{-3}$) **(c, i, o)**, ($\alpha = 1 \times 10^{-2}$, $\beta = 1 \times 10^{-2}$) (d, j, p), ($\alpha = 5 \times 10^{-2}$, $\beta = 5 \times 10^{-3}$) **(e, k, q)** and ($\alpha = 5 \times 10^{-3}$, $\beta = 5 \times 10^{-3}$) **(f, l, r)**, respectively.

The TV based algorithm has great performance when the image demonstrate piecewise continuous behavior. All reconstructions are accurate and the background noise is suppressed well. When it comes to the images with the vessel phantom (Figure 8 (g)-(l)), those original optical energy depositions are a little bit more complex than the dots. The reconstruction results show that the image reconstructed by TV-L_p algorithm is better than the TV-GD ones. As in Figure 8(h), TV-GD image has some noises in the background and the edge of the vessel is blurred. While the TV-L_p images with different parameter settings both have high-resolution results. As for the real MRI image, it has very detailed information. As expected, both two groups of parameter setting $\alpha = \beta$ have the most accurate result among them. The increasing weighting of L_p-norm condition can provide more detail information and prevent the reconstructed image emerging plantlike artifacts. The details such as edges and fine structures are well preserved in both reconstructions. The reconstruction results show that the TV-GD reconstructed image has severely paintinglike staircase artifacts with some loss in fine details. From our observation, TV-L_p algorithm with the parameter setting $\alpha = \beta$ preserves the fine features better than the TV-GD one. α and β are the regularization parameters determining the trade-off between the data consistency and the sparsity. It is revealed from the above simulation that the parameter setting $\alpha = \beta$ is a better strategy. In this parameter setting, the TV-L_p algorithm provides a 3 dB improvement in the PSNR over the TV-GD algorithm based on our calculation.

Limited-view and irregular-view simulation
In the real application of PAI, due to the restrains of the shape or the size of the imaging object, a full angular scanning sometimes is hard to achieve. We evaluate the performance of the TV-L_p method in limited-view case, line-view case and unequal-view case.

The simulation setup and the reconstructed image is shown in Figure 9. In the limited-view simulation (Figure 9(a)), the scanning angular range is set to 150° and the angular step is 3°, so 50-view photoacoustic signals are obtained. In the line-view simulation (Figure 9(c)), the transducer array with 60 transducers is placed in the right side of the imaging object and the interval between two transducer elements is 1.49 mm. It

Figure 9 Limited-view, line-view and un-equal angel step scanning reconstruction results. (a) refers to the location of transducers in the limited-view simulation. **(b)** refers to the reconstructed image from the limited-view data. **(c)** refers to the location of transducers in the line-view simulation. **(d)** refers to the reconstructed image from the line-view data. **(e)** refers to the location of transducers in the un-equal angel step scanning. **(f)** refers to the reconstructed image from the un-equal angel step scanning data.

is revealed in Figure 9(b) and Figure 9(d) that the quality of the TV-L_p reconstruction is not much affected by the limitation of the sampling angle. Because the sampling angle is limited, the information definite is partly missed, yet the TV-L_p method can still provide a satisfying reconstruction. In un-equal angel step scanning, we randomly choose 30 sampling points from a 60-view projection and use these 30-view un-equal angle step data for image reconstruction. The result is shown in Figure 9(f). As we can see from the image, the reconstruction result can still maintain a very high quality.

In-vitro experiments
Experiment setup
We carry out the experiments on *in-vitro* signals to demonstrate the proposed TV-L_p algorithm's performance in the practical application.

The framework of the experiment platform is shown in Figure 10. In this platform, an Nd:YAG laser generator (Continum, Surelite I) is used to emit the laser pulse. The wavelength of the laser is 532 nm. A single laser pulse is generated at the frequency of 10 Hz and last 6-7 ns. The incident laser pulse is emitted towards the top of the phantom through a concave lens with the diameter of 5 cm. The setup of the lens enlarges the illumination area and lead to the pulse energy reduction in the illumination area. The energy is about 6.47 mJcm^{-2}, which is lower than the ANSI laser radiation safety standard (20 mJcm^{-2}) [1]. Signal acquisition is done by a water-immersion ultrasound transducer (Panametric, V383-SU). The transducer is a linearly unfocused one at 3.5 MHz (-6 dB bandwidth at 45%). A digital stepping motor (GCD-0301 M) is used to rotate the ultrasound transducer around the phantom placed in water. The scanning radius is 38 mm. The received analog ultrasound signals are amplified by a pulse receiver (Panametric, 5900PR). An oscilloscope (Agilent, 54622D) with the sampling frequency of 16.67 MHz is set to transform the received signals into digital ones. Both the laser generator and the digital motor are controlled by the computer through the serial interface.


Figure 10 The scheme of the experiment platform.

The transformed digital data is transported to the computer through the general purpose interface bus (GPIB).

The imaged phantom we used in the experiment is made by gelatin cylinder. It is shown in Figure 11. There are two different phantoms. The radius of the phantom is 25 mm. The left one is made by two rubber bars with 1 mm diameter that embedded as the optical absorbers. The right one utilizes leaf which pretends as vein and tissue as the optical absorbers.

In the experiment, the transducer tends to measure the photoacoustic signal in-plane only, and the reconstruction is also in 2D. The cross-sectional image in any plane is mainly determined by the measured data in the same plane, and a set of circular measurement data on the same plane would be sufficient to reconstruct a good image. We use the deconvlution calculation before the reconstruction to eliminate the transducer's impulse response influence.

Experiment result

In the experiment, 90-view and 30-view data are collected for reconstruction. The images are constructed by the FBP, the TV-GD and the TV-L_p algorithms, respectively. The reconstruction results are shown in Figure 12. The left column of the Figure 12 is reconstructed from 90-view data. When the sampling data is sufficient, all three algorithms are effective. With respect to the locations and sizes, the optical absorbers are all well reconstructed in the figure. While the FBP reconstructed image is not as clear as the images reconstructed by the iterative algorithms. When we reconstruct the image with a small of sampling angles (right column of Figure 11), the artifacts start to emerge in the FBP reconstructed image and the quality of the image is severely affected. But the TV-GD and the TV-L_p algorithms can still provide high-contrast images with less noise. In Figure 12(f), it is shown that the image reconstructed by the TV-L_p algorithm outperforms other algorithms in image contrast and noise suppression. The structure of optical absorbers is clear and the noise in the background is well-suppressed. The sparse-view of sampling has barely any influence on the quality of the TV-L_p reconstructed image.

In vitro imaging of a leaf vein is also performed to further demonstrate the advantages of the TV-L_p algorithm. The reconstruction result is shown in Figure 13. As the


Figure 11 The photos of the imaging samples.

Figure 12 The rubber phantom reconstruction results from experiment data. From left to right, the reconstructed images of rubber sample from 90-view and 30-view experiment data respectively. The first to third rows refer to the images reconstructed by the FBP **(a, b)**, the TV-GD **(c, d)** and the TV-Lp **(e, f)** algorithm respectively.

structure of the phantom is more complex, the FBP are deeply influenced by the artifact and fail to reconstruct the accurate image both under 90-view and 30-view sampling circumstance. It is shown in the figure, that the TV-GD and TV-L_p algorithms can still reconstruct the image in a high contrast level. But when the data is insufficient, there is some noise emerging in the background. TV-L_p algorithms can suppress the noise better than the TV-GD one. The optical absorber in TV-L_p one is more distinct than that in the image by TV-GD algorithm.

Quantitative comparisons

We use the L_1-norm algorithm to reconstruct the image of 180-view data from the leaf vein phantom. As the sampling view is efficient, the reconstructed image is used as a "standard" one. We calculate the histograms of the difference between the reconstructed

Figure 13 The leaf vein phantom reconstruction results from experiment data. From left to right, the reconstructed images of leaf vein phantom from 90-view and 30-view sampling experiment data respectively. The first to third rows refer to the images reconstructed by the FBP **(a, b)**, the TV-GD **(c, d)** and the TV-Lp **(e, f)** algorithm respectively.

one and the "standard" one as shown in Figure 14. Figure 14 (a)-(c) are the difference histograms between the standard and the images reconstructed by the FBP, the TV-GD and the TV-L_p, respectively, with 30-view data. In Figure 14, two CS-based algorithms have a large number of pixels with small ranges of difference with the standard one, which suggests that these two algorithms can reconstruct the image more accurately. In the case of the TV-L_p algorithm, the major part of the pixel difference is in the range from 0 to 0.1.

Figure 14 (See legend on next page.)

> (See figure on previous page.)
> **Figure 14 The amplitudes histograms of the difference between reconstruction results and the "standard" one.** We use the image reconstructed by L1-norm with data from all 180 transducer elements as the standard. **(a)** Histograms image of the FBP algorithm. **(b)** Histograms image of the TV-GD algorithm. **(c)** Histograms image of the TV-Lp algorithm.

From this experiment, the results demonstrate that the TV-L_p method can outperform the TV-GD one in the field of the image quality.

From the experiment result noted above, it is safe to say that the TV-L_p algorithm would have better performance in sparse-view PAI than other algorithms. It could provide stable and accurate reconstruction in both sufficient data sampling and sparse-view sampling situation.

Conclusion

Aiming to reduce the scanning time and enhance the imaging quality of the photo-acoustic image reconstruction, we proposed the TV-L_p algorithm that applies the total variation method and nonconvex optimization method to the PAI. The main idea of the algorithm is to apply L_p-norm nonconvex optimization along with the total variation method. In the proposed algorithm, the Barzilai-Borwein step size selection method is adopted to provide faster convergence and smaller calculation. The effectiveness and universality of the algorithm is demonstrated through the numerical simulations. The numerical simulations show that the TV-L_p algorithm provides good imaging quality in sparse-view sampling situation. The algorithm convergence, the robustness to noise and the tunable parameters are also discussed. The simulation result reveals that the TV-L_p algorithm is a stable image reconstruction method with fast convergence and small computational cost. The TV-L_p algorithm is further investigated through some experiments using gelatin-made phantom. Compared with the result of other popular image reconstruction method, the TV-L_p imaging algorithm has significant advantage on contrast and noise suppression. From the discussion noted above, it could be concluded that the TV-L_p algorithm may be a practical algorithm for sparse-view photoacoustic imaging reconstruction.

References

1. Wang LV: **Tutorial on photoacoustic microscopy and computed tomography.** *IEEE J Sel Top Quantum Electron* 2008, **14**(1):171–179.
2. Xu M, Wang LV: **Photoacoustic imaging in biomedicine.** *Rev Sci Instrum* 2006, **77**(4):041101-1–041101-22.
3. Li C, Wang LV: **Photoacoustic tomography and sensing in biomedicine.** *Phys Med Biol* 2009, **54**(19):R59–R97.
4. Wang LV: **Prospects of photoacoustic tomography.** *Med Phys* 2008, **35**(12):5758–5767.
5. Kruger RA, Reinecke DR, Kruger GA: **Thermoacoustic computed tomography—technical considerations.** *Med Phys* 1999, **26**(9):1832–1837.
6. Kruger RA, Liu P, Fang Y, Appledorn CR: **Photoacoustic ultrasound (PAUS)-reconstruction tomography.** *Med Phys* 1995, **22**(10):1605–1609.
7. Guo B, Li J, Zmuda H, Sheplak M: **Multyfrequency microwave-induced thermal acoustic imaging for breast cancer detection.** *IEEE Trans Ultrason Ferroelectr Freq Control* 2007, **54**(11):2000–2010.
8. Pramanik M, Ku G, Li C, Wang LV: **Design and evaluation of a novel breast cancer detection system combining both thermoacoustic (TA) and photoacoustic (PA) tomography.** *Med Phys* 2008, **35**(6):2218–2223.
9. Zhang EZ, Laufer JG, Pedley RB, Beard PC: *In vivo* **high-resolution 3D photoacoustic imaging of superficial vascular anatomy.** *Phys Med Biol* 2009, **54**(4):1035–1046.
10. Niederhauser JJ, Jaeger M, Lemor R, Weber P, Frenz M: **Combined ultrasound and optoacoustic system for real-time high-contrast vascular imaging** *in vivo.* *IEEE Trans Med Imaging* 2005, **24**(4):436–440.
11. Zerda A, Paulus YM, Teed R, Bodapati S, Dollberg Y, Khuri-Yakub BT, Blumenkranz BS, Moshfeghi DM, Gambhir SS: **Photoacoustic ocular imaging.** *Opt Lett* 2010, **35**(3):270–272.
12. Xu M, Wang LV: **Time-domain reconstruction for thermoacoustic tomography in a spherical geometry.** *IEEE Trans Med Imaging* 2002, **21**(7):814–822.
13. Xu M, Xu Y, Wang LV: **Time-domain reconstruction algorithms and numerical simulations for thermoacoustic tomography in various geometries.** *IEEE Trans Biomed Eng* 2003, **50**(9):1086–1099.
14. Xu Y, Feng DZ, Wang LV: **Exact frequency-domain reconstruction for thermoacoustic tomography—I: Planar geometry.** *IEEE Trans Med Imaging* 2002, **21**(7):823–828.
15. Xu Y, Xu M, Wang LV: **Exact frequency-domain reconstruction for thermoacoustic tomography—II: Cylindrical geometry.** *IEEE Trans Med Imaging* 2002, **21**(7):829–833.
16. Xu M, Wang LV: **Pulsed-microwave-induced thermoacoustic tomography: Filtered back-projection in a circular measurement configuration.** *Med Phys* 2002, **29**(8):1661–1669.
17. Xu M, Wang LV: **Universal back-projection algorithm for photoacoustic computed tomography.** *Phys Rev E* 2005, **71**(1):016706-1–016706-7.
18. Zhang C, Wang Y: **Deconvolution reconstruction of full-view and limited-view photoacoustic tomography: a simulation study.** *J Opt Soc Am A* 2008, **25**(10):2436–2443.
19. Zhang C, Li C, Wang LV: **Fast and robust deconvolution-based image reconstruction for photoacoustic tomography in circular geometry experimental validation.** *IEEE Photonics J* 2010, **2**(1):57–66.
20. Liao CK, Li ML, Li PC: **Optoacoustic imaging with synthetic aperture focusing and coherence weighting.** *Opt Lett* 2004, **29**(21):2506–2508.
21. Modgil D, Rivière PJ: **Implementation and comparison of reconstruction algorithms for 2D optoacoustic tomography using a linear array.** In *Proceedings of SPIE 6856:2008; San Jose.* 2008:68561D-1–68561D-12.
22. Zhang J, Anastasio MA, Riviere PJ, Wang LV: **Effects of different imaging models on least-squares image reconstruction accuracy in photoacoustic tomography.** *IEEE Trans Med Imaging* 2009, **28**(11):1781–1790.
23. Paltauf G, Viator JA, Prahl SA, Jacques SL: **Iterative reconstruction algorithm for optoacoustic imaging.** *J Acoust Soc Am* 2002, **112**(4):1536–1544.
24. Jose J, Willemink R, Steenbergen W, Slump C, van Leeuwen T, Manohar S: **Speed-of-sound compensated photoacoustic tomography for accurate imaging.** *Med Phys* 2012, **39**:7262–7281.
25. Huang C, Wang K, Nie L, Wang LV, Anastasio M: **Full-Wave Iterative Image Reconstruction in Photoacoustic Tomography With Acoustically Inhomogeneous Media.** *IEEE Trans Med Imaging* 2013, **32**(6):1097–1110.
26. Provost J, Lesage F: **The application of compressed sensing for photo-acoustic tomography.** *IEEE Trans Med Imaging* 2009, **28**(4):585–594.
27. Guo Z, Li C, Song L, Wang LV: **Compressed sensing in photoacoustic tomography in vivo.** *J Biomed Opt* 2010, **15**(2):021311-1–021311-6.
28. Rosenthal A, Jetzfellner T, Razansky D, Ntziachristos V: **Efficient framework for model-based tomographic image reconstruction using wavelet packets.** *IEEE Trans Med Imaging* 2012, **31**(7):1346–1357.
29. Meng J, Wang LV, Ying L, Liang D, Song L: **Compressed-sensing photoacoustic computed tomography in vivo with partially known support.** *Opt Express* 2012, **20**(15):16510–16523.
30. Meng J, Wang LV, Ying L, Liang D, Song L: *In vivo* **optical-resolution photoacoustic computed tomography with compressed sensing.** *Opt Lett* 2012, **37**(22):4573–4575.
31. Yao L, Jiang H: **Photoacoustic image reconstruction from few-detector and limited-angle data.** *Biomed Opt Express* 2011, **2**(9):2649–2654.
32. Yao L, Jiang H: **Enhancing finite element-based photoacoustic tomography using total variation minimization.** *Appl Optics* 2011, **50**(25):5031–5041.
33. Wang K, Su R, Oraevsky AA, Anastasio M: **Investigation of iterative image reconstruction in three-dimensional optoacoustic tomography.** *Phys Med Biol* 2012, **57**(17):5399–5423.
34. Zhang Y, Wang Y, Zhang C: **Total variation based gradient descent algorithm for sparse-view photoacoustic image reconstruction.** *Ultrasonics* 2012, **52**(8):1046–1055.
35. Chartrand R: **Exact reconstruction of sparse signals via nonconvex minimization.** *IEEE Trans Signal Proc Let* 2007, **14**(10):707–710.
36. Majumdar A, Ward RK: **An algorithm for sparse MRI reconstruction by Schatten *p*-norm minization.** *Magn Reson Imaging* 2011, **29**(3):408–417.

37. Majumdar A: **Improved dynamic MRI reconstruction by exploiting sparsity and rank-deficiency.** *Magn Reson Imaging* 2013, **31**(5):789–795.
38. Ma S, Yin W, Zhang Y, Chakraborty A: **An efficient algorithm for compressed MR imaging using total variation and wavelets.** In *Proceedings IEEE Conference on Computer Vision Pattern Recongnition: 2008; Anchorage.* 2008:1–8.
39. Ye X, Chen Y, Huang F: **Computational acceleration for MR image reconstruction in partially parallel imaging.** *IEEE Trans Med Imaging* 2011, **30**(5):1055–1063.
40. Finch D, Haltmeier M, Rakesh: **Inversion of spherical means and the wave equation in even dimensions.** *SIAM J Appl Math* 2007, **68**(2):392–412.
41. Yu Z, Noo F, Dennerlein F, Wunderlich A, Lauritsch G, Hornegger J: **Simulation tools for two-dimensional experiments in x-ray computed tomography using the FORBILD head phantom.** *Phys Med Biol* 2012, **57**(13):237–252.