

TempCo

Analysis implemented	TPCB	VTEMP	CTEMP	VHTR	CHTR	RFL	PDSU	TEMPCO	VQPA	VQP
Standard deviation		>0.01	> .001		>300*				N/A	
Average										
Minimum				5k	10k	14k	325			
Maximum				60k 62k	58k 60k				0 if 1 st run 60 if 2 nd	
Max - Min							350			200 in 2 nd

Items in red fail if criteria is met or exceeded

Items in black need further inspection to determine pass/fail

TempCo

Analysis implemented	FFE	DCL	PDADC	CPTP	CPTF	CPTD	VBC	VFA	XFA
Standard deviation		> .002V				>2* median			
Average									
Minimum		0				100			10k
Maximum									56k
Max - Min	5e-10 2e-10					20% of avg. at 30C		15k 1 st run	16k

Items in red fail if criteria is met or exceeded

Items in black need further inspection to determine pass/fail