Device Comfort and Usability Survey: Physiotherapist

This survey aims to evaluate the comfort, usability, and overall user experience of the stroke rehabilitation device. There are no right or wrong answers. Your honest feedback will help improve the design and usability of the device. All responses will be kept confidential.

Participant Information

Name: Ling Wei Hsuen
Age: 23
Gender: Female
Occupation: Student
Date of Test: 2 Apr 2026
Tester Name: Nadia and Ron
Total Time Using Device: 40 mins

Section A: General Usability

Please rate the following statements on a scale of 1 (Strongly Disagree) to 5 (Strongly Agree).

The device is easy to operate for patient setup. Rating (1–5): 5 Comments: The device is simple and easy to use
The system functions are well integrated. Rating (1–5): 5 Comments: All the system features are collated and easy to navigate.
The device fits well into clinical workflow. Rating (1–5): 5 Comments: The device works well for warm up and treatments.
The controls/interface were intuitive. Rating (1–5): 5 Comments: The UI and buttons were very intuitive for a new user to navigate.
I felt confident operating the device. Rating (1–5): 5 Comments: The intuitive UI makes me confident in using it.

Section B: Setup, Adjustment and Handling

The device is easy to set up for patients. Rating (1–5): 5 Comments: It was easy to set up the device.
Adjustments (e.g. leg length, straps) are efficient and intuitive. Rating (1–5): 5 Comments: The straps were intuitive to help the patient in.
Getting patients into the device is safe and manageable. Rating (1–5): 5 Comments: It was quite effortless to get the patient into the device.

Section C: Fit and Physical Comfort

The device accommodates different leg sizes well. Rating (1–5): 5 Comments: The straps in the device is accommodating.
The device provides adequate support and alignment. Rating (1–5): 5 Comments: The device holds up the patient's leg well.
The fit can be adjusted to avoid pressure points. Rating (1–5): 5 Comments: Fit can be adjusted by shifting or using the straps.

Section D: Exercise Experience and Effectiveness

The device enables clinically appropriate movement patterns. Rating (1–5): 5 Comments: There were a large variety of warm up and exercises available.
The range of motion is adequate for rehabilitation goals. Rating (1–5): 5 Comments: The range of motion allows easy warm ups and treatments.
The device supports passive exercises effectively. Rating (1–5): 5 Comments:
The device supports active exercises effectively. Rating (1–5): 5 Comments:
The device reduces therapist physical workload. Rating (1–5): 5 Comments:

Section E: Safety and Confidence

The device appears safe for patient use. Rating (1–5): 5 Comments:
The device remains stable during operation. Rating (1–5): 5 Comments: The device moves as required.
Emergency controls are accessible and reliable. Rating (1–5): 5 Comments: The emergency buttons on the app is intuitive.
I trust the device for supervised use. Rating (1–5): 5 Comments: I agree.
I trust the device for unsupervised use. Rating (1–5): 5 Comments: I agree.

Section F: Device Design and Practicality

The device is practical for rehabilitation use (e.g. space, storage). Rating (1–5): 5 Comments: It is small enough and light enough for storage.
The weight and portability are acceptable. Rating (1–5): 5 Comments: The device is easy to move when necessary.
The device is durable and reliable. Rating (1–5): 5 Comments: The device works as needed.
The device is easy to disinfect (e.g. wipe down) and maintain. Rating (1–5): 5 Comments: It was easy to wipe it down with the wipes, to keep the device clean.

Section G: Effort and Workload

Operating the device reduces physical strain. Rating (1–5): 5 Comments: The machine moves the patient on its own.
The device reduces workload during therapy sessions. Rating (1–5): 5 Comments: The devices gives me time to do other things as it works by itself.
The device is not mentally demanding to operate. Rating (1–5): 5 Comments: The device moves on it's own and allows me to work on other patients while it runs.

Section G: Overall Feedback

What did you like most about the device?
What did you find uncomfortable or difficult?
What improvements would you suggest?
Would you be willing to use this device regularly? (Yes / No)

Remarks: This survey is an adapted hybrid usability survey based on recognised standards and validated constructs, refined using current lower-limb rehabilitation-device literature.

References

[1] J. Brooke, “SUS: A quick and dirty usability scale,” in *Usability Evaluation in Industry*, P. W. Jordan, B. Thomas, B. A. Weerdmeester, and A. L. McClelland, Eds. London, U.K.: Taylor & Francis, 1996, pp. 189–194.

[2] L. Demers, R. Weiss-Lambrou, and B. Ska, “The Quebec User Evaluation of Satisfaction with Assistive Technology (QUEST 2.0): An overview and recent progress,” *Disability and Rehabilitation*, vol. 24, no. 1–3, pp. 21–30, 2002.

[3] S. G. Hart and L. E. Staveland, “Development of NASA-TLX (Task Load Index): Results of empirical and theoretical research,” in *Advances in Psychology*, vol. 52, P. A. Hancock and N. Meshkati, Eds. Amsterdam, Netherlands: North-Holland, 1988, pp. 139–183.

[4] ISO 62366-1, *Medical devices—Application of usability engineering to medical devices—Part 1: Application of usability engineering to medical devices*, Geneva, Switzerland: International Organization for Standardization, 2015.

[5] ISO 62366-1 Amendment 1, *Medical devices—Application of usability engineering to medical devices—Amendment 1*, Geneva, Switzerland: International Organization for Standardization, 2020.

[6] D. R. Louie and J. J. Eng, “Powered robotic exoskeletons in post-stroke rehabilitation of gait: A scoping review,” *Journal of NeuroEngineering and Rehabilitation*, vol. 13, no. 53, 2016.

Device Comfort and Usability Survey: Physiotherapist

This survey aims to evaluate the comfort, usability, and overall user experience of the stroke rehabilitation device. There are no right or wrong answers. Your honest feedback will help improve the design and usability of the device. All responses will be kept confidential.

Participant Information

Name: Chloe Wong
Age: 22
Gender: Female
Occupation: Student
Date of Test: 3 Apr 2026
Tester Name: Nadia and Ron
Total Time Using Device: 15 minutes

Section A: General Usability

Please rate the following statements on a scale of 1 (Strongly Disagree) to 5 (Strongly Agree).

The device is easy to operate for patient setup. Rating (1–5): __4__ Comments: _____
The system functions are well integrated. Rating (1–5): __4__ Comments: _____
The controls/interface were intuitive. Rating (1–5): __5__ Comments: __UI is very simple and basic_____
I felt confident operating the device. Rating (1–5): __4__ Comments: _____

Section B: Setup, Adjustment and Handling

The device is easy to set up for patients. Rating (1–5): __4__ Comments: _____
Adjustments (e.g. leg length, straps) are efficient and intuitive. Rating (1–5): __3__ Comments: Might be a little difficult for patients who are shorter/have shorter legs
Getting patients into the device is safe and manageable. Rating (1–5): __4__ Comments: Might depend on how mobile the patient is but I think it would be much easier compared to traditional physio techniques

Section C: Fit and Physical Comfort

The device provides adequate support and alignment. Rating (1–5): __3___ Comments: _My leg is a bit short, so when it extends it might feel a bit too long
The fit can be adjusted to avoid pressure points. Rating (1–5): __5___ Comments: _____

Section D: Exercise Experience and Effectiveness

The device reduces therapist physical workload. Rating (1–5): __4___ Comments: _____

Section E: Safety and Confidence

The device appears safe for patient use. Rating (1–5): __5___ Comments: _____
The device remains stable during operation. Rating (1–5): __3___ Comments: It was a bit shaky at times
Emergency controls are accessible and reliable. Rating (1–5): __5___ Comments: _____
I trust the device for supervised use. Rating (1–5): __5___ Comments: _____

Section F: Device Design and Practicality

The device is durable and reliable. Rating (1–5): __3___ Comments: _____
The device is easy to disinfect (e.g. wipe down) and maintain. Rating (1–5): ___5__ Comments: _____

Section G: Effort and Workload

Operating the device reduces physical strain. Rating (1–5): __4___ Comments: _____
The device reduces workload during therapy sessions. Rating (1–5): __4___ Comments: _____
The device is not mentally demanding to operate. Rating (1–5): __1___ Comments: _____

Section G: Overall Feedback

What did you like most about the device? Very easy to set up and use; intuitive and simple UI; also does not require much effort from both physiotherapist and patient
What did you find uncomfortable or difficult?
What improvements would you suggest?

The setup for the device could be more inclusive for people of different sizes and proportions
Would you be willing to use this device regularly if recommended by a therapist? (Yes / No) Yes

Remarks: This survey is an adapted hybrid usability survey based on recognised standards and validated constructs, refined using current lower-limb rehabilitation-device literature.

References

[1] J. Brooke, “SUS: A quick and dirty usability scale,” in *Usability Evaluation in Industry*, P. W. Jordan, B. Thomas, B. A. Weerdmeester, and A. L. McClelland, Eds. London, U.K.: Taylor & Francis, 1996, pp. 189–194.

[2] L. Demers, R. Weiss-Lambrou, and B. Ska, “The Quebec User Evaluation of Satisfaction with Assistive Technology (QUEST 2.0): An overview and recent progress,” *Disability and Rehabilitation*, vol. 24, no. 1–3, pp. 21–30, 2002.

[3] S. G. Hart and L. E. Staveland, “Development of NASA-TLX (Task Load Index): Results of empirical and theoretical research,” in *Advances in Psychology*, vol. 52, P. A. Hancock and N. Meshkati, Eds. Amsterdam, Netherlands: North-Holland, 1988, pp. 139–183.

[4] ISO 62366-1, *Medical devices—Application of usability engineering to medical devices—Part 1: Application of usability engineering to medical devices*, Geneva, Switzerland: International Organization for Standardization, 2015.

[5] ISO 62366-1 Amendment 1, *Medical devices—Application of usability engineering to medical devices—Amendment 1*, Geneva, Switzerland: International Organization for Standardization, 2020.

[6] D. R. Louie and J. J. Eng, “Powered robotic exoskeletons in post-stroke rehabilitation of gait: A scoping review,” *Journal of NeuroEngineering and Rehabilitation*, vol. 13, no. 53, 2016.

Device Comfort and Usability Survey: Physiotherapist

This survey aims to evaluate the comfort, usability, and overall user experience of the stroke rehabilitation device. There are no right or wrong answers. Your honest feedback will help improve the design and usability of the device. All responses will be kept confidential.

Participant Information

Name: Abigail Tan Huey Ern
Age: 22
Gender: Female
Occupation: Student
Date of Test: 3 Apr 2026
Tester Name:
Total Time Using Device:

Section A: General Usability

Please rate the following statements on a scale of 1 (Strongly Disagree) to 5 (Strongly Agree).

The device is easy to operate for patient setup. Rating (1–5): __5__ Comments: _____
The system functions are well integrated. Rating (1–5): __4__ Comments: _____
The controls/interface were intuitive. Rating (1–5): __5__ Comments: _____
I felt confident operating the device. Rating (1–5): __5__ Comments: _____

Section B: Setup, Adjustment and Handling

The device is easy to set up for patients. Rating (1–5): __5__ Comments: _____
Adjustments (e.g. leg length, straps) are efficient and intuitive. Rating (1–5): __5__ Comments: _____
Getting patients into the device is safe and manageable. Rating (1–5): __3__ Comments: ___Depends on their flexibility_____

Section C: Fit and Physical Comfort

The device provides adequate support and alignment. Rating (1–5): __4___ Comments: _____
The fit can be adjusted to avoid pressure points. Rating (1–5): ___2__ Comments: ____Nothing much I can do about the fit_____

Section D: Exercise Experience and Effectiveness

The device reduces therapist physical workload. Rating (1–5): __4___ Comments: _____

Section E: Safety and Confidence

The device appears safe for patient use. Rating (1–5): __3___ Comments: _____
The device remains stable during operation. Rating (1–5): __3___ Comments: _____
Emergency controls are accessible and reliable. Rating (1–5): __5___ Comments: _____
I trust the device for supervised use. Rating (1–5): __4___ Comments: _____

Section F: Device Design and Practicality

The device is durable and reliable. Rating (1–5): __3___ Comments: _____
The device is easy to disinfect (e.g. wipe down) and maintain. Rating (1–5): __5___ Comments: _____

Section G: Effort and Workload

Operating the device reduces physical strain. Rating (1–5): __4___ Comments: _____Just the lifting of the patient's leg_____
The device reduces workload during therapy sessions. Rating (1–5): ___5__ Comments: _____
The device is not mentally demanding to operate. Rating (1–5): ___5__ Comments: _____

Section G: Overall Feedback

What did you like most about the device? I just have to press buttons and it will generate the chart for me
What did you find uncomfortable or difficult? Nothing
What improvements would you suggest? A space to prop up the tablet for when I leave and within the patient's reach for the stop button.

Would you be willing to use this device regularly if recommended by a therapist? (Yes / No) Yes

Remarks: This survey is an adapted hybrid usability survey based on recognised standards and validated constructs, refined using current lower-limb rehabilitation-device literature.

References

[1] J. Brooke, “SUS: A quick and dirty usability scale,” in *Usability Evaluation in Industry*, P. W. Jordan, B. Thomas, B. A. Weerdmeester, and A. L. McClelland, Eds. London, U.K.: Taylor & Francis, 1996, pp. 189–194.

[2] L. Demers, R. Weiss-Lambrou, and B. Ska, “The Quebec User Evaluation of Satisfaction with Assistive Technology (QUEST 2.0): An overview and recent progress,” *Disability and Rehabilitation*, vol. 24, no. 1–3, pp. 21–30, 2002.

[3] S. G. Hart and L. E. Staveland, “Development of NASA-TLX (Task Load Index): Results of empirical and theoretical research,” in *Advances in Psychology*, vol. 52, P. A. Hancock and N. Meshkati, Eds. Amsterdam, Netherlands: North-Holland, 1988, pp. 139–183.

[4] ISO 62366-1, *Medical devices—Application of usability engineering to medical devices—Part 1: Application of usability engineering to medical devices*, Geneva, Switzerland: International Organization for Standardization, 2015.

[5] ISO 62366-1 Amendment 1, *Medical devices—Application of usability engineering to medical devices—Amendment 1*, Geneva, Switzerland: International Organization for Standardization, 2020.

[6] D. R. Louie and J. J. Eng, “Powered robotic exoskeletons in post-stroke rehabilitation of gait: A scoping review,” *Journal of NeuroEngineering and Rehabilitation*, vol. 13, no. 53, 2016.

Device Comfort and Usability Survey: Physiotherapist

This survey aims to evaluate the comfort, usability, and overall user experience of the stroke rehabilitation device. There are no right or wrong answers. Your honest feedback will help improve the design and usability of the device. All responses will be kept confidential.

Participant Information

Name: HUANG ZICHEN
Age: 19
Gender: Male
Occupation: Student
Date of Test: 3 Apr 2026
Tester Name: Ron and Nadia
Total Time Using Device: 15 mins

Section A: General Usability

Please rate the following statements on a scale of 1 (Strongly Disagree) to 5 (Strongly Agree).

The device is easy to operate for patient setup. Rating (1–5): __5___ Comments: _____
The system functions are well integrated. Rating (1–5): ___5__ Comments: _____
The controls/interface were intuitive. Rating (1–5): ___5__ Comments: _____
I felt confident operating the device. Rating (1–5): ___5__ Comments: _____

Section B: Setup, Adjustment and Handling

The device is easy to set up for patients. Rating (1–5): __4___ Comments: _____
Adjustments (e.g. leg length, straps) are efficient and intuitive. Rating (1–5): __3___ Comments: _____
Getting patients into the device is safe and manageable. Rating (1–5): __5___ Comments: _____

Section C: Fit and Physical Comfort

The device provides adequate support and alignment. Rating (1–5): __5__ Comments: _____
The fit can be adjusted to avoid pressure points. Rating (1–5): _5___ Comments: _____

Section D: Exercise Experience and Effectiveness

The device reduces therapist physical workload. Rating (1–5): __5__ Comments: _____

Section E: Safety and Confidence

The device appears safe for patient use. Rating (1–5): _5___ Comments: _____
The device remains stable during operation. Rating (1–5): ___5_ Comments: _____
Emergency controls are accessible and reliable. Rating (1–5): __5__ Comments: _____
I trust the device for supervised use. Rating (1–5): __5__ Comments: _____

Section F: Device Design and Practicality

The device is durable and reliable. Rating (1–5): __4__ Comments: _____
The device is easy to disinfect (e.g. wipe down) and maintain. Rating (1–5): __5__ Comments: _____

Section G: Effort and Workload

Operating the device reduces physical strain. Rating (1–5): __5__ Comments: _____
The device reduces workload during therapy sessions. Rating (1–5): __5__ Comments: _____
The device is not mentally demanding to operate. Rating (1–5): __5__ Comments: _____

Section G: Overall Feedback

What did you like most about the device?
What did you find uncomfortable or difficult?
What improvements would you suggest?

Would you be willing to use this device regularly if recommended by a therapist? (Yes / No)

Remarks: This survey is an adapted hybrid usability survey based on recognised standards and validated constructs, refined using current lower-limb rehabilitation-device literature.

References

[1] J. Brooke, “SUS: A quick and dirty usability scale,” in *Usability Evaluation in Industry*, P. W. Jordan, B. Thomas, B. A. Weerdmeester, and A. L. McClelland, Eds. London, U.K.: Taylor & Francis, 1996, pp. 189–194.

[2] L. Demers, R. Weiss-Lambrou, and B. Ska, “The Quebec User Evaluation of Satisfaction with Assistive Technology (QUEST 2.0): An overview and recent progress,” *Disability and Rehabilitation*, vol. 24, no. 1–3, pp. 21–30, 2002.

[3] S. G. Hart and L. E. Staveland, “Development of NASA-TLX (Task Load Index): Results of empirical and theoretical research,” in *Advances in Psychology*, vol. 52, P. A. Hancock and N. Meshkati, Eds. Amsterdam, Netherlands: North-Holland, 1988, pp. 139–183.

[4] ISO 62366-1, *Medical devices—Application of usability engineering to medical devices—Part 1: Application of usability engineering to medical devices*, Geneva, Switzerland: International Organization for Standardization, 2015.

[5] ISO 62366-1 Amendment 1, *Medical devices—Application of usability engineering to medical devices—Amendment 1*, Geneva, Switzerland: International Organization for Standardization, 2020.

[6] D. R. Louie and J. J. Eng, “Powered robotic exoskeletons in post-stroke rehabilitation of gait: A scoping review,” *Journal of NeuroEngineering and Rehabilitation*, vol. 13, no. 53, 2016.

Device Comfort and Usability Survey: Physiotherapist

This survey aims to evaluate the comfort, usability, and overall user experience of the stroke rehabilitation device. There are no right or wrong answers. Your honest feedback will help improve the design and usability of the device. All responses will be kept confidential.

Participant Information

Name: Shandrico Julian Lim
Age: 19
Gender: Male
Occupation: Student
Date of Test: 3 April 2026
Tester Name: Shandrico Julian Lim
Total Time Using Device: 15 minutes

Section A: General Usability

Please rate the following statements on a scale of 1 (Strongly Disagree) to 5 (Strongly Agree).

The device is easy to operate for patient setup. Rating (1–5): __5__ Comments: _____
The system functions are well integrated. Rating (1–5): __5__ Comments: _____
The controls/interface were intuitive. Rating (1–5): __5__ Comments: _____
I felt confident operating the device. Rating (1–5): __5__ Comments: _____

Section B: Setup, Adjustment and Handling

The device is easy to set up for patients. Rating (1–5): __5__ Comments: _____
Adjustments (e.g. leg length, straps) are efficient and intuitive. Rating (1–5): __5__ Comments: _____
Getting patients into the device is safe and manageable. Rating (1–5): __5__ Comments: _____

Section C: Fit and Physical Comfort

The device provides adequate support and alignment. Rating (1–5): __5__ Comments: _____
The fit can be adjusted to avoid pressure points. Rating (1–5): __5__ Comments: _____

Section D: Exercise Experience and Effectiveness

The device reduces therapist physical workload. Rating (1–5): __5__ Comments: _____

Section E: Safety and Confidence

The device appears safe for patient use. Rating (1–5): __5__ Comments: _____
The device remains stable during operation. Rating (1–5): __5__ Comments: _____
Emergency controls are accessible and reliable. Rating (1–5): __5__ Comments: _____
I trust the device for supervised use. Rating (1–5): __5__ Comments: _____

Section F: Device Design and Practicality

The device is durable and reliable. Rating (1–5): __5__ Comments: _____
The device is easy to disinfect (e.g. wipe down) and maintain. Rating (1–5): __5__ Comments: _____

Section G: Effort and Workload

Operating the device reduces physical strain. Rating (1–5): __5__ Comments: _____
The device reduces workload during therapy sessions. Rating (1–5): __5__ Comments: _____
The device is not mentally demanding to operate. Rating (1–5): __5__ Comments: _____

Section G: Overall Feedback

What did you like most about the device? It is easy to operate and intuitive
What did you find uncomfortable or difficult? -
What improvements would you suggest? would be better if it can move right and left

Would you be willing to use this device regularly if recommended by a therapist? (Yes / No) Yes

Remarks: This survey is an adapted hybrid usability survey based on recognised standards and validated constructs, refined using current lower-limb rehabilitation-device literature.

References

[1] J. Brooke, “SUS: A quick and dirty usability scale,” in *Usability Evaluation in Industry*, P. W. Jordan, B. Thomas, B. A. Weerdmeester, and A. L. McClelland, Eds. London, U.K.: Taylor & Francis, 1996, pp. 189–194.

[2] L. Demers, R. Weiss-Lambrou, and B. Ska, “The Quebec User Evaluation of Satisfaction with Assistive Technology (QUEST 2.0): An overview and recent progress,” *Disability and Rehabilitation*, vol. 24, no. 1–3, pp. 21–30, 2002.

[3] S. G. Hart and L. E. Staveland, “Development of NASA-TLX (Task Load Index): Results of empirical and theoretical research,” in *Advances in Psychology*, vol. 52, P. A. Hancock and N. Meshkati, Eds. Amsterdam, Netherlands: North-Holland, 1988, pp. 139–183.

[4] ISO 62366-1, *Medical devices—Application of usability engineering to medical devices—Part 1: Application of usability engineering to medical devices*, Geneva, Switzerland: International Organization for Standardization, 2015.

[5] ISO 62366-1 Amendment 1, *Medical devices—Application of usability engineering to medical devices—Amendment 1*, Geneva, Switzerland: International Organization for Standardization, 2020.

[6] D. R. Louie and J. J. Eng, “Powered robotic exoskeletons in post-stroke rehabilitation of gait: A scoping review,” *Journal of NeuroEngineering and Rehabilitation*, vol. 13, no. 53, 2016.

Device Comfort and Usability Survey: Physiotherapist

This survey aims to evaluate the comfort, usability, and overall user experience of the stroke rehabilitation device. There are no right or wrong answers. Your honest feedback will help improve the design and usability of the device. All responses will be kept confidential.

Participant Information

Name: Reagan Moses Widjaja
Age: 20
Gender: Male
Occupation: Student
Date of Test: 3 Apr 2026
Tester Name: Nadia and Ron
Total Time Using Device: 15 minutes

Section A: General Usability

Please rate the following statements on a scale of 1 (Strongly Disagree) to 5 (Strongly Agree).

The device is easy to operate for patient setup. Rating (1–5): _5___ Comments: ___Good GUI_____
The system functions are well integrated. Rating (1–5): __5___ Comments: _____
The controls/interface were intuitive. Rating (1–5): ___5__ Comments: _____
I felt confident operating the device. Rating (1–5): __5___ Comments: _____

Section B: Setup, Adjustment and Handling

The device is easy to set up for patients. Rating (1–5): __5___ Comments: _____
Adjustments (e.g. leg length, straps) are efficient and intuitive. Rating (1–5): ___5__ Comments: _____
Getting patients into the device is safe and manageable. Rating (1–5): __4___ Comments: _____

Section C: Fit and Physical Comfort

The device provides adequate support and alignment. Rating (1–5): __5__ Comments: _____
The fit can be adjusted to avoid pressure points. Rating (1–5): __5__ Comments: _____

Section D: Exercise Experience and Effectiveness

The device reduces therapist physical workload. Rating (1–5): __5__ Comments: _____

Section E: Safety and Confidence

The device appears safe for patient use. Rating (1–5): ___5_ Comments: _____
The device remains stable during operation. Rating (1–5): __5__ Comments: _____
Emergency controls are accessible and reliable. Rating (1–5): __5__ Comments: _____
I trust the device for supervised use. Rating (1–5): __5__ Comments: _____

Section F: Device Design and Practicality

The device is durable and reliable. Rating (1–5): __5__ Comments: _____
The device is easy to disinfect (e.g. wipe down) and maintain. Rating (1–5): __5__ Comments: _____

Section G: Effort and Workload

Operating the device reduces physical strain. Rating (1–5): __5__ Comments: _____
The device reduces workload during therapy sessions. Rating (1–5): __5__ Comments: _____
The device is not mentally demanding to operate. Rating (1–5): _5___ Comments: _____

Section G: Overall Feedback

What did you like most about the device? User friendly, good GUI
What did you find uncomfortable or difficult? NIL
What improvements would you suggest? Instructions on GUI itself

Would you be willing to use this device regularly if recommended by a therapist? (Yes / No) Yes

Remarks: This survey is an adapted hybrid usability survey based on recognised standards and validated constructs, refined using current lower-limb rehabilitation-device literature.

References

[1] J. Brooke, “SUS: A quick and dirty usability scale,” in *Usability Evaluation in Industry*, P. W. Jordan, B. Thomas, B. A. Weerdmeester, and A. L. McClelland, Eds. London, U.K.: Taylor & Francis, 1996, pp. 189–194.

[2] L. Demers, R. Weiss-Lambrou, and B. Ska, “The Quebec User Evaluation of Satisfaction with Assistive Technology (QUEST 2.0): An overview and recent progress,” *Disability and Rehabilitation*, vol. 24, no. 1–3, pp. 21–30, 2002.

[3] S. G. Hart and L. E. Staveland, “Development of NASA-TLX (Task Load Index): Results of empirical and theoretical research,” in *Advances in Psychology*, vol. 52, P. A. Hancock and N. Meshkati, Eds. Amsterdam, Netherlands: North-Holland, 1988, pp. 139–183.

[4] ISO 62366-1, *Medical devices—Application of usability engineering to medical devices—Part 1: Application of usability engineering to medical devices*, Geneva, Switzerland: International Organization for Standardization, 2015.

[5] ISO 62366-1 Amendment 1, *Medical devices—Application of usability engineering to medical devices—Amendment 1*, Geneva, Switzerland: International Organization for Standardization, 2020.

[6] D. R. Louie and J. J. Eng, “Powered robotic exoskeletons in post-stroke rehabilitation of gait: A scoping review,” *Journal of NeuroEngineering and Rehabilitation*, vol. 13, no. 53, 2016.

Device Comfort and Usability Survey: Physiotherapist

This survey aims to evaluate the comfort, usability, and overall user experience of the stroke rehabilitation device. There are no right or wrong answers. Your honest feedback will help improve the design and usability of the device. All responses will be kept confidential.

Participant Information

Name: Sebastian Mc Eon Espina Besonia
Age: 25
Gender: M
Occupation: Student
Date of Test: 03/04/ 26
Tester Name: Nadia and Ron
Total Time Using Device: 15 minutes

Section A: General Usability

Please rate the following statements on a scale of 1 (Strongly Disagree) to 5 (Strongly Agree).

The device is easy to operate for patient setup. Rating (1–5): __5__ Comments: _____
The system functions are well integrated. Rating (1–5): __5__ Comments: _____
The controls/interface were intuitive. Rating (1–5): __5__ Comments: _____
I felt confident operating the device. Rating (1–5): __4__ Comments: _____

Section B: Setup, Adjustment and Handling

The device is easy to set up for patients. Rating (1–5): __5__ Comments: _____
Adjustments (e.g. leg length, straps) are efficient and intuitive. Rating (1–5): __5__ Comments: _____
Getting patients into the device is safe and manageable. Rating (1–5): __5__ Comments: _____

Section C: Fit and Physical Comfort

The device provides adequate support and alignment. Rating (1–5): __4__ Comments: _____
The fit can be adjusted to avoid pressure points. Rating (1–5): __4__ Comments: _____

Section D: Exercise Experience and Effectiveness

The device reduces therapist physical workload. Rating (1–5): __5__ Comments: _____

Section E: Safety and Confidence

The device appears safe for patient use. Rating (1–5): __5__ Comments: _____
The device remains stable during operation. Rating (1–5): __3__ Comments: _____
Emergency controls are accessible and reliable. Rating (1–5): __5__ Comments: _____
I trust the device for supervised use. Rating (1–5): __4__ Comments: _____

Section F: Device Design and Practicality

The device is durable and reliable. Rating (1–5): __4__ Comments: _____
The device is easy to disinfect (e.g. wipe down) and maintain. Rating (1–5): __5__ Comments: _____

Section G: Effort and Workload

Operating the device reduces physical strain. Rating (1–5): __5__ Comments: _____
The device reduces workload during therapy sessions. Rating (1–5): __5__ Comments: _____
The device is not mentally demanding to operate. Rating (1–5): __5__ Comments: _____

Section G: Overall Feedback

What did you like most about the device? I like how easy it is to customize the treatment plan.
What did you find uncomfortable or difficult? I think the structure can be more secure.
What improvements would you suggest? The structure might need more than 2 mounting points

Would you be willing to use this device regularly if recommended by a therapist? (Yes / No) Yes

Remarks: This survey is an adapted hybrid usability survey based on recognised standards and validated constructs, refined using current lower-limb rehabilitation-device literature.

References

[1] J. Brooke, “SUS: A quick and dirty usability scale,” in *Usability Evaluation in Industry*, P. W. Jordan, B. Thomas, B. A. Weerdmeester, and A. L. McClelland, Eds. London, U.K.: Taylor & Francis, 1996, pp. 189–194.

[2] L. Demers, R. Weiss-Lambrou, and B. Ska, “The Quebec User Evaluation of Satisfaction with Assistive Technology (QUEST 2.0): An overview and recent progress,” *Disability and Rehabilitation*, vol. 24, no. 1–3, pp. 21–30, 2002.

[3] S. G. Hart and L. E. Staveland, “Development of NASA-TLX (Task Load Index): Results of empirical and theoretical research,” in *Advances in Psychology*, vol. 52, P. A. Hancock and N. Meshkati, Eds. Amsterdam, Netherlands: North-Holland, 1988, pp. 139–183.

[4] ISO 62366-1, *Medical devices—Application of usability engineering to medical devices—Part 1: Application of usability engineering to medical devices*, Geneva, Switzerland: International Organization for Standardization, 2015.

[5] ISO 62366-1 Amendment 1, *Medical devices—Application of usability engineering to medical devices—Amendment 1*, Geneva, Switzerland: International Organization for Standardization, 2020.

[6] D. R. Louie and J. J. Eng, “Powered robotic exoskeletons in post-stroke rehabilitation of gait: A scoping review,” *Journal of NeuroEngineering and Rehabilitation*, vol. 13, no. 53, 2016.

Device Comfort and Usability Survey: Physiotherapist

This survey aims to evaluate the comfort, usability, and overall user experience of the stroke rehabilitation device. There are no right or wrong answers. Your honest feedback will help improve the design and usability of the device. All responses will be kept confidential.

Participant Information

Name: Ahmad Rizqi Hidayat
Age: 21
Gender: Male
Occupation: Student
Date of Test: 3rd April 2026
Tester Name: Nadia
Total Time Using Device: 15 minutes

Section A: General Usability

Please rate the following statements on a scale of 1 (Strongly Disagree) to 5 (Strongly Agree).

The device is easy to operate for patient setup. Rating (1–5): __4__ Comments: _____
The system functions are well integrated. Rating (1–5): __4__ Comments: _____
The controls/interface were intuitive. Rating (1–5): __5__ Comments: _____
I felt confident operating the device. Rating (1–5): __5__ Comments: _____

Section B: Setup, Adjustment and Handling

The device is easy to set up for patients. Rating (1–5): __4__ Comments: _____
Adjustments (e.g. leg length, straps) are efficient and intuitive. Rating (1–5): __4__ Comments: _____
Getting patients into the device is safe and manageable. Rating (1–5): __4__ Comments: _____

Section C: Fit and Physical Comfort

The device provides adequate support and alignment. Rating (1–5): __4__ Comments: _____
The fit can be adjusted to avoid pressure points. Rating (1–5): __4__ Comments: _____

Section D: Exercise Experience and Effectiveness

The device reduces therapist physical workload. Rating (1–5): __4__ Comments: _____

Section E: Safety and Confidence

The device appears safe for patient use. Rating (1–5): __5__ Comments: _____
The device remains stable during operation. Rating (1–5): __4__ Comments: _____
Emergency controls are accessible and reliable. Rating (1–5): __5__ Comments: _____
I trust the device for supervised use. Rating (1–5): __4__ Comments: _____

Section F: Device Design and Practicality

The device is durable and reliable. Rating (1–5): __4__ Comments: _____
The device is easy to disinfect (e.g. wipe down) and maintain. Rating (1–5): __5__ Comments: _____

Section G: Effort and Workload

Operating the device reduces physical strain. Rating (1–5): __5__ Comments: _____
The device reduces workload during therapy sessions. Rating (1–5): __4__ Comments: _____
The device is not mentally demanding to operate. Rating (1–5): __4__ Comments: _____

Section G: Overall Feedback

What did you like most about the device? The device is relatively easy to set up
What did you find uncomfortable or difficult? The structure is slightly shaky
What improvements would you suggest? Adding more pivot points to the sturcture

Would you be willing to use this device regularly if recommended by a therapist? (Yes / No) Yes

Remarks: This survey is an adapted hybrid usability survey based on recognised standards and validated constructs, refined using current lower-limb rehabilitation-device literature.

References

[1] J. Brooke, “SUS: A quick and dirty usability scale,” in *Usability Evaluation in Industry*, P. W. Jordan, B. Thomas, B. A. Weerdmeester, and A. L. McClelland, Eds. London, U.K.: Taylor & Francis, 1996, pp. 189–194.

[2] L. Demers, R. Weiss-Lambrou, and B. Ska, “The Quebec User Evaluation of Satisfaction with Assistive Technology (QUEST 2.0): An overview and recent progress,” *Disability and Rehabilitation*, vol. 24, no. 1–3, pp. 21–30, 2002.

[3] S. G. Hart and L. E. Staveland, “Development of NASA-TLX (Task Load Index): Results of empirical and theoretical research,” in *Advances in Psychology*, vol. 52, P. A. Hancock and N. Meshkati, Eds. Amsterdam, Netherlands: North-Holland, 1988, pp. 139–183.

[4] ISO 62366-1, *Medical devices—Application of usability engineering to medical devices—Part 1: Application of usability engineering to medical devices*, Geneva, Switzerland: International Organization for Standardization, 2015.

[5] ISO 62366-1 Amendment 1, *Medical devices—Application of usability engineering to medical devices—Amendment 1*, Geneva, Switzerland: International Organization for Standardization, 2020.

[6] D. R. Louie and J. J. Eng, “Powered robotic exoskeletons in post-stroke rehabilitation of gait: A scoping review,” *Journal of NeuroEngineering and Rehabilitation*, vol. 13, no. 53, 2016.

Device Comfort and Usability Survey: Physiotherapist

This survey aims to evaluate the comfort, usability, and overall user experience of the stroke rehabilitation device. There are no right or wrong answers. Your honest feedback will help improve the design and usability of the device. All responses will be kept confidential.

Participant Information

Name: Ron
Age: 25
Gender: Male
Occupation: Student
Date of Test: 3 Apr 2026
Tester Name: Nadia
Total Time Using Device: 15 minutes

Section A: General Usability

Please rate the following statements on a scale of 1 (Strongly Disagree) to 5 (Strongly Agree).

The device is easy to operate for patient setup. Rating (1–5): __4___ Comments: _____
The system functions are well integrated. Rating (1–5): __4___ Comments: _____
The controls/interface were intuitive. Rating (1–5): __4___ Comments: _____
I felt confident operating the device. Rating (1–5): __5___ Comments: _____

Section B: Setup, Adjustment and Handling

The device is easy to set up for patients. Rating (1–5): __5___ Comments: _____
Adjustments (e.g. leg length, straps) are efficient and intuitive. Rating (1–5): __4___ Comments: _____
Getting patients into the device is safe and manageable. Rating (1–5): __4___ Comments: _____

Section C: Fit and Physical Comfort

The device provides adequate support and alignment. Rating (1–5): __3___ Comments: _____
The fit can be adjusted to avoid pressure points. Rating (1–5): __4___ Comments: _____

Section D: Exercise Experience and Effectiveness

The device reduces therapist physical workload. Rating (1–5): __5___ Comments: _____

Section E: Safety and Confidence

The device appears safe for patient use. Rating (1–5): __4___ Comments: _____
The device remains stable during operation. Rating (1–5): __3___ Comments: _____
Emergency controls are accessible and reliable. Rating (1–5): __5___ Comments: _____
I trust the device for supervised use. Rating (1–5): __5___ Comments: _____

Section F: Device Design and Practicality

The device is durable and reliable. Rating (1–5): __4___ Comments: _____
The device is easy to disinfect (e.g. wipe down) and maintain. Rating (1–5): __5___ Comments: _____

Section G: Effort and Workload

Operating the device reduces physical strain. Rating (1–5): __5___ Comments: _____
The device reduces workload during therapy sessions. Rating (1–5): _5___ Comments: _____
The device is not mentally demanding to operate. Rating (1–5): __5___ Comments: _____

Section G: Overall Feedback

What did you like most about the device? It was easy and quick to set up
What did you find uncomfortable or difficult? NA
What improvements would you suggest? Having more exercises

Would you be willing to use this device regularly if recommended by a therapist? (Yes / No) Yes

Remarks: This survey is an adapted hybrid usability survey based on recognised standards and validated constructs, refined using current lower-limb rehabilitation-device literature.

References

[1] J. Brooke, “SUS: A quick and dirty usability scale,” in *Usability Evaluation in Industry*, P. W. Jordan, B. Thomas, B. A. Weerdmeester, and A. L. McClelland, Eds. London, U.K.: Taylor & Francis, 1996, pp. 189–194.

[2] L. Demers, R. Weiss-Lambrou, and B. Ska, “The Quebec User Evaluation of Satisfaction with Assistive Technology (QUEST 2.0): An overview and recent progress,” *Disability and Rehabilitation*, vol. 24, no. 1–3, pp. 21–30, 2002.

[3] S. G. Hart and L. E. Staveland, “Development of NASA-TLX (Task Load Index): Results of empirical and theoretical research,” in *Advances in Psychology*, vol. 52, P. A. Hancock and N. Meshkati, Eds. Amsterdam, Netherlands: North-Holland, 1988, pp. 139–183.

[4] ISO 62366-1, *Medical devices—Application of usability engineering to medical devices—Part 1: Application of usability engineering to medical devices*, Geneva, Switzerland: International Organization for Standardization, 2015.

[5] ISO 62366-1 Amendment 1, *Medical devices—Application of usability engineering to medical devices—Amendment 1*, Geneva, Switzerland: International Organization for Standardization, 2020.

[6] D. R. Louie and J. J. Eng, “Powered robotic exoskeletons in post-stroke rehabilitation of gait: A scoping review,” *Journal of NeuroEngineering and Rehabilitation*, vol. 13, no. 53, 2016.

Device Comfort and Usability Survey: Physiotherapist

This survey aims to evaluate the comfort, usability, and overall user experience of the stroke rehabilitation device. There are no right or wrong answers. Your honest feedback will help improve the design and usability of the device. All responses will be kept confidential.

Participant Information

Name: Nadia Rahman
Age: 22
Gender: Female
Occupation: Student
Date of Test: 3 April
Tester Name: Ron and Nadia
Total Time Using Device: 15 Min

Section A: General Usability

Please rate the following statements on a scale of 1 (Strongly Disagree) to 5 (Strongly Agree).

The device is easy to operate for patient setup. Rating (1–5): _5___ Comments: _____
The system functions are well integrated. Rating (1–5): __4___ Comments: _____
The controls/interface were intuitive. Rating (1–5): __4___ Comments: _____
I felt confident operating the device. Rating (1–5): __4___ Comments: _____

Section B: Setup, Adjustment and Handling

The device is easy to set up for patients. Rating (1–5): ___5__ Comments: _____
Adjustments (e.g. leg length, straps) are efficient and intuitive. Rating (1–5): __4___ Comments: _____
Getting patients into the device is safe and manageable. Rating (1–5): __4___ Comments: _____

Section C: Fit and Physical Comfort

The device provides adequate support and alignment. Rating (1–5): __5__ Comments: _____
The fit can be adjusted to avoid pressure points. Rating (1–5): __4__ Comments: _____

Section D: Exercise Experience and Effectiveness

The device reduces therapist physical workload. Rating (1–5): __4__ Comments: _____

Section E: Safety and Confidence

The device appears safe for patient use. Rating (1–5): __5__ Comments: _____
The device remains stable during operation. Rating (1–5): __4__ Comments: _____
Emergency controls are accessible and reliable. Rating (1–5): __5__ Comments: _____
I trust the device for supervised use. Rating (1–5): __5__ Comments: _____

Section F: Device Design and Practicality

The device is durable and reliable. Rating (1–5): __5__ Comments: _____
The device is easy to disinfect (e.g. wipe down) and maintain. Rating (1–5): __4__ Comments: _____

Section G: Effort and Workload

Operating the device reduces physical strain. Rating (1–5): __5__ Comments: _____
The device reduces workload during therapy sessions. Rating (1–5): __5__ Comments: _____
The device is not mentally demanding to operate. Rating (1–5): __5__ Comments: _____

Section G: Overall Feedback

What did you like most about the device? I just have to let it run and monitor the patients.
What did you find uncomfortable or difficult? Getting the patient's leg into the device can be a bit challenging
What improvements would you suggest? The UI can have more auditory cues.

Would you be willing to use this device regularly if recommended by a therapist? (Yes / No) Yes

Remarks: This survey is an adapted hybrid usability survey based on recognised standards and validated constructs, refined using current lower-limb rehabilitation-device literature.

References

[1] J. Brooke, “SUS: A quick and dirty usability scale,” in *Usability Evaluation in Industry*, P. W. Jordan, B. Thomas, B. A. Weerdmeester, and A. L. McClelland, Eds. London, U.K.: Taylor & Francis, 1996, pp. 189–194.

[2] L. Demers, R. Weiss-Lambrou, and B. Ska, “The Quebec User Evaluation of Satisfaction with Assistive Technology (QUEST 2.0): An overview and recent progress,” *Disability and Rehabilitation*, vol. 24, no. 1–3, pp. 21–30, 2002.

[3] S. G. Hart and L. E. Staveland, “Development of NASA-TLX (Task Load Index): Results of empirical and theoretical research,” in *Advances in Psychology*, vol. 52, P. A. Hancock and N. Meshkati, Eds. Amsterdam, Netherlands: North-Holland, 1988, pp. 139–183.

[4] ISO 62366-1, *Medical devices—Application of usability engineering to medical devices—Part 1: Application of usability engineering to medical devices*, Geneva, Switzerland: International Organization for Standardization, 2015.

[5] ISO 62366-1 Amendment 1, *Medical devices—Application of usability engineering to medical devices—Amendment 1*, Geneva, Switzerland: International Organization for Standardization, 2020.

[6] D. R. Louie and J. J. Eng, “Powered robotic exoskeletons in post-stroke rehabilitation of gait: A scoping review,” *Journal of NeuroEngineering and Rehabilitation*, vol. 13, no. 53, 2016.