

Evaluation / Calibration Record

Date: September 26, 2022
Sensor Type: Seapoint Turbidity Meter
Serial Number(s): 10961

A calibration was performed using formazin turbidity standard.

As received, the sensitivity of this Seapoint Turbidity Meter was 14% lower than specification.

Adjustments were made; it now passes or exceeds the specifications stated in the supplied User Manual.

If you have any questions please contact me at 603/642-4921 or seapoint@seapoint.com.

T. O'Brien

09/26/22

Signature

Date



PO Box 368 • Exeter, NH 03833 • USA
Tel: (603) 642-4921 • Fax: (603) 642-4922
seapoint@seapoint.com • www.seapoint.com