

TEMPEST: Porewater Sulfide

2025 Samples Plate 12

2025-10-24

Run Information

```
##### Run information - PLEASE CHANGE
Date_Run = "20250926" #Date that instrument was run
Plate = "12"
Year = "2025"
Run_by = "Zoe Read" #Instrument user
Script_run_by = "Zoe Read" #Code user
Project = "TEMPEST"
run_notes = "Std Curve r2 is below cutoff: 0.978154
Removed Std 5.
Samples 109 and 133 still have high CVs after removing one duplicate.
Dup & respective sample had H2S values of 0.
" #any notes from the run

##### STDs to remove manually - NA unless standards have high CVs
stds_to_remove <- c("B03", "C03", "E01", "F01", "F02", "F03") #Ex. c("A02", "B02")

Chkstds_to_remove <- c(NA)

MCs_to_remove <- c(NA)

##Sample data that was entered incorrectly
# The Old ID is the original, incorrectly-entered ID and the New ID is the correct ID to change it to.
# Old_ID_1 = "17 Dup"
# New_ID_1 = "16 Dup"
cat(run_notes)

## Std Curve r2 is below cutoff: 0.978154
## Removed Std 5.
## Samples 109 and 133 still have high CVs after removing one duplicate.
## Dup & respective sample had H2S values of 0.
##

## [1] "Std Curve R squared: 0.9782"
```





