

TEMPEST: Porewater SO₄/Cl

202606 1-23 Samples

2026-09-02

Contents

0.1	Run Information	2
0.2	Assess Standard Curves	3
0.3	Assess Check Standards	4
0.4	Assess Blanks	5
0.5	Assess Duplicates	7
0.6	Calculate mmol/L concentrations & salinity, add dilutions	8
0.7	Assess Analytical Spikes	9
0.8	Check if samples within the range of the standard curve	9
0.9	Visualize Data by Plot	11
0.10	Export Processed Data	11

##Add Required Packages

0.1 Run Information

```
##### Run information - PLEASE CHANGE
Sample_Year = "2026"
Date_Run = "2026-08-26" #Date that instrument was run
Run_by = "Zoe Read" #Instrument user
Script_run_by = "Zoe Read" #Code user
run_notes = "
" #any notes from the run

##Sample data that was entered incorrectly
# The Old ID is the original, incorrectly-entered ID and the New ID is the correct ID to change it to.
Old_ID_1 = NA
New_ID_1 = NA

##### File Names - PLEASE CHANGE
#file path and name for raw summary data file
raw_file_name_cl = "Raw Data/COMPASS_TEMPEST_202606_1-23_C1.txt"
raw_file_name_so4 = "Raw Data/COMPASS_TEMPEST_202606_1-23_S04.txt"

#Sample ID Path
SampleID_path = "Raw Data/TEMPEST_S04_20260501-20260612_corrected.csv"

#file path and name of processed data file
processed_file_name = "Processed Data/COMPASS_TEMPEST_Processed_C1_S04_202606_1-23.csv"


##### Log Files - PLEASE CHECK

#qaqc log file path for this year copied over from COMPASS GitHub
Log_path = "Raw Data/COMPASS_Synoptic_C1_S04_QAQClog_2026.csv"


##Set Up Code - constants and QAQC cutoffs
##Import Sample Data
##Fix Sample IDs entered wrong
```

0.2 Assess Standard Curves


Chloride Std Curve



Sulfate Std Curve



```
## [1] "QAQC log file exists and has been read into the code."
```



```
## [1] "Cl Curve r2 GOOD"
```

```
## [1] "SO4 Curve r2 GOOD"
```


0.3 Assess Check Standards

```
## # A tibble: 5 x 5
##   sample_ID mean_Cl    sd_Cl    cv_Cl flag_Cl
##   <chr>      <dbl>    <dbl>    <dbl> <chr>
## 1 Standard 1  4.96  0.000212  0.0000428 Chloride Check Standard RSD within Ra-
## 2 Standard 2  9.99  0.00884   0.000885 Chloride Check Standard RSD within Ra-
## 3 Standard 3  19.6 NA      NA      <NA>
## 4 Standard 4  99.8 NA      NA      <NA>
## 5 Standard 5  200. NA      NA      <NA>
```

```
## # A tibble: 5 x 5
##   sample_ID mean_SO4    sd_SO4    cv_SO4 flag_SO4
##   <chr>      <dbl>    <dbl>    <dbl> <chr>
## 1 Standard 1  0.522  0.00354  0.00677 Sulfate Check Standard RSD within Range-
## 2 Standard 2  0.996  0.00460  0.00462 Sulfate Check Standard RSD within Range-
## 3 Standard 3  1.90 NA      NA      <NA>
## 4 Standard 4  9.93 NA      NA      <NA>
## 5 Standard 5  20.0 NA      NA      <NA>
```

```
## [1] ">80% of Chloride Check Standards have RSD within range - PROCEED"
```

```
## [1] ">80% of Sulfate Check Standards have RSD within range - PROCEED"
```




```
## [1] ">80% of Chloride Check Standards are within range of expected concentration - PROCEED"
```

```
## [1] ">80% of Sulfate Check Standards are within range of expected concentration - PROCEED"
```

0.4 Assess Blanks

```
## [1] ">80% of Chloride Blank concentrations are lower 25% quartile of samples"
```

```
## [1] ">80% of Sulfate Blank concentrations are lower 25% quartile of samples"
```




```
## Chloride blanks mean ppm:
```

```
## [1] 0.002225
```

```
## Sulfate blanks mean ppm:
```


```
## [1] 0.0001875
```

0.5 Assess Duplicates



```
## [1] ">80% of Chloride Duplicates have a CV <10% - PROCEED"
```

```
## [1] ">80% of Sulfate Duplicates have a CV <10% - PROCEED"
```



```
## [1] ">80% of Chloride Duplicates have a percent difference <10% - PROCEED"
```

```
## [1] ">80% of Sulfate Duplicates have a percent difference <10% - PROCEED"
```

0.6 Calculate mmol/L concentrations & salinity, add dilutions


```
# Convert ppm to mmol/L
all_dat$S04_Conc_mM <- (all_dat$S04_ppm / s_mw)
all_dat$Cl_Conc_mM <- (all_dat$Cl_ppm / cl_mw)

# Calculate Salinity
# calculated using the Knudsen equation
# Salinity = 0.03 + 1.8050 * Chlorinity
# Ref: A Practical Handbook of Seawater Analysis by Strickland & Parsons (P. 11)
# =((1.807*Cl_ppm)+0.026)/1000
all_dat$salinity <- ((1.8070 * all_dat$Cl_ppm) + 0.026) / 1000

#Need to determine dilution factors for your samples
#for TEMPEST this depends on the sample so...
all_dat$Dilution <- 50
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "TMP"), 50, all_dat$Dilution)
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "EST_SourceWater"), 100, all_dat$Dilution)
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "SW_SourceWater"), 100, all_dat$Dilution)
```

```
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "FATE"), 100, all_dat$Dilution)
# head(all_dat)
```

0.7 Assess Analytical Spikes



```
## [1] ">80% of S04 spikes have a recovery between the high and low cutoff - PROCEED"
```

0.8 Check if samples within the range of the standard curve

```
## Sample Flagging
```




Table 1: SO4 samples

SO4_Conc_flag	Percent_samples
Within_Range	100


Table 2: Cl samples

Cl_Conc_flag	Percent_samples
Within_Range	100

```
##Merge with Sample ID list to get actual sample IDs
##Pull in active porewater inventory
##Check samples against metadata

## ***All sample IDs are present in metadata.***
```

0.9 Visualize Data by Plot



0.10 Export Processed Data

#end