```
Thermo Scientific SII for Xcalibur Method

---- Overview ----
Name: New Instrument Method
Comment:
Run time: 17.000 [min]
Instrument: Vanquish on desktop-vcjjm2l
Description:
---- Script ----
initial     Instrument Setup
            ColumnComp.PrehtLeft.ReadyTempDelta: 1.00 [°C]
            ColumnComp.PrehtLeft.TempCtrl: On
            ColumnComp.PrehtLeft.Temperature.Nominal: 40.00 [°C]
            ColumnComp.PrehtLeft.EquilibrationTime: 1.0 [min]
            ColumnComp.CC.Mode: StillAir
            ColumnComp.CC.ReadyTempDelta: 0.50 [°C]
            ColumnComp.CC.TempCtrl: On
            ColumnComp.CC.Temperature.Nominal: 40.00 [°C]
            ColumnComp.CC.EquilibrationTime: 1.0 [min]
            SamplerModule.Sampler.SyncWithPump: On
            SamplerModule.Sampler.DrawSpeed: 1.600 [µl/s]
            SamplerModule.Sampler.DispenseSpeed: 5.000 [µl/s]
            SamplerModule.Sampler.PunctureOffset: 0 [µm]
            SamplerModule.Sampler.WashTime: 5.0 [s]
            SamplerModule.Sampler.WashSpeed: 50.0 [µl/s]
            SamplerModule.Sampler.InjectWashMode: Both
            SamplerModule.Sampler.Pump: "Pump"
            SamplerModule.Sampler.InjectMode: Normal
            SamplerModule.TempCtrl: On
            SamplerModule.Temperature.Nominal: 4.0 [°C]
            PumpModule.Pump.%A.Equate: "%A "
            PumpModule.Pump.%B.Equate: "%B acetonitrile and formic 0.1%"
            PumpModule.Pump.%C.Equate: "%C water and formic 0.1%"
            PumpModule.Pump.%D.Equate: "%D "
            PumpModule.Pump.Pressure.LowerLimit: 0 [bar]
            PumpModule.Pump.Pressure.UpperLimit: 600 [bar]
            PumpModule.Pump.MaximumFlowRampUp: 1.00 [ml/min²]
            PumpModule.Pump.MaximumFlowRampDown: 1.00 [ml/min²]
0.000 [min] Inject Preparation
            Wait ColumnComp.Ready And SamplerModule.Sampler.Ready And PumpModule.Pump.Ready
0.000 [min] Inject
            SamplerModule.Sampler.Inject
0.000 [min] Start Run
            PumpModule.Pump.Pump_Pressure.AcqOn
0.000 [min] Run
            PumpModule.Pump.Flow.Nominal: 0.400 [ml/min]
            PumpModule.Pump.%B.Value: 0.0 [%]
            PumpModule.Pump.%C.Value: 100.0 [%]
            PumpModule.Pump.%D.Value: 0.0 [%]
            PumpModule.Pump.Curve: 5
```

```
Thermo Scientific SII for Xcalibur Method

1.000 [min]
            PumpModule.Pump.Flow.Nominal: 0.400 [ml/min]
            PumpModule.Pump.%B.Value: 0.0 [%]
            PumpModule.Pump.%C.Value: 100.0 [%]
            PumpModule.Pump.%D.Value: 0.0 [%]
            PumpModule.Pump.Curve: 5
9.000 [min]
            PumpModule.Pump.Flow.Nominal: 0.400 [ml/min]
            PumpModule.Pump.%B.Value: 98.0 [%]
            PumpModule.Pump.%C.Value: 2.0 [%]
            PumpModule.Pump.%D.Value: 0.0 [%]
            PumpModule.Pump.Curve: 5
11.000 [min]
            PumpModule.Pump.Flow.Nominal: 0.400 [ml/min]
            PumpModule.Pump.%B.Value: 98.0 [%]
            PumpModule.Pump.%C.Value: 2.0 [%]
            PumpModule.Pump.%D.Value: 0.0 [%]
            PumpModule.Pump.Curve: 5
12.000 [min]
            PumpModule.Pump.Flow.Nominal: 0.400 [ml/min]
            PumpModule.Pump.%B.Value: 98.0 [%]
            PumpModule.Pump.%C.Value: 2.0 [%]
            PumpModule.Pump.%D.Value: 0.0 [%]
            PumpModule.Pump.Curve: 5
            PumpModule.Pump.Flow.Nominal: 0.400 [ml/min]
            PumpModule.Pump.%B.Value: 0.0 [%]
            PumpModule.Pump.%C.Value: 100.0 [%]
            PumpModule.Pump.%D.Value: 0.0 [%]
            PumpModule.Pump.Curve: 5
17.000 [min]
            PumpModule.Pump.Flow.Nominal: 0.400 [ml/min]
            PumpModule.Pump.%B.Value: 0.0 [%]
            PumpModule.Pump.%C.Value: 100.0 [%]
            PumpModule.Pump.%D.Value: 0.0 [%]
            PumpModule.Pump.Curve: 5
17.000 [min] Stop Run
            PumpModule.Pump.Pump_Pressure.AcqOff
```

Method Summary

Method Settings

Application Mode: **Small Molecule**
Method Duration (min): **15**

Global Parameters

Ion Source

Ion Source Type: **H-ESI**
Spray Voltage: **Static**
Positive Ion (V): **3500**
Negative Ion (V): **2000**
Gas Mode: **Static**
Sheath Gas (Arb): **50**
Aux Gas (Arb): **10**
Sweep Gas (Arb): **1**
Ion Transfer Tube Temp (°C): **325**
Vaporizer Temp (°C): **350**
APPI Lamp: **Not in Use**
Use Ion Source Settings from Tune: **False**
FAIMS Mode: **Not Installed**

MS Global Settings

Infusion Mode: **Liquid Chromatography**
Expected LC Peak Width (s): **6**
Advanced Peak Determination: **True**
Mild Trapping: **False**
Default Charge State: **1**
Enable Xcalibur AcquireX method modifications: **True**
Internal Mass Calibration: **EASY-IC™**
Mode: **Scan-to-Scan**

Divert Valve A

Time (min)	Position
0	1-2

0.5	1-6
15	1-2

Experiment #1 [MS]

Start Time (min): **0**
End Time (min): **15**

Master Scan:

Full Scan

Orbitrap Resolution: **120000**
Scan Range (m/z): **60-900**
RF Lens (%): **60**
AGC Target: **Standard**
Maximum Injection Time Mode: **Auto**
Microscans: **1**
Data Type: **Profile**
Polarity: **Positive**
Source Fragmentation: **Disabled**
Use EASY-IC™: **On**
Scan Description:

Filters:

Dynamic Exclusion

Dynamic Exclusion Mode: **Auto**

Targeted Mass Exclusion

Mass List

Mass List type: **m/z**
Time Mode: **Start/End Time**
Include Intensity Threshold: **True**
Add Mass List Targets Determined by Xcalibur AcquireX: **True**

Compound	m/z	t start (min)	t stop (min)	Intensity Threshold
	524.2649	0	15	1E+20

Exclusion mass width: **ppm**

Low: **5**
High: **5**

Apex Detection

Desired Apex Window (%): **30**

Intensity

Intensity Threshold: **5.0e4**

Data Dependent

Data Dependent Mode: **Cycle Time**
Time between Master Scans (sec): **0.8**

Scan Event Type 1:

Scan:

$ddMS^2$

Multiplex Ions: **False**
Isolation Window (m/z): **1.5**
Isolation Offset: **Off**
Collision Energy Type: **Normalized**
HCD Collision Energy (%): **20,30,40**
Orbitrap Resolution: **30000**
Scan Range Mode: **Auto**
AGC Target: **Standard**
Maximum Injection Time Mode: **Auto**
Microscans: **1**
Data Type: **Centroid**
Use EASY-IC™: **Off**
Scan Description: