

Synoptic CB: Porewater SO₄/Cl

May 2024 Samples

2025-09-11

Contents

0.1	Run Information	2
0.2	Assess Standard Curves	3
0.3	Assess Check Standards	4
0.4	Assess Blanks	5
0.5	Assess Duplicates	7
0.6	Calculate mmol/L concentrations & salinity, add dilutions	8
0.7	Assess Analytical Spikes	9
0.8	Check if samples within the range of the standard curve	9
0.9	Check to see if samples run match metadata & merge info	10
0.10	Visualize Data by Plot	11
0.11	Export Processed Data	11

```
##Add Required Packages
```

0.1 Run Information

```
##### Run information - PLEASE CHANGE
Date_Run = "2024-06-04" #Date that instrument was run
Run_by = "Zoe Read" #Instrument user
Script_run_by = "Zoe Read" #Code user
run_notes = "Some sample IDs are missing from metadata:
GWI_202405_UP_LYSC_45CM - this has values of 0 for S04 and Cl so maybe there was no sample.
MSM_202405_UP_LYSA_10CM" #any notes from the run
samples <- c("GCW", "GWI", "MSM", "SWH") #whatever identifies your samples within the same names
samples_pattern <- paste(samples, collapse = "|")
#samples_pattern <- "GCW" #use this instead of the line above if you have only one site code

##### File Names - PLEASE CHANGE
#file path and name for raw summary data file
raw_file_name_cl = "Raw Data/COMPASS_Synoptic_CB_MonMon_202405_Cl.txt"
raw_file_name_so4 = "Raw Data/COMPASS_Synoptic_CB_MonMon_202405_S04.txt"

#file path and name of processed data file
processed_file_name = "Processed Data/COMPASS_SynopticCB_PW_Processed_Cl_S04_202405.csv"

##### Log Files - PLEASE CHECK
#downloaded metadata csv - downloaded from Google drive as csv for this year
Raw_Metadata = "Raw Data/COMPASS_SynopticCB_PW_SampleLog_2024.csv"

#qaqc log file path for this year
Log_path = "Raw Data/COMPASS_Synoptic_Cl_S04_QAQClog_2024.csv"
```


```
##Set Up Code - constants and QAQC cutoffs
```

```
##Read in metadata and create similar sample IDs for matching to samples
```


```
##Import Sample Data
```

0.2 Assess Standard Curves


Chloride Std Curve



Sulfate Std Curve



```
## [1] "QAQC log file exists and has been read into the code."
```



```
## [1] "Cl Curve r2 GOOD"
```


```
## [1] "SO4 Curve r2 GOOD"
```

0.3 Assess Check Standards

```
## # A tibble: 5 x 5
##   sample_ID  mean_Cl    sd_Cl    cv_Cl flag_Cl
##   <chr>       <dbl>     <dbl>     <dbl> <chr>
## 1 Standard 1  5.13  0.0830  0.0162 Chloride Check Standard RSD within Range - ~
## 2 Standard 2  10.4   0.122   0.0118 Chloride Check Standard RSD within Range - ~
## 3 Standard 3  20.6   0.180   0.00875 Chloride Check Standard RSD within Range - ~
## 4 Standard 4  102.    0.993   0.00972 Chloride Check Standard RSD within Range - ~
## 5 Standard 5  203.    2.11    0.0104  Chloride Check Standard RSD within Range - ~

## # A tibble: 5 x 5
##   sample_ID  mean_SO4   sd_SO4   cv_SO4 flag_SO4
##   <chr>       <dbl>     <dbl>     <dbl> <chr>
## 1 Standard 1  0.499  0.0143  0.0287 Sulfate Check Standard RSD within Range - ~
## 2 Standard 2  1.02   0.0156  0.0154 Sulfate Check Standard RSD within Range - ~
## 3 Standard 3  2.04   0.0298  0.0146 Sulfate Check Standard RSD within Range - ~
## 4 Standard 4  10.3   0.234   0.0228 Sulfate Check Standard RSD within Range - ~
## 5 Standard 5  20.2   0.188   0.00928 Sulfate Check Standard RSD within Range - ~


## [1] ">80% of Chloride Check Standards have RSD within range - PROCEED"
## [1] ">80% of Sulfate Check Standards have RSD within range - PROCEED"
```



```
## [1] ">80% of Chloride Check Standards are within range of expected concentration - PROCEED"
## [1] ">80% of Sulfate Check Standards are within range of expected concentration - PROCEED"
```

0.4 Assess Blanks

```
## [1] ">80% of Chloride Blank concentrations are lower 25% quartile of samples"
## [1] ">80% of Sulfate Blank concentrations are lower 25% quartile of samples"
```




```
## Chloride blanks mean ppm:
```

```
## [1] 0.0208
```

```
## Sulfate blanks mean ppm:
```


```
## [1] 0.005310526
```

0.5 Assess Duplicates



```
## [1] ">80% of Chloride Duplicates have a CV <10% - PROCEED"
```

```
## [1] ">80% of Sulfate Duplicates have a CV <10% - PROCEED"
```



```
## [1] ">80% of Chloride Duplicates have a percent difference <10% - PROCEED"
```

```
## [1] ">80% of Sulfate Duplicates have a percent difference <10% - PROCEED"
```

0.6 Calculate mmol/L concentrations & salinity, add dilutions

```
# Convert ppm to mmol/L
all_dat$SO4_Conc_mM <- (all_dat$SO4_ppm / s_mw)
all_dat$Cl_Conc_mM <- (all_dat$Cl_ppm / cl_mw)

# Calculate Salinity
# calculated using the Knudsen equation
# Salinity = 0.03 + 1.8050 * Chlorinity
# Ref: A Practical Handbook of Seawater Analysis by Strickland & Parsons (P. 11)
# =((1.807*Cl_ppm)+0.026)/1000
all_dat$salinity <- ((1.8070 * all_dat$Cl_ppm) + 0.026) / 1000

#Need to determine dilution factors for your samples
#for Steph / COMPASS this depends on the site so...
all_dat$Dilution <- 1
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "MSM") & str_detect(all_dat$sample_ID, "UP"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "MSM") & str_detect(all_dat$sample_ID, "TR"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "MSM") & str_detect(all_dat$sample_ID, "WC"),
```


```

all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "MSM") & str_detect(all_dat$sample_ID, "SW"), 50,
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GCW") & str_detect(all_dat$sample_ID, "UP"), 100,
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GCW") & str_detect(all_dat$sample_ID, "TR"), 100,
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GCW") & str_detect(all_dat$sample_ID, "WC"), 100,
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GCW") & str_detect(all_dat$sample_ID, "SW"), 100,
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GWI") & str_detect(all_dat$sample_ID, "UP"), 100,
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GWI") & str_detect(all_dat$sample_ID, "TR"), 100,
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GWI") & str_detect(all_dat$sample_ID, "WC"), 100,
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GWI") & str_detect(all_dat$sample_ID, "SW"), 100,
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "SWH"), 50, all_dat$Dilution)

# head(all_dat)

```

0.7 Assess Analytical Spikes



```
## [1] ">80% of S04 spikes have a recovery between the high and low cutoff - PROCEED"
```

0.8 Check if samples within the range of the standard curve

```
## Sample Flagging
```




Table 1: SO4 samples

SO4_Conc_flag	Percent_samples
Within_Range	99.2957746
bdl	0.7042254

Table 2: Cl samples


Cl_Conc_flag	Percent_samples
Within_Range	99.2957746
bdl	0.7042254

0.9 Check to see if samples run match metadata & merge info

```
## Some sample IDs are missing from metadata.
```

```
## [1] "GWI_202405_UP_LYSC_45CM" "MSM_202405_UP_LYSA_10CM"
```

0.10 Visualize Data by Plot



0.11 Export Processed Data

```
#end
```