

# Synoptic CB: Porewater SO<sub>4</sub>/Cl

October 2024 Samples

2025-09-05

## Contents

0.1	Run Information . . . . .	2
0.2	Assess Standard Curves . . . . .	3
0.3	Assess Check Standards . . . . .	4
0.4	Assess Blanks . . . . .	5
0.5	Assess Duplicates . . . . .	6
0.6	Calculate mmol/L concentrations & salinity, add dilutions . . . . .	8
0.7	Assess Analytical Spikes . . . . .	9
0.8	Check if samples within the range of the standard curve . . . . .	9
0.9	Check to see if samples run match metadata & merge info . . . . .	10
0.10	Visualize Data by Plot . . . . .	11
0.11	Export Processed Data . . . . .	11

##Add Required Packages

## 0.1 Run Information

```
##### Run information - PLEASE CHANGE
Date_Run = "2025-06-18" #Date that instrument was run
Run_by = "Zoe Read" #Instrument user
Script_run_by = "Zoe Read" #Code user
run_notes = "Some Cl/SO4 standards and checks were lower than the expected concentration" #any notes
samples <- c("GCW", "GWI", "MSM", "SWH") #whatever identifies your samples within the same names
samples_pattern <- paste(samples, collapse = "|")
#samples_pattern <- "GCW" #use this instead of the line above if you have only one site code

##### File Names - PLEASE CHANGE
#file path and name for raw summary data file
raw_file_name_cl = "Raw Data/COMPASS_Synoptic_CB_MonMon_202410_Cl.txt"
raw_file_name_so4 = "Raw Data/COMPASS_Synoptic_CB_MonMon_202410_SO4.txt"

#file path and name of processed data file
processed_file_name = "Processed Data/COMPASS_SynopticCB_PW_Processed_Cl_SO4_202410.csv"

##### Log Files - PLEASE CHECK
#downloaded metadata csv - downloaded from Google drive as csv for this year
Raw_Metadata = "Raw Data/COMPASS_SynopticCB_PW_SampleLog_2024.csv"

#qaqc log file path for this year
Log_path = "Raw Data/COMPASS_Synoptic_Cl_SO4_QAQClog_2024.csv"
```


##Set Up Code - constants and QAQC cutoffs

##Read in metadata and create similar sample IDs for matching to samples


##Import Sample Data

## 0.2 Assess Standard Curves


Chloride Std Curve



Sulfate Std Curve



```
## [1] "QAQC log file exists and has been read into the code."
```



```
## [1] "Cl Curve r2 GOOD"
```

```
## [1] "SO4 Curve r2 GOOD"
```


### 0.3 Assess Check Standards

```
## # A tibble: 5 x 5
##   sample_ID mean_Cl sd_Cl cv_Cl flag_Cl
##   <chr>      <dbl> <dbl> <dbl> <chr>
## 1 Standard 1  4.78 0.0475 0.00993 Chloride Check Standard RSD within Range - ~
## 2 Standard 2  9.68 0.0371 0.00383 Chloride Check Standard RSD within Range - ~
## 3 Standard 3 19.5 0.103 0.00527 Chloride Check Standard RSD within Range - ~
## 4 Standard 4 101. 1.06 0.0104 Chloride Check Standard RSD within Range - ~
## 5 Standard 5 200. 0.207 0.00103 Chloride Check Standard RSD within Range - ~
```

```
## # A tibble: 5 x 5
##   sample_ID mean_S04 sd_S04 cv_S04 flag_S04
##   <chr>      <dbl> <dbl> <dbl> <chr>
## 1 Standard 1  0.375 0.0352 0.0937 Sulfate Check Standard RSD within Range - ~
## 2 Standard 2  0.882 0.0177 0.0201 Sulfate Check Standard RSD within Range - ~
## 3 Standard 3  1.90 0.0123 0.00648 Sulfate Check Standard RSD within Range - ~
## 4 Standard 4 10.2 0.0931 0.00914 Sulfate Check Standard RSD within Range - ~
## 5 Standard 5 20.0 0.0436 0.00218 Sulfate Check Standard RSD within Range - ~
```

```
## [1] ">80% of Chloride Check Standards have RSD within range - PROCEED"
```

```
## [1] ">80% of Sulfate Check Standards have RSD within range - PROCEED"
```




```
## [1] ">80% of Chloride Check Standards are within range of expected concentration - PROCEED"
```

```
## [1] "<80% of Sulfate Check Standards are within range of expected concentration - REASSESS"
```

#### 0.4 Assess Blanks

```
## [1] ">80% of Chloride Blank concentrations are lower 25% quartile of samples"
```

```
## [1] ">80% of Sulfate Blank concentrations are lower 25% quartile of samples"
```



```
## Chloride blanks mean ppm:
```


```
## [1] 0.004233333
```

```
## Sulfate blanks mean ppm:
```

```
## [1] 0.001833333
```

## 0.5 Assess Duplicates


```
## Warning: Removed 1 row containing missing values or values outside the scale range
## ('geom_bar()').
```



```
## [1] ">80% of Chloride Duplicates have a CV <10%"
```

```
## [1] ">80% of Sulfate Duplicates have a CV <10%"
```

```
## Warning: Removed 1 row containing missing values or values outside the scale range  
## ('geom_bar()').
```



```
## [1] ">80% of Chloride Duplicates have a percent difference <10% - PROCEED"
```

```
## [1] ">80% of Sulfate Duplicates have a percent difference <10% - PROCEED"
```

## 0.6 Calculate mmol/L concentrations & salinity, add dilutions

```
# Convert ppm to mmol/L
all_dat$S04_Conc_mM <- (all_dat$S04_ppm / s_mw)
all_dat$Cl_Conc_mM <- (all_dat$Cl_ppm / cl_mw)

# Calculate Salinity
# calculated using the Knudsen equation
# Salinity = 0.03 + 1.8050 * Chlorinity
# Ref: A Practical Handbook of Seawater Analysis by Strickland & Parsons (P. 11)
# =((1.807*Cl_ppm)+0.026)/1000
all_dat$salinity <- ((1.8070 * all_dat$Cl_ppm) + 0.026) / 1000

#Need to determine dilution factors for your samples
#for Steph / COMPASS this depends on the site so...
all_dat$Dilution <- 1
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "MSM") & str_detect(all_dat$sample_ID, "UP"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "MSM") & str_detect(all_dat$sample_ID, "TR"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "MSM") & str_detect(all_dat$sample_ID, "WC"),
```



```

all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "MSM") & str_detect(all_dat$sample_ID, "SW"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GCW") & str_detect(all_dat$sample_ID, "UP"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GCW") & str_detect(all_dat$sample_ID, "TR"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GCW") & str_detect(all_dat$sample_ID, "WC"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GCW") & str_detect(all_dat$sample_ID, "SW"),


all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GWI") & str_detect(all_dat$sample_ID, "UP"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GWI") & str_detect(all_dat$sample_ID, "TR"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GWI") & str_detect(all_dat$sample_ID, "WC"),
all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "GWI") & str_detect(all_dat$sample_ID, "SW"),

all_dat$Dilution <- ifelse(str_detect(all_dat$sample_ID, "SWH"), 50, all_dat$Dilution)

# head(all_dat)

```

## 0.7 Assess Analytical Spikes



```
## [1] ">80% of S04 spikes have a recovery between the high and low cutoff - PROCEED"
```

## 0.8 Check if samples within the range of the standard curve

```
## Sample Flagging
```




Table 1: SO4 samples

SO4_Conc_flag	Percent_samples
Within_Range	91.489362
bdl	8.510638


Table 2: Cl samples

Cl_Conc_flag	Percent_samples
Within_Range	87.23404
bdl	12.76596

## 0.9 Check to see if samples run match metadata & merge info

```
## All sample IDs are present in metadata.
```

## 0.10 Visualize Data by Plot



## 0.11 Export Processed Data

#end