

Synoptic CB: Porewater Sulfide

November 2025 Samples Plate 4 20251114

2025-12-01

Run Information

```
##### Run information - PLEASE CHANGE
Date_Run = "20251114" #Date that instrument was run
Plate = "4"
Month = "Nov"
Year = "2025"
Run_by = "Zoe Read" #Instrument user
Script_run_by = "Zoe Read" #Code user
Project = "COMPASS"
run_notes = "MC 10 was bad.
Samples 25, 27, and 30 still had high CVs after one replicate was removed.
Sup and respective sample both had H2S values of 0.
Spk was bad.
50 uL of spike added.
" #any notes from the run

##### STDs to remove manually - NA unless standards have high CVs
stds_to_remove <- c("C02", "D03", "E03") #Ex. c("A02", "B02")

Chkstds_to_remove <- c(NA)

MCs_to_remove <- c(NA)

##Sample data that was entered incorrectly
# The Old ID is the original, incorrectly-entered ID and the New ID is the correct ID to change it to.
# Old_ID_1 = "17 Dup"
# New_ID_1 = "16 Dup"
cat(run_notes)

## MC 10 was bad.
## Samples 25, 27, and 30 still had high CVs after one replicate was removed.
## Sup and respective sample both had H2S values of 0.
## Spk was bad.
## 50 uL of spike added.
##

## [1] "Std Curve R squared: 0.9868"
```





