

Synoptic_CB_Nutrients_202308

August 2023

2025-06-25

Contents

0.1	Import Data & Clean	3
0.2	Assessing standard Curves	3
0.3	Dilution Corrections - ensure the latest dilution is kept	7
0.4	Analyze the Check Standards	8
0.5	Analyze Blanks	9
0.6	Analyze Duplicates	10
0.7	Spikes	11
0.8	Matrix Effects	12
0.9	Unit Converted Data Column Added (mg/L to uM)	12
0.10	Sample Flagging - Within range of standard curve	12
0.11	Pull out sample id information	12
0.12	Pulling Rhizon Samples	12
0.13	Check to see if samples run match metadata & merge info	12
0.14	Visualize Data	14
0.15	Export Processed Data	16

##Run Information

```
cat("Run Information: Input by User") #lets you know what section you're in
```

Run Information: Input by User

```
#set the run date & user name
```

```
run_date <- "20240114"
```

```
sample_year <- 2023
```

```
sample_month <- 08
```

```
user <- "Stephanie Wilson"
```

```
#identify the files you want to read in
```

```
#read in as a list to accomodate ultiple runs in a month
```

```
NOx_files <- c("Raw Data/SEAL_COMPASS_Synoptic_NOx_Aug2023_1.csv",
```

```
              "Raw Data/SEAL_COMPASS_Synoptic_NOx_Aug2023_2.csv",
```

```
              "Raw Data/SEAL_COMPASS_Synoptic_NOx_Aug2023_3.csv")
```

```
NH3_PO4_files <- c("Raw Data/SEAL_COMPASS_Synoptic_NH3_PO4_Aug2023_1.csv",
```

```
                  "Raw Data/SEAL_COMPASS_Synoptic_NH3_PO4_Aug2023_2.csv",
```

```
                  "Raw Data/SEAL_COMPASS_Synoptic_NH3_PO4_Aug2023_3.csv")
```

```
# Define the file path for QAQC log file - NO Need to change just check year
```

```
file_path <- "Raw Data/SEAL_COMPASS_Synoptic_QAQC_Log_2023.csv"
```

```
final_path <- "Processed Data/COMPASS_Synoptic_Nutrients_202308.csv"
```

```
#record any notes about the run or anything other info here:
```

```
run_notes <- "Duplicates are out of range because of detection limit, too small of values"
```

```
#Set up file path for metadata
```

```
#downloaded metadata csv - downloaded from Google drive as csv for this year
```

```
#https://docs.google.com/spreadsheets/d/1HCAN0_q6y17x0RUXVzID09hVal-RfwWc/edit?usp=sharing&ouid=10899
```

```
Raw_Metadata = "Raw Data/COMPASS_SynopticCB_PW_SampleLog_2023.csv"
```

```
##Setup
```

```
##Read in metadata and create similar sample IDs for matching to samples
```

0.1 Import Data & Clean


0.2 Assessing standard Curves

```
#Pull out standards data
```

```
## Assess Standard Curves
```


```
#Plot standards data
```

```
## Assess Standard Curves
```




```
## 'geom_smooth()' using formula = 'y ~ x'
```

NH3 Standard Curve



```
## 'geom_smooth()' using formula = 'y ~ x'
```

PO4 Standard Curve



```
## [1] "NOx Curve r2 GOOD - PROCEED"
```

```
## [1] "NH3 Curve r2 GOOD - PROCEED"
```

```
## [1] "PO4 Curve r2 GOOD - PROCEED"
```

```
## [1] "QAQC log file exists and has been read into the code."
```




Table 1: Average Slope by Analyte

Test	avg_slope
NH3	1.912
NOx	1.247
PO4	2.320

0.3 Dilution Corrections - ensure the latest dilution is kept


```
## Dilution Corrections
```

```
## Duplicated samples: GCW_202308_WC_LysA_45cm, MSM_202308_TR_LysA_20cm, MSM_202308_TR_LysB_20cm, MSM_202308_TR_LysC_20cm
```

```
## Dilution Present, Need to Correct
```

```
#Check NOx Reduction Efficiency
```

```
## Assess Reduction Efficiency
```



```
## [1] "Mean NOx Reduction Efficiency >95% - PROCEED"
```

```
## [1] 95.1625
```


0.4 Analyze the Check Standards

```
## Analyze Check Standards
```

```
## [1] "NOx Check Standard RSD within Range - PROCEED"
```

```
## [1] "NH3 Check Standard RSD within Range - PROCEED"
```

```
## [1] "PO4 Check Standard RSD within Range - PROCEED"
```



```
## [1] ">60% of NOx Check Standards are within range of expected concentration - PROCEED"
```

```
## [1] ">60% of NH3 Check Standards are within range of expected concentration - PROCEED"
```

```
## [1] ">60% of PO4 Check Standards are within range of expected concentration - PROCEED"
```


0.5 Analyze Blanks

Assess Blanks

[1] ">60% of NOx Blank concentrations are lower than the lower 25% quartile of samples - PROCEED"

[1] ">60% of NH3 Blank concentrations are lower than the lower 25% quartile of samples - PROCEED"

[1] ">60% of PO4 Blank concentrations are lower than the lower 25% quartile of samples- PROCEED"




Table 2: Mean Concentration of Blanks

Test	Blank_Mean_Conc
NOx	0.0088
NH3	-0.0020
PO4	0.0022

0.6 Analyze Duplicates

```
## Analyze Duplicates
```

```
## [1] ">60% of NOx Duplicates have a CV <10% - PROCEED"
```

```
## [1] ">60% of NH3 Duplicates have a CV <10% - PROCEED"
```

```
## [1] ">60% of PO4 Duplicates have a CV <10% - PROCEED"
```


```
## Warning: Using 'size' aesthetic for lines was deprecated in ggplot2 3.4.0.
```

```
## i Please use 'linewidth' instead.
```

```
## This warning is displayed once every 8 hours.
```

```
## Call 'lifecycle::last_lifecycle_warnings()' to see where this warning was
```

```
## generated.
```




0.7 Spikes

[1] ">60% of Spikes have a CV <50% - PROCEED"

[1] ">60% of Spikes have a CV <50% - PROCEED"

[1] ">60% of Spikes have a CV <50% - PROCEED"



0.8 Matrix Effects

```
## [1] "NO NOx Matrix Effect, PROCEED"
```

```
## [1] "NO NH3 Matrix Effect, PROCEED"
```

```
## [1] "NO PO4 Matrix Effect, PROCEED"
```

0.9 Unit Converted Data Column Added (mg/L to uM)

0.10 Sample Flagging - Within range of standard curve

```
## Sample Flagging
```

0.11 Pull out sample id information

```
## Sample Processing
```

```
## Warning: Expected 5 pieces. Missing pieces filled with 'NA' in 36 rows [20, 21, 22, 46,
## 47, 48, 84, 85, 86, 109, 110, 111, 130, 131, 132, 168, 169, 170, 199, 200,
## ...].
```

```
## Warning: There was 1 warning in 'mutate()'.
## i In argument: 'Samp_Time = ym(Samp_Time)'.
## Caused by warning:
## ! 9 failed to parse.
```

0.12 Pulling Rhizon Samples

0.13 Check to see if samples run match metadata & merge info

```
## Check Sample IDs with Metadata
```

```
## Some sample IDs are missing from metadata.
```


```
## [1] "GCW_202308_UP_LysA_20cm" "GCW_202308_UP_LysB_10cm"
## [3] "GCW_202308_UP_LysB_20cm" "GCW_202308_UP_LysC_10cm"
## [5] "MSM_202038_TR_LysB_20cm" "MSM_202038_WC_LysA_10cm"
## [7] "MSM_202038_WC_LysA_20cm" "SWH_202308_UPCON_LysA_10cm"
## [9] "SWH_202308_UPCON_LysA_20cm" "SWH_202308_UPCON_LysA_45cm"
## [11] "SWH_202308_UPCON_LysB_10cm" "SWH_202308_UPCON_LysB_20cm"
## [13] "SWH_202308_UPCON_LysB_45cm" "SWH_202308_UPCON_LysC_10cm"
## [15] "SWH_202308_UPCON_LysC_20cm" "SWH_202308_UPCON_LysC_45cm"
## [17] "SWH_202308_WC_LysB_20cm" "SWH_202308_WC_LysC_20cm"
## [19] "GCW_202038_WC_LysB_10cm" "GCW_202038_WC_LysC_45cm"
## [21] "SWH_202038_UPCON_LysA_45cm"
```

```
## Warning: Expected 5 pieces. Missing pieces filled with 'NA' in 36 rows [20, 21, 22, 46,  
## 47, 48, 84, 85, 86, 109, 110, 111, 130, 131, 132, 168, 169, 170, 199, 200,  
## ...].
```


```
## Warning: There was 1 warning in 'mutate()'.  
## i In argument: 'Samp_Time = ym(Samp_Time)'.  
## Caused by warning:  
## ! 9 failed to parse.
```


0.14 Visualize Data

Visualize Data



Porewater NH3





0.15 Export Processed Data

#end