

Synoptic CB: Porewater Sulfide

May 2026 Samples Ran on 20260526

2026-05-29

Run Information

###things that need to be changed

Date_Run = "20260526"

plates<- c("Plate1","Plate2","Plate3","Plate4","Plate5","Plate6")

Month = "May"

Year = "2026"

Sample_Month_Year = "202605"

Run_by = "Zoe Read" #Instrument user

Script_run_by = "Zoe Read" #Code user


Project = "COMPASS"

Run_notes="Used Plate 6 Std Curve. 1/6 dups bad, probably because the original value was 0. 3/6 spks were 1

#Stds that should be excluded

stds_to_remove<-data.frame(Plate=c("Plate6"),IDs=c("Std 3"))

stds_to_remove<-NA



Checking STD Data against QAQC file






Table 1: Best std curve to use:

Date	Project	R2	Slope	Intercept	Top_STD	Plate
20260526	COMPASS	0.9808451	0.0125426	0.101497	100	Plate6


Matrix Check QAQC




Sample triplicate means and sd dev before bad reps removed



Remove bad reps




Check the dups for QAQC



```
## [1] ">60% of Duplicates are within <10%"
```

Check the spks for QAQC




```
## [1] "<60% of Spikes are out of range - REASSESS"
```

Check to see if samples run match metadata & merge info

```
## ***All sample IDs are present in metadata.***
```

Visualize Data by Plot

Samples: Sulfide



###END