


DETAIL A SCALE 4:1



## **NOTES:**

В

- MATERIAL CERTIFICATION IS REQUIRED PARTS SHALL BE FREE OF ANY PROCESS
- CONTAMINANTS
- PARTS SHALL BE BURR FREE
- PARTS SHALL BE PACKAGED AS TO PREVENT SHIPPING DAMAGE
- MED-4800 MASTERBATCH PIGMENT ADDED (GREEN) TO SILICONE OVERMOLD
- PINOUT: TIP-GREEN (C) RING-CLEAR (D)



Open source implantable devices

THE INFORMATION CONTAINED IN THIS DRAWING IS PUBLICLY RELEASED BY COSMIC UNDER AN OPEN SOURCE LICENSE, COSMIC IS NOT LIABLE FOR ANY ASPECT OF USE OF THIS DRAWING, INCLUDING EDITS, FABRICATION, AND REPRODUCTION OF THIS DESIGN.

| DO NOT SCALE DRAWING  INTERPRET DIMENSIONS AND TOLERANCES PER ASME Y14.5M - 2018  DIMENSIONS IN INCHES UNLESS OTHERWISE SPECIFIED | THIRD ANGLE PROJECTION |
|---|------------------------|
| MATERIAL  | FINISH                 |

DOC. NAME

## Electrode Cable, IM, MES

|   | PROJECT | DATE     | SIZE | DOC. NUMBER                            | REV |
|---|---------|----------|------|--|-----|
|   | NNP     | 7/9/2024 | A    | \\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\ | 1.0 |
| ı | SHEET   | SCALE    | Α    | NNP-DWG-145-004-000                    | 1.0 |
|   | 1 OF 1  | 2:1      |      |  |     |
|   |         |          |      |  |     |