




4

3

2

1




NOTE:

- USE "VERTREL XSI" SOLVENT TO EXPAND SILICONE TUBING PRIOR TO INSERTING COILS.



THE INFORMATION CONTAINED IN THIS DRAWING IS PUBLICLY RELEASED BY COSMIC UNDER AN OPEN SOURCE LICENSE. COSMIC IS NOT LIABLE FOR ANY ASPECT OF USE OF THIS DRAWING, INCLUDING EDITS, FABRICATION, AND REPRODUCTION OF THIS DESIGN.

DO NOT SCALE DRAWING  
INTERPRET DIMENSIONS AND TOLERANCES PER  
ASME Y14.5M - 2018  
DIMENSIONS IN INCHES UNLESS OTHERWISE SPECIFIED



MATERIAL

FINISH

DOC. NAME  
SA, Cable Body, Insulated SS Filars,  
Red-Clear

PROJECT NNP	DATE 7/10/2024	SIZE <b>B</b>	DOC. NUMBER NNP-DWG-145-007-002	REV 1.0
SHEET 1 OF 1	SCALE 2:1			

ITEM NO.	DESCRIPTION
1	Filar, SS, Red-Clear
2	Cable, Insulating Silicone Tubing
3	Cable Body, Insulated SS Filars, Red-Clear

4

3

2

1