
2





DETAIL A SCALE 4:1



## NOTES:

- 1. MATERIAL CERTIFICATION IS REQUIRED
- PARTS SHALL BE FREE OF ANY PROCESS CONTAMINANTS
- 3. PARTS SHALL BE BURR FREE
- 4. PARTS SHALL BE PACKAGED AS TO PREVENT SHIPPING DAMAGE
- 5. MED-4800 MASTERBATCH PIGMENT ADDED (RED) TO SILICONE OVERMOLD
- 6. PINOUT: TIP-ELECTRODE (RED&CLEAR) RING (NOT CONNECTED)



Open source implantable devices

THE INFORMATION CONTAINED IN THIS DRAWING IS PUBLICLY RELEASED BY COSMIIC UNDER AN OPEN SOURCE LICENSE. COSMIIC IS NOT LIABLE FOR ANY ASPECT OF USE OF THIS DRAWING, INCLUDING EDITS, FABRICATION, AND REPRODUCTION OF THIS DESIGN.

DO NOT SCALE DRAWING

INTERPRET DIMENSIONS AND TOLERANCES PER
ASME Y14.5M - 2018

DIMENSIONS IN INCHES UNLESS OTHERWISE SPECIFIED

MATERIAL

THIRD ANGLE PROJECTION

THIRD ANGLE PROJECTION

FINISH

DOC. NAME

## Electrode Cable, IM, Stim

	PROJECT	DATE	SIZE	DOC. NUMBER	REV
	NNP	7/9/2024			1.0
ı	SHEET	SCALE	A	NNP-DWG-145-005-000	1.0
	1 OF 1	2:1			

A

A

2

.