Preprocessing I: Within Subject

John Ashburner

Pre-processing Overview



Contents

* **Preliminaries**
  * **Rigid-Body and Affine Transformations**
  * **Optimisation and Objective Functions**
  * **Transformations and Interpolation**
* Realignment & Coregistration

Rigid-Body Transformations

* Assume that brain of the same subject doesn’t change shape or size in the scanner.
  * Head can move, but remains the same shape and size.
  * Some exceptions:
    * Image distortions.
    * Brain slops about slightly because of gravity.
    * Brain growth or atrophy over time.
* If the subject’s head moves, we need to correct the images.
  * Do this by image registration.

Image Registration

Two components:

- **Registration** - i.e. Optimise the parameters that describe a spatial transformation between the source and reference images

- **Transformation** - i.e. Re-sample according to the determined transformation parameters

2D Affine Transforms



* Translations by t_x and t_y
  * $x_1 = x_0 + t_x$
  * $y_1 = y_0 + t_y$
* Rotation around the origin by Θ radians
  * $x_1 = \cos(\Theta)\, x_0 + \sin(\Theta)\, y_0$
  * $y_1 = -\sin(\Theta)\, x_0 + \cos(\Theta)\, y_0$
* Zooms by s_x and s_y
  * $x_1 = s_x\, x_0$
  * $y_1 = s_y\, y_0$
* Shear
  * $x_1 = x_0 + h\, y_0$
  * $y_1 = y_0$

2D Affine Transforms



* Translations by t_x and t_y
  * $x_1 = 1\ x_0 + 0\ y_0 + t_x$
  * $y_1 = 0\ x_0 + 1\ y_0 + t_y$
* Rotation around the origin by Θ radians
  * $x_1 = \cos(\Theta)\ x_0 + \sin(\Theta)\ y_0 + 0$
  * $y_1 = -\sin(\Theta)\ x_0 + \cos(\Theta)\ y_0 + 0$
* Zooms by s_x and s_y:
  * $x_1 = s_x\ x_0 + 0\ y_0 + 0$
  * $y_1 = 0\ x_0 + s_y\ y_0 + 0$
* Shear
  * $x_1 = 1\ x_0 + h\ y_0 + 0$
  * $y_1 = 0\ x_0 + 1\ y_0 + 0$

3D Rigid-body Transformations

* A 3D rigid body transform is defined by:
  * 3 translations - in X, Y & Z directions
  * 3 rotations - about X, Y & Z axes
* The order of the operations matters

$$\begin{pmatrix} 1 & 0 & 0 & X_{trans} \\ 0 & 1 & 0 & Y_{trans} \\ 0 & 0 & 1 & Zt_{rans} \\ 0 & 0 & 0 & 1 \end{pmatrix} \times \begin{pmatrix} 1 & 0 & 0 & 0 \\ 0 & \cos\Phi & \sin\Phi & 0 \\ 0 & -\sin\Phi & \cos\Phi & 0 \\ 0 & 0 & 0 & 1 \end{pmatrix} \times \begin{pmatrix} \cos\Theta & 0 & \sin\Theta & 0 \\ 0 & 1 & 0 & 0 \\ -\sin\Theta & 0 & \cos\Theta & 0 \\ 0 & 0 & 0 & 1 \end{pmatrix} \times \begin{pmatrix} \cos\Omega & \sin\Omega & 0 & 0 \\ -\sin\Omega & \cos\Omega & 0 & 0 \\ 0 & 0 & 1 & 0 \\ 0 & 0 & 0 & 1 \end{pmatrix}$$

Translations | Pitch about x axis | Roll about y axis | Yaw about z axis

Voxel-to-world Transforms

* Affine transform associated with each image
  * Maps from voxels ($x=1..n_x$, $y=1..n_y$, $z=1..n_z$) to some world co-ordinate system. e.g.,
    * Scanner co-ordinates - images from DICOM toolbox
    * T&T/MNI coordinates - spatially normalised
* Registering image B (source) to image A (target) will update B's voxel-to-world mapping
  * Mapping from voxels in A to voxels in B is by
    * A-to-world using M_A, then world-to-B using M_B^{-1}
    * $M_B^{-1} M_A$

Left- and Right-handed Coordinate Systems

* NIfTI format files are stored in either a left- or right-handed system
  * Indicated in the header
* Talairach & Tournoux uses a right-handed system
* Mapping between them sometimes requires a flip
  * Affine transform has a negative determinant





Optimisation

* Image registration is done by optimisation.
* Optimisation involves finding some “best” parameters according to an “objective function”, which is either minimised or maximised
* The “objective function” is often related to a probability based on some model



Objective Functions



* Intra-modal
  * Mean squared difference (minimise)
  * Normalised cross correlation (maximise)



* Inter-modal (or intra-modal)
  * Mutual information (maximise)
  * Normalised mutual information (maximise)
  * Entropy correlation coefficient (maximise)

Simple Interpolation

* Nearest neighbour
  * Take the value of the closest voxel
* Tri-linear
  * Just a weighted average of the neighbouring voxels
  * $f_5 = f_1 x_2 + f_2 x_1$
  * $f_6 = f_3 x_2 + f_4 x_1$
  * $f_7 = f_5 y_2 + f_6 y_1$



B-spline Interpolation

A continuous function is represented by a linear combination of basis functions



2D B-spline basis functions of degrees 0, 1, 2 and 3



B-splines are piecewise polynomials



Nearest neighbour and trilinear interpolation are the same as B-spline interpolation with degrees 0 and 1.

Contents

* Preliminaries
* **Realignment & Coregistration**
  * **Realignment by minimising mean-squared difference**
  * **Residual artifacts and distortion correction**
  * **Coregistration by maximising mutual information**

Mean-squared Difference







* Minimising mean-squared difference works for intra-modal registration (realignment)
* Simple relationship between intensities in one image, versus those in the other
  * Assumes normally distributed differences

Residual Errors from aligned fMRI

* Re-sampling can introduce interpolation errors
  * especially tri-linear interpolation
* Gaps between slices can cause aliasing artefacts
* Slices are not acquired simultaneously
  * rapid movements not accounted for by rigid body model
* Image artefacts may not move according to a rigid body model
  * image distortion
  * image dropout
  * Nyquist ghost

* Functions of the estimated motion parameters can be modelled as confounds in subsequent analyses

Movement by Distortion Interaction of fMRI

* Subject disrupts B_0 field, rendering it inhomogeneous
    * distortions in phase-encode direction
* Subject moves during EPI time series
* Distortions vary with subject orientation
    * shape varies



Movement by distortion interaction

Original position



After rotation



Correcting for distortion changes using Unwarp



Estimate movement parameters.

Estimate reference from mean of all scans.

Estimate new distortion fields for each image:

- estimate rate of change of field with respect to the current estimate of movement parameters in **pitch** and **roll**.



Unwarp time series.



Andersson et al, 2001

Coregistration

- Inter-modal registration.
- Match images from same subject but different modalities:
  - anatomical localisation of single subject activations
  - achieve more precise spatial normalisation of functional image using anatomical image.





Coregistration maximises Mutual Information





* Used for between-modality registration
* Derived from joint histograms
* $MI = \int_{ab} P(a,b) \log_2 [P(a,b)/(P(a) P(b))]$
  * Related to entropy: MI = -H(a,b) + H(a) + H(b)
    * Where $H(a) = -\int_a P(a) \log_2 P(a)$ and $H(a,b) = -\int_a P(a,b) \log_2 P(a,b)$

References

* Friston et al. *Spatial registration and normalisation of images*.
  Human Brain Mapping 3:165-189 (1995).
* Collignon et al. *Automated multi-modality image registration based on information theory*. IPMI'95 pp 263-274 (1995).
* Thévenaz et al. *Interpolation revisited*.
  IEEE Trans. Med. Imaging 19:739-758 (2000).
* Andersson et al. *Modeling geometric deformations in EPI time series*.
  Neuroimage 13:903-919 (2001).