

NO	PRODUCT SPECIFICATIONS	Wishes	Requirements	Standards	Company	Subcontractor	Distributors	End User	Production	Distribution	Use	Disposal
1	To minimize the electrical power usage		✓					✓			✓	
2	To operate all the time(24 hours) by the domestic AC power ratings(230V , 50 Hz)		✓								✓	
3	The place which the device to be installed, must be dry and clean			✓				✓			✓	
4	The heat dissipation must be minimum		✓								✓	
5	Life expectancy of 3 years	✓						✓			✓	
6	To make the interconnections between product and operator as possible as.		✓					✓			✓	
7	Avoiding monitoring circumstances that required lengthy concentrations	✓						✓			✓	
8	Maintenance and repair of the product must be easy		✓		✓	✓			✓		✓	
9	Production cost must be less than 6000.00 Rupees	✓			✓	✓			✓			
10	The packing should guarantee the protection of the product		✓		✓		✓			✓		
11	Weight should be less than 600 grams		✓		✓			✓			✓	
12	Can be easily installed and operated, by a person with low technical skill and knowledge.		✓		✓						✓	
13	The color of the product must be a light color	✓			✓	✓			✓		✓	
14	The product must ensure about 100% accuracy of the measurements		✓		✓			✓	✓		✓	
15	The product should guarantee 100% safety of the patient		✓		✓	✓		✓	✓		✓	
16	The product must be easy to install	✓						✓			✓	
17	Should not contain harmful materials		✓		✓							✓
18	Metals, glass, plastics are majorly used in the manufacture of product.		✓		✓	✓			✓			✓
19	The product is packaged using air bubble material to comply with standards to provide a safe environment from exterior forces during distribution.		✓				✓			✓		
20	Includes a leaflet of user instructions in the packaging		✓		✓		✓			✓	✓	
21	The product must be reuse after repairs and maintenance		✓								✓	
22	The product must be able to recycle		✓		✓						✓	✓