
Last modified on 8/18/2023 at 11:31 AM

Printed on 12/8/2023 at 12:24 PM

# Dymin\_RT\_Main.vi





TE Stop

[132] Threshold (counts)

Theshold (counts)

TF **Dymin On?** 

Laser Wait (us)

[DBL] **DyMIN Step Time (us)** 

Max AOM Voltage

Last modified on 8/18/2023 at 11:31 AM

Printed on 12/8/2023 at 12:24 PM



#### error in

<B>error in</B> can accept error information wired from VIs previously called. Use this information to decide if any functionality should be bypassed in the event of errors from other VIs.

Right-click the <B>error in</B> control on the front panel and select <B>Explain Error</B> or <B>Explain Warning</B> from the shortcut menu for more information about the error.



#### status

<B>status</B> is TRUE (X) if an error occurred or FALSE (checkmark) to indicate a warning or that no error occurred.

Right-click the <B>error in</B> control on the front panel and select <B>Explain Error</B> or <B>Explain Warning</B> from the shortcut menu for more information about the error.

# 132

#### code

<B>code</B> is the error or warning code.

Right-click the <B>error in</B> control on the front panel and select <B>Explain Error</B> or <B>Explain Warning</B> from the shortcut menu for more information about the error.

# abc

#### source

<B>source</B> describes the origin of the error or warning.

Right-click the <B>error in</B> control on the front panel and select <B>Explain Error</B> or <B>Explain Warning</B> from the shortcut menu for more information about the error.



#### Calibrate?



**AOM Units** 



AOM P or V



AOM P or V



## error out

<B>error out</B> passes error or warning information out of a VI to be used by other VIs.

Right-click the <B>error out</B> indicator on the front panel and select <B>Explain Error</B> or <B>Explain Warning</B> from the shortcut menu for more information about the error.



132

code



**Source** 



# **AOM Voltage**



Max AOM Voltage

[132]

Time (#80MHz periods)



**132** dT (# 80MHz periods)



PIX CLK Measured (bits)

PIX CLK Measured (us)

Last modified on 8/18/2023 at 11:31 AM

Printed on 12/8/2023 at 12:24 PM

132

PIX CLK period (bits)

DBL


PIX CLK period (us)

[DBL]

All Step Times (us)



**Max AOM Voltage** 



Add an option to use old calibration data if I want to
Just save the fit paramters in a file for later with date created in filename

Multiplexer inputs from FPGA DIOs:

Multiplexer digital input names an A0-DIO7


SS A1 A0 CK1 CK2 -> Output

CK=CKBAR=DIO9=0 always

CK=CKBAR=DIO9=0 always


CS=DIO10=0 (unless High Z desired)

1000 -> NR2 = C/AO1



Last modified on 8/18/2023 at 11:31 AM

Printed on 12/8/2023 at 12:24 PM



Dymin\_RT\_Main.vi Q:\GitHub\DyminSTED4all\LabviewCode\Dymin\_RT\_Main.vi **\$** 

Printed on 12/8/2023 at 12:24 PM


Last modified on 8/18/2023 at 11:31 AM

Dymin\_RT\_Main.vi

Q:\GitHub\DyminSTED4all\LabviewCode\Dymin\_RT\_Main.vi


Last modified on 8/18/2023 at 11:31 AM

Printed on 12/8/2023 at 12:24 PM




Last modified on 8/18/2023 at 11:31 AM

Printed on 12/8/2023 at 12:24 PM



Dymin\_RT\_Main.vi Q:\GitHub\DyminSTED4all\LabviewCode\Dymin\_RT\_Main.vi Last modified on 8/18/2023 at 11:31 AM Printed on 12/8/2023 at 12:24 PM



Last modified on 8/18/2023 at 11:31 AM

Printed on 12/8/2023 at 12:24 PM



### Calculate\_aomVolts.vi

Q:\GitHub\DyminSTED4all\LabviewCode\Calculate\_aomVolts.vi



### CalculateTimes.vi

Q:\GitHub\DyminSTED4all\LabviewCode\CalculateTimes.vi



## Calibrate\_power.vi

Q:\GitHub\DyminSTED4all\LabviewCode\Calibrate\_power.vi



# myRIO v1.0 Reset FPGA.vi

C:\Program Files (x86)\National Instruments\LabVIEW 2016\vi.lib\myRIO\Common\Instrument Driver Framework\myRIO v1.0\myRIO v1.0 Reset FPGA.vi



# myRIO v1.0 Timer IRQ Info FGV.vi

C:\Program Files (x86)\National Instruments\LabVIEW 2016\vi.lib\myRIO\Common\Instrument Driver Framework\myRIO v1.0\IRQ\vis\myRIO v1.0 Timer IRQ Info FGV.vi



### **Error Cluster From Error Code.vi**

C:\Program Files (x86)\National Instruments\LabVIEW 2016\vi.lib\Utility\error.llb\Error Cluster From Error Code.vi



# myRIO v1.0 ISR Agent.vi

C:\Program Files (x86)\National Instruments\LabVIEW 2016\vi.lib\myRIO\Common\Instrument Driver Framework\myRIO v1.0\IRQ\vis\myRIO v1.0 ISR Agent.vi



#### FPGA Ref Manager.vi

C:\Program Files (x86)\National Instruments\LabVIEW 2016\vi.lib\myRIO\Common\Instrument Driver Framework\Utilities\vis\FPGA Ref Manager.vi



#### myRIO v1.0 IRQ Info FGV.vi

C:\Program Files (x86)\National Instruments\LabVIEW 2016\vi.lib\myRIO\Common\Instrument Driver Framework\myRIO v1.0\IRQ\vis\myRIO v1.0 IRQ Info FGV.vi



## Calibrate\_control.ctl

Q:\GitHub\DyminSTED4all\LabviewCode\Calibrate\_control.ctl



#### Utilities.lvlib:Reset myRIO.vi

C:\Program Files (x86)\National Instruments\LabVIEW 2016\vi.lib\myRIO\Instrument Drivers\Utilities\vis\ Reset myRIO.vi



# AOM\_PorV\_control\_single.ctl

Q:\GitHub\DyminSTED4all\LabviewCode\AOM\_PorV\_control\_single.ctl