



**METRIC**

**INSTRUMENTS/GAUGE HISTORY CARD**

<b>Gauge Number</b> : Test 2001				<b>Instrument / Gauge Name</b> : Plain Plug Gauge				<b>Permissible Size</b> :		
<b>Gauge Serial No</b> : 222				<b>Range / Size</b> : 20.5/20.9 - mm				GO : 20.5255/20.5465		
<b>Calibration Source</b> : In house				<b>Frequency of Calibration</b> : 6				NOGO : 20.8895/20.9105		
				<b>Department</b> : N/A						

SI No.	Calibration Date	Calibration Status	Next Calibration	Certificate Status	Certificate No	Observed Error	Calibration At	Calibrated by	Verified by	Remarks
1	7-Sep-23		7-Mar-24			GO : 0/0 NOGO : 0/0				
2	4-Sep-23	OK	4-Mar-24	Received	Metric/18/2023	GO 20.528/20.529 NOGO 20.9/20.901	In House	Bala M		