

				CALIBRA	TION	CERTIF	ICATE					
Certific	Certificate No. Metric/18/2023					DEVICE UNDER CALURDATION DETAILS						
CUSTOMER DETAILS						— DEVICE UNDER CALIBRATION DETAILS						
Name METRIC						Description		Plain Plug Gauge				
Address Avadi, , Ch		nennai			Make		Gnn					
					SIZE		20.5/20.9 mm					
					Gauge No		Test 2001					
					Gauge Serial No		222					
Issue D	ate			Calibrated on		4-Sep-23						
Received Condition of DUC Satisfac			Satisfacto	ory		Next Due		4-Mar-24				
ENVIRONMENTAL CONDITION												
Temperature $20 \pm 2^{\circ}\text{C}$			$20 \pm 2^{\circ}C$	Hum		dity $50 \pm 10\%$		Drawing		No. :		
Calibration Procedure No				Standard Ref :				Uncertainty:				
MASTER USED FOR CALIBRATION												
SI. No.	SI. No. Name of the Master Used			Id No	Calibration Report No		Valid Upto		Traceability			
1	1 Outside Micrometer (ODM/A/001	Tii	Tii221217-5-9		17-Dec-23		Tii Techno Testing Serv		

CALIBRATION RESULTS

Parameter	Nominal Size	Permissible Size Min	Permissible Size Max	Observed Size Min	Observed Size Max	Cal status
GO	20.5000	20.5255	20.5465	20.5280	20.5290	ACCEPTED
NOGO	20.9000	20.8895	20.9105	20.9000	20.9010	ACCEPTED

GAUGE IS ACCEPTABLE

Remarks:

This Certificate is Digitally Signed and does not require any seal or signature

----End of the Certificate----

Bala M

Calibrated by Authorised by