AB-220 Certified sample calibration certificate

Fictional document for Northfield Lab Supply, a fictional company created for a portfolio project. Not for real use.

Item	Value
Document ID	NLS-DOC-AB220-CERT-SAMPLE-R1
Product	Analytical Balance AB-220
Applies to	NLS-AB220-CRT
Revision	R1
Effective date	2026-03-01
Status	current

SAMPLE - not valid for any unit

This is a sample of the certificate supplied with the AB-220 Certified. Each unit's own certificate is issued with that unit and carries its serial number.

Item	Value
Certificate number	SAMPLE-0000
Instrument	Analytical Balance AB-220 Certified
Serial number	SAMPLE
Date of calibration	SAMPLE
Conditions	21 °C, 45 % humidity

Results

Test weight (g)	Reading (g)	Error (mg)	Uncertainty (mg)
10.0000	10.0001	+0.1	0.2
50.0000	49.9999	-0.1	0.2
100.0000	100.0001	+0.1	0.2
200.0000	200.0000	0.0	0.3

Readings are traceable to national standards through the test weights used. The certificate does not replace calibration checks by the user.