Serological pipettes 10 ml certificate of analysis (typical lot)

Fictional document for Northfield Lab Supply, a fictional company created for a portfolio project. Not for real use.

Item	Value
Document ID	NLS-DOC-SP10-COA-R1
Product	Serological Pipettes 10 ml
Applies to	NLS-SP10-P
Revision	R1
Effective date	2026-01-01
Status	current

Typical lot results

These are typical results for a representative lot, supplied for information. They are not a guarantee for any particular lot.

Test	Result
Sterility	Pass
Pyrogens (endotoxin)	Below 0.01 EU/ml
Volume accuracy	Within ±1 % of nominal