


Title: Pressure Sensor App Communication Protocol and steps

Date: 6/05/2021

Revision: X03

Compatible PCB: Sensor_Glove-X01



Turn-on Instructions:

Supply a voltage potential of 3.4V to 5.5V between the VBAT+ and VBAT- connection points. The system is still off though. Voltage is supplied to the battery powered components, however, the nRF52 is not powered. To enable the linear regulator, you have to use the magnet on the popsicle stick to turn on the magnetic sensor. Hold the magnet over the magnetic sensor for a couple seconds. There will be an orange LED that will blink twice to notify the user that the patch is on and enabled. Now the system will begin advertising for three minutes. If the user does not connect to the patch within three minutes, the patch will shutdown requiring the user to use the magnet to connect to the patch again.

BLE GATT Database:

Configuration Service: CE:13:10:B0:F0:74:3E:A1:EE:45:30:38:D4:8C:AD:8A

-Settings Characteristic: 06:0D:00:CE:13:10:B0:F0:74:3E:A1:EE:45:30:38:AA:46:AD:8A

Data type: uint8_t [10]

Bluetooth Header	Module	Sub Message #1	Sub Message #2	Sub Message #3	Sub Message #4	Sub Message #5	Sub Message #6	Sub Message #7	Bluetooth Footer
------------------	--------	----------------	----------------	----------------	----------------	----------------	----------------	----------------	------------------

-Response Characteristic: 12:0F:00:CE:13:10:B0:F0:74:3E:A1:EE:45:30:38:AB:46:AD:8A

Data type: uint8_t [4]

Response #1	Response #2	Response #3	Response #4
-------------	-------------	-------------	-------------

Pressure Service: CE:13:10:B0:F0:74:3E:A1:EE:45:30:38:D7:8C:AD:8A

-Pressure Characteristic: 12:1D:00:CE:13:10:B0:F0:74:3E:A1:EE:45:30:38:B0:46:AD:8A

Data type: uint8_t [12]

CIN #1 [0]	CIN #1 [1]	CIN #1 [2]	CIN #2 [0]	CIN #2 [0]	CIN #2 [1]	CIN #2 [2]	CIN #3 [0]	CIN #3 [1]	CIN #3 [2]	CIN #4 [0]	CIN #4 [1]	CIN #4 [2]
------------------	------------------	------------------	------------------	------------------	------------------	------------------	------------------	------------------	------------------	------------------	------------------	------------------

The capacitance measurement are 24-bit result registers in binary format. The data is encoded in two's complement format. The result of the measurement can be calculated using the following formula:

$$\text{Capacitance (pF)} = ((\text{Two's Complement (measurement[23:0])}) / 2^{19}) + C_{\text{CAPDAC}}$$

BLE Command Instructions

1. Connect to Device.
 - a. Device name: SG-P-X01
 - b. Manufacturer name: GT-BITNG
2. Discover all attributes and services.
3. Enable all characteristic notifications.
4. Write the following command to the SETTINGS CHARACTERISTIC
 - a. Enable Bluetooth Advertising after disconnection.

BLUETOOTH HEADER	BLUETOOTH MODULE	BLUETOOTH ENABLE ADVERTISING AFTER DISCONNECTION COMMAND	0	0	0	0	0	0	BLUETOOTH FOOTER
0XB8	0X06	0X18	0	0	0	0	0	0	0XBB

- b. The Response Characteristic should read the following after the previous command:

0x00	0x00	0x00	0x01
------	------	------	------

5. Write the following command to the SETTINGS CHARACTERISTIC
 - a. Enable the measurement channel for the FDC1004.

BLUETOOTH HEADER	FDC1004 MODULE	FDC1004 ENABLE CHANNEL COMMAND	CHANNEL	0	0	0	0	0	BLUETOOTH FOOTER
0XB8	0X09	0X0C	X	0	0	0	0	0	0XBB

0XBA	0X09	0X0E	0	0	0	0	0	0	0XBB
------	------	------	---	---	---	---	---	---	------

* This command is to set internal variables to monitor the transmission of data from the FDC1004 to the nRF52.

- b. The Response Characteristic should read the following after the previous command:

0x00	0x00	0x00	0x01
------	------	------	------

- c. The previous command will be implemented. After implementation the following command, the response will be sent to indicate that the temperature data collection has finished.

0x00	0x00	0x00	0x0D
------	------	------	------

9. Write the following command to the SETTINGS CHARACTERISTIC

- a. Change the sampling rate if you desire.

BLUETOOTH HEADER	NRF52 MODULE	NRF52 RTC CLOCK COMMAND	NRF52 RTC SENSOR SET COUNTER	COUNTER[0]	COUNTER[1]	0	0	0	BLUETOOTH FOOTER
0XBA	0X01	0X04	0X08	X	X	0	0	0	0XBB

The X values signify a byte in an unit16_t data type. Counter[0] is the MSB and Counter[1] is the LSB. The counter / 8 = seconds per sample.

- b. The Response Characteristic should read the following after the previous command:

0x00	0x00	0x00	0x01
------	------	------	------

10. Write the following command to the SETTINGS CHARACTERISTIC

- a. Start the internal timer to regulate the sampling rate.

BLUETOOTH HEADER	NRF52 MODULE	NRF52 RTC CLOCK COMMAND	NRF52 RTC SENSOR START	0	0	0	0	0	BLUETOOTH FOOTER
0XBA	0X01	0X04	0X09	0	0	0	0	0	0XBB

* Only call this function once to start the timer to regulate the sampling rate.

Data is written to the pressure characteristic above at the desired sampling rate. After data is written to the characteristic, a notification is sent to the BLE Client.

- b. The Response Characteristic should read the following after the previous command:

0x00	0x00	0x00	0x01
------	------	------	------

To stop data transfer of pressure, send the following commands:

1. Write the following command to the SETTINGS CHARACTERISTIC

- a. Stop instant pressure data recording.

BLUETOOTH HEADER	FDC1004 MODULE	FDC1004 STOP DATA COLLECTION COMMAND	0	0	0	0	0	0	BLUETOOTH FOOTER
0XBA	0X09	0X0F	0	0	0	0	0	0	0XBB

* This function only stops pressure data being transferred. It does not stop other types of data collection methods going on.

- b. The Response Characteristic should read the following after the previous command:

0x00	0x00	0x00	0x01
------	------	------	------

- c. The previous command will be implemented. After implementation the following command, the response will be sent to indicate that the pressure data collection has finished.

0x00	0x00	0x00	0x0E
------	------	------	------

2. Write the following command to the SETTINGS CHARACTERISTIC

- a. Stop the internal timer that regulates the sampling rate.

BLUETOOTH HEADER	NRF52 MODULE	NRF52 RTC CLOCK COMMAND	NRF52 RTC SENSOR STOP	0	0	0	0	0	BLUETOOTH FOOTER
0XBA	0X01	0X04	0X10	0	0	0	0	0	0XBB

* Only call this function once to stop all timer enabled sensor data collection.

- b. The Response Characteristic should read the following after the previous command:

0x00	0x00	0x00	0x01
------	------	------	------

3. Write the following command to the SETTINGS CHARACTERISTIC

- a. Shutdown patch.

BLUETOOTH HEADER	NRF52 MODULE	NRF52 COMMON COMMAND	NRF52 VCC LDO DISABLE	0	0	0	0	0	BLUETOOTH FOOTER
0XBA	0X01	0X06	0X07	0	0	0	0	0	0XBB

- b. The response characteristic should read the following after the previous command:

0x00	0x00	0x00	0x01
------	------	------	------