

Channel Calibration Report

2/21/2012

Channel Name : Nozzle dP 0-5inWC

Calibration Date : 1/20/2012

Expiration Date : 1/20/2013

Calibration Enabled? : Yes

If the calibration is expired, apply calibration and return warning.

Sample Size : 1000

Sample Rate : 1200 Hz

Additional Information :

Reference	Uncalibrated	Calibrated	Reference - Calibrated
0.0000E+0	1.8524E-3		
2.0000E-1	1.5843E-1		
4.0000E-1	3.5293E-1		
8.0000E-1	7.5105E-1		
1.2000E+0	1.1553E+0		
1.6000E+0	1.5531E+0		
2.0000E+0	1.9495E+0		
5.0000E+0	4.5000E+0		
Units: inWC			

History :

1/20/2012 CH : Calibration values edited

1/20/2012 CH : Calibration values edited

1/20/2012 CH : Calibrated