

Graph Signal Processing and Interpolation for EEG Channel Reconstruction

Carlos Saldivia, *Student Member, IEEE*

Abstract—This paper studies EEG channel reconstruction under missing-channel conditions using a graph signal processing (GSP) benchmark with reproducible execution artifacts. Following the GSP perspective popularized by Ortega and collaborators, we model EEG electrodes as graph vertices and evaluate reconstruction quality through both spatial and temporal criteria. We compare geometric baselines, graph-based estimators, and graph-time regularization methods under scenario-driven masking protocols. The current B2 stage includes full-scale batched runs and consolidated publication tables with MAE, RMSE, DTW, and SNR statistics. Results indicate that graph-aware and TV/time methods remain consistently competitive, while ranking differences are scenario-dependent and therefore require variance-aware interpretation. We also report the current BGSRP status transparently: strong proxy performance in Python, with strict MATLAB/GSPBox 1:1 equivalence still pending.

Index Terms—EEG reconstruction, graph signal processing, interpolation, missing channels, temporal regularization

- A cross-family comparison including geometric baselines, graph-based interpolation, and temporal regularization methods.
- A publication-oriented artifact set with frozen configurations, raw outputs, and summary tables.

At the current B2/B3/B4 stage, we evaluate performance using MAE, RMSE, DTW, and SNR under scenario-driven masking conditions (4 scenarios, 10%, 20%, 30% y 40% missing) with repeated runs and consolidated variability statistics.

This is still an evolving manuscript, so we intentionally keep a transparent scope statement: most of the benchmark is consolidated, but strict BGSRP equivalence against the original MATLAB/GSPBox environment remains a documented limitation rather than a completed claim.

I. INTRODUCTION

Electroencephalography (EEG) recordings frequently exhibit missing channels due to contact issues, amplifier saturation, movement artifacts, or quality-control rejection. In practical studies, this is not a minor inconvenience: missing channels can distort spatial interpretations, degrade biomarkers, and reduce the reliability of downstream decoding pipelines. Because of that, channel interpolation has been studied for decades, from classical spatial interpolation to graph-based reconstruction.

Graph Signal Processing (GSP) provides a natural framework for this problem by representing electrodes as graph vertices and modeling inter-channel relationships through weighted edges. The general perspective is consistent with the graph-signal literature associated with Ortega and collaborators [?], [?], where smoothness, bandlimitedness, and spectral structure are used to replace purely geometric interpolation with graph-constrained estimation.

Even though many methods have been proposed in the literature [?], [?], [?], in practice it is still difficult to compare them fairly in a single reproducible pipeline. Different papers use different masks, datasets, and evaluation conventions, so conclusions are often hard to transfer to real EEG workflows. This thesis-paper track tries to reduce that gap with a unified benchmark and publication-ready artifacts.

The main contributions of this manuscript are:

- A unified and reproducible pipeline that covers graph construction, missing-channel simulation, reconstruction, and quantitative evaluation.