LABEL-BACKED VIDEO ANALYSIS

Seizure detection report

A concise, human-readable summary of one completed video run at the selected analysis threshold.

Recording	Sz2PG-sample
Dataset	VSViG-demo
Run ID	run-8c93f2a1
Processing	NVIDIA GeForce RTX 4070 Laptop GPU (CUDA)
Threshold	0.75 - exploratory, stored run unchanged

POSSIBLE EVENTS	MATCHED	RUNTIME
3	**1**	**2m 03s**
Model detections	Workbook timing	End-to-end

Interpretation

At threshold **0.75**, the model identified **3 possible seizure events**. One event matched the workbook timing and two were model-only. This threshold is exploratory and does not change the saved model run.

Workbook ground truth

Label field	Workbook value	How it is used
EEG onset	68.00 s	Evaluation boundary
Clinical onset	75.00 s	Event matching reference
Seizure end	Not provided	No duration was invented

Evaluation metrics

Metric	Value	Plain meaning
Sensitivity	0.745	Detected labelled seizure windows
Specificity	0.502	Rejected labelled non-seizure windows
Precision	0.320	Model-positive windows that were labelled positive
F1	0.448	Balance of sensitivity and precision
AUROC	0.739	Probability ranking across thresholds

SeizureAI RESEARCH PROTOTYPE - NOT A MEDICAL DEVICE

PREDICTION-ONLY VIDEO ANALYSIS

Seizure detection report

A report designed for runs without a workbook. It clearly separates model predictions from ground-truth evaluation.

Recording	home-monitoring-015
Dataset	pilot-uploads
Run ID	run-11a742c0
Processing	CPU - Intel Core i7
Threshold	0.50 - exploratory

POSSIBLE EVENTS	PEAK SCORE	RUNTIME
7	**91.2%**	**8m 41s**
Model detections	Highest window	End-to-end

Interpretation

The model identified **7 possible seizure events**. No workbook was supplied, so sensitivity, precision, F1 and other label-backed metrics cannot be calculated. These events require expert review.

Detected events

Event	Start	End	Duration	Peak probability
1	00:42.0	00:49.5	7.5 s	0.820
2	01:13.5	01:21.0	7.5 s	0.910
3	02:34.0	02:39.5	5.5 s	0.770
4	04:10.0	04:18.0	8.0 s	0.804
5-7	See appendix	-	-	3 additional events

Why metrics are unavailable: Accuracy-style metrics need verified ground-truth labels. This report intentionally does not treat model predictions as truth.

RUN COMPARISON

Threshold experiment report

Run A and Run B are recalculated from saved probabilities. Difference is Run A minus Run B.

Configuration	Run A	Run B
Recording	Sz2PG-sample	Sz2PG-sample
Model	Video v1.0	Video v1.0
Device	CUDA - RTX 4070	CPU - Intel i7
Threshold	0.75	0.30

Fair-comparison check	Compatible: same recording, model version and label mode. Threshold and processing device differ by design.

Comparison table

Measure	Run A	Run B	A - B	Meaning
Sensitivity	0.745	0.821	-0.076	Run A worsened
Specificity	0.702	0.502	+0.200	Run A improved
Precision	0.510	0.320	+0.190	Run A improved
F1	0.605	0.461	+0.144	Run A improved
False positives/hour	5.2	12.8	-7.6	Run A improved
Detected events	3	11	-8	Neutral
Total runtime	123.5 s	521.0 s	-397.5 s	Run A improved

Human-readable conclusion

Run A used a stricter threshold and produced fewer events, higher precision, higher specificity and fewer false positives per hour, but sensitivity decreased. CUDA processing completed approximately **4.2 times faster**. No single run is universally better; the preferred threshold depends on the validation objective.

Research-use notice: Outputs are experimental decision-support information, not a diagnosis. Review all detections and ground-truth alignment with qualified clinical and research personnel.