



AALBORG UNIVERSITY
STUDENT REPORT

Evaluation of Electrotactile Feedback Schemes in Combination with Myoelectric Prosthetic Control – Closing the Loop

Master Thesis
Biomedical Engineering & Informatics -
Spring 2019
Project group: 19gr10407

Christian Korfitz Mortensen, Martin Alexander Garenfeld



4th Semester, Project

School of Medicine and Health

Biomedical Engineering and Informatics

Fredrik Bajers Vej 7A

9220 Aalborg

Title:

Evaluation of Electrotactile Feedback Schemes in Combination with Myoelectric Prosthetic Control – Closing the Loop

Project period:

P3, Fall 2018

01/02/2019 - 6/6/2019

Project group:

10gr10407

Participants:

Christian Korfitz Mortensen
Martin Alexander Garenfeld

Abstract:

text

Supervisors:

Strahinja Dosen
Jakob Lund Dideriksen

Pages: ??

Appendix: ??

Handed in: 6/6/2019

The content of this report is freely available, but publication (with reference) may only be done with agreement with the authors.

Preface

Contents

Part I Paper	1
Part II Worksheets	11
1 Background	12
1.1 Sensory Feedback Stimulation	13
1.2 State of Art in Electrotactile Feedback	15
1.3 Closing the Loop	17
1.4 Feedback Stimulation Setup	18
1.5 Electromyography	19
1.6 Data Processing	22
1.7 Pattern Recognition	23
1.8 Proportional Control	25
1.9 Performance Evaluation	25
2 Study Objective	27
3 Methods	28
3.1 Study Design	28
3.2 Feedback Configurations	30
3.3 The Virtual Closed-Loop Prosthesis	33
3.4 Acquiring Control System Training Data	34
3.5 Data Processing	36
3.6 The Prosthetic Control System	39
3.7 Online Control Training and Test	40
3.8 Determination of Sensory Thresholds	42
3.9 Feedback Training	43
3.10 Combining Control and Sensory Feedback	45
4 Results	47
A Appendices	53
A.1 Experiment Protocol	53
A.2 Short Experiment Description	61

Part I

Paper

Evaluation of Electrotactile Feedback Schemes in Combination with Electromyographic Control

Christian Korfitz Mortensen and Martin Alexander Garenfeld

Abstract—The restoration of intuitive and meaningful proprioceptive feedback of myoelectric prosthetic state is an important task to enhance embodiment and user satisfaction, hence lowering the demand for visual attention for prosthetic control in everyday tasks. Therefore, two different configurations for conveying two DoF prosthetic state information of wrist rotation and hand opening through electrotactile feedback stimulation were developed and evaluated in a simulated closed-loop prosthesis. A spatially based configuration was made conveying information by changes the activation of pads in an electrode array placed circumferentially around the contra-lateral arm. The other, amplitude based, used various levels of amplitude to specific pads to convey information of the prosthetic state.

14 able-bodied subjects were trained and evaluated through a blinded target reaching test in using both feedback configurations following a minimal training session. The completion rate for visual feedback (99 %) significantly outperformed both the spatially based (87 %) and the amplitude based (93 %) configurations. The amplitude feedback configuration yielded a significantly higher completion rate ($p = 0.044$) than the spatially based and was also preferred by 64 % of the subjects. However, both feedback schemes were reported to be useful and intuitive manifesting that both configurations are ready to be tested in less discriminative cases.

Index Terms—Closed-Loop, myoelectric Prosthesis, electrotactile feedback, electrode array, prosthetic state, Plumbus.

I. INTRODUCTION

THE loss of an upper limb can be an incredibly traumatic and life-changing event with the consequence of a significantly reduced quality of life due to restrictions in function, sensation and appearance [1], [2]. In an effort to restore pre-trauma functionality, prosthetics of various functionality and complexity have been introduced to replace the missing limb [3]. However, despite advancements in prosthetic technologies 25% of users choose to abandon their myoelectric prosthetic device [4]. A major reason for the low user satisfaction is found in the lack of exteroceptive and proprioceptive feedback provided by commercially available devices [1], [5]. Presently, merely one commercially available device (VINCENT evolution 2, Vincent Systems GmbH, DE), provides the user with feedback information of grasping force through a feedback interface [6].

The missing sensory feedback can cause the prosthetic hand to feel more unnatural and awkward [7]. Furthermore, the user mainly relies on visual feedback [7], [8], which is a need prosthetic users have shown a strong desire to decrease in order to enhance easiness and naturalness of use [9]. In a survey by Peerdeeman et al. [5], it was found that secondly

Christian Korfitz Mortensen and Martin Alexander Garenfeld are both undergrad. students at the School of Medicine and Health at Aalborg University, Denmark (email: ckmo14@student.aau.dk, mgaren14@student.aau.dk).

to receiving proportional grasp force feedback, prosthetic positional state feedback was of the highest priority. Visual independence can be achieved by providing the user with proprioceptive information through somatosensory feedback. This might facilitate the prosthetic device to be adopted by the user as an integrated part of their body, enhancing the feeling of embodiment and restoring the once physiologically closed loop [8], [10]–[12].

Various means of recreating the sensory feedback has been sought through either invasive and non-invasive approaches that translate information from sensors in the prosthesis to new sensory sites. Invasive methods, termed somatotopical feedback, aim to recreate the localization of the prior sensory experience by directly stimulating the nerves, which conveyed that particular sensory modality in the lost limb. This is, however, a complicated solution and multiple aspects, like long term effect, have yet to be investigated. [1], [8]. Substitution feedback utilizes various tactors (pressure, vibrational, temperature, electrotactile, etc.) and their use can either be modality matched using e.g. pressure as a substitute for grasp force [13] or non-modality matched via e.g. vibration for grasp force [14], [15]. Electrotactile feedback uses small electrical currents to activate skin afferents eliciting sensory sensations, which can be modulated in multiple parameters such as pulse width, amplitude, and frequency to convey feedback information along with the possibility of using multiple feedback channels [12]. As commercially available upper-limb prosthetics have multiple degrees of freedom (DoF's) [16] the need for multiple feedback channels is present to accommodate the amount of information which needs to be provided in a meaningful way.

In cases where two information variables are being conveyed e.g. grasping force and hand aperture using frequency and amplitude modulation in electrotactile stimulation [17] or pulse interval and stimulation frequency in vibrotactile stimulation [18], results have shown that one stimulator is not sufficient for users to distinguish between two modalities. In 2014, Witteveen et al. [19] provided sensory feedback of grasping force and hand aperture through a single vibrator and an array of vibrotactile actuators, respectively. Results showed that identification of stiffness for four virtual objects was around 60 %. Although the percentage was rather low, the feedback configuration proved better compared to no feedback showing that multichannel feedback helps distinguishability when conveying feedback of more than one information variable. [19] However, the use of multiple vibrotactile actuators might be less feasible and practical to implement in prosthetics, due to their size and greater power consumption compared to electrotactile stimulation.

The flexibility of electrotactile stimulation makes it de-

sirable and its use has earlier been proven useful in cases of conveying force feedback from pressure sensors on a prosthetic hand or from sensors in artificial skin [20], [21]. However, the possibilities in electrotactile feedback have also been investigated with regards to communication information on states of a multi DoF prosthesis. Strbac et al. [11] presented a novel electrotactile feedback stimulation interface, which could be used to convey information about the current state of a multi-DoF prosthesis. The system was comprised of four different dynamic stimulation patterns communicating the states of four different DoF's through a 16 multi-pad array electrode. The state of three different DoF's were communicated by altering the electrodes activated in a specific pattern. The fourth pattern communicated grasp force by modulating the stimulation frequency. Tests of the stimulation design showed that six amputees were able to recognize the stimulation pattern of the four DoF's with an average accuracy of 86 %. [11] However, it was not tested how well these stimulation patterns was aiding the user when combined with prosthetic control.

To the authors' knowledge, no one has fully closed the neural afferent/efferent loop, when investigating the usability of electrotactile feedback for restoring proprioceptive aspects during the use of a myoelectric prosthesis. Furthermore, based on the multiple parameters that can be modulated in electrotactile feedback, the question of which parameters that are most useful to convey tactile information on prosthetic motion states, is still unanswered. This study will, therefore, investigate how different electrotactile feedback modalities support prosthetic control when conveying proprioceptive sensory feedback of prosthetic states using a stimulation setup. Two novel stimulation configurations that delivered feedback regarding motion states of a two DoF virtual myoelectric prosthesis were investigated; one based on spatial activation of differently located pads in an electrode array, and one based on modulating the current amplitude of the electrode pads.

In Section II a description of the two novel feedback configurations will be given. Subsequently, the study design, closed-loop prosthesis, control system and experimental protocol will be presented. Results of the experiment will be reported in section III. Finally, the significance of this study and its results will be presented in Sections IV and V.

II. METHODS

A. Feedback Configurations

The main objective of the study was to evaluate two novel electrotactile feedback configurations' intuitiveness in providing proprioceptive information of a two DoF myoelectric prosthesis.




Fig. 1. Image of the 16 multi-pad electrode array used for stimulation. It consisted of 16 circular cathode pads, which each shared a common anode.

The DoF's used were wrist rotation and closed hand, where each DoF was divided into four motion states, where a unique feedback was provided for the various states in each DoF. The electrode array used to deliver electrical stimulation can be seen in figure 1. The array electrode consisted of a single anode pad and 16 circular cathode pads of which the activation and current amplitude could be modulated individually. The pads comprised of conductive Ag/AgCl traces imprinted in a 150 μm thick polyester layer. All pads were covered with conductive hydrogel (AG702, Axelgaard, Denmark) to enhance skin-electrode contact. The array electrode was placed around the contralateral lower arm of the dominant arm such that the end pads had a maximum gap of three cm centrally on the posterior side, when using a pronated arm as reference position. The following sections will present the two developed feedback configurations.

1) *Spatial configuration:* The motivation behind the spatial configuration was to communicate wrist rotation by spatially rotate anteriorly placed active electrode pads and to communicate closed hand by narrowing the distance between posteriorly placed pads.




Fig. 2. Illustration of the developed spatial configuration. The levels written next to the pads pairs corresponded to the level of the prosthetic state; the higher the level, the higher the prosthetic state of the given movement was. For left-handed subjects supination and pronation were reversed.

This feedback design was chosen in order to intuitively mimic the directions of the motions in the included DoF's. An illustration of the spatial configuration can be seen in figure 2. The pads were divided two groups each responsible for conveying information about a single DoF. The anteriorly placed pads were allocated for wrist rotation and the posteriorly placed for closed hand. The pads were furthermore paired such that each pair would represent one of four motion states in one DoF. For wrist rotation the pads were connected in side by side pairs. For right-handed subjects the activation of pads pairs would rotate laterally when increasing motion states during supination and rotate medially during pronation. The reversed activation was apparent for left-handed subjects. For the closed hand DoF the pairs consisted of oppositely located pads on the medial and lateral sides. When increasing motion states the active pairs would move posteriorly and the distance between active pads would become shorter. When the prosthesis was in a combined DoF state, the pads pairs corresponding to the level of the prosthetic state of each DoF would be active. Thus, a maximum of eight pads could be active simultaneously.

2) *Amplitude configuration:* The incentive behind the amplitude configuration was to convey information about the increase of prosthetic states by gaining the amplitude in electrode pad groups of four.




Fig. 3. Illustration of the developed amplitude scheme. Different groups of four electrode pads were active during supination, pronation and closed hand, respectively. The amplitude of the active pads would increase with the increase of the prosthetic state; the higher the prosthetic state level the higher the current amplitude of the given pads. For left-handed subjects supination and pronation were reversed.

The areas of active pads allocated for the various motions was similar to the spatial configuration to intuitively resemble the prosthesis motions. An illustration of the amplitude configuration can be seen in figure 3. The eight most anteriorly placed pads were used for wrist rotation and the four most posteriorly placed pads for closed hand. The eight pads used during wrist rotation were split such that the four most laterally placed were used during supination and four most medially placed were used during pronation for right-handed subjects. The pad activation was reversed for left-handed subjects. As

the prosthetic state of a given movement would increase the the current amplitude in the pads corresponding to that movement would increase. When in combined DoF motion state, the pads corresponding to the level of the prosthetic state of each DoF would be active in the relative amplitude level. Thus, a maximum of eight pads could be active concurrently.

The choice of groups of four electrode pads was decided upon to exploit the highest number of pads in the electrode array, while maintaining a symmetric distribution of possible active pads.

B. Study Design

To investigate the usability of the developed feedback schemes in combination with control an experiment was conducted which evaluated the usefulness of the feedback schemes when eliminating visual dependency. For this purpose 14 able-bodied subjects (12 male and 2 female - 13 right-handed and 1 left-handed with a mean age of 26.1 ± 2.4) were recruited. Included subjects signed an informed consent form and meet inclusion criteria stated in the experimental protocol, which was ethically approved by the ethical committee of Region Nordjylland, Denmark (approval number N-20150075). Each subject was introduced, trained and finally evaluated in understanding both the spatially based scheme and the amplitude based scheme. However, the order of which feedback scheme the subject would be trained/tested in was randomized. Figure 4 illustrates the chronological flow of stages in the experiment, where the first block focused on developing a subject specific prosthetic control system and the second block focused on training and evaluating the use of the sensory feedback schemes.

During the first block, EMG data was initially acquired and used to train a control system, which was used for a simulate prosthetic control. Subsequently, was a stage where the subject was made familiar with the control system. Finally, the achieved prosthetic control was evaluated through a target reaching test. Afterwards, a series of subjects sensory thresholds were determined for use of conveying electrotactile feedback. The subject then began the stages of familiarizing and training with a feedback scheme followed by re-familiarization of control in combination with feedback. Finally, a evaluation test of using the sensory feedback in combination for control was made. The entire sensory feedback block was then repeated using the remaining feedback scheme.

C. Virtual Closed-Loop Prosthesis

Investigating the usability of the two sensory configurations in a closed-loop scenario required these to be interfaced with a prosthetic device, which accommodated the actuation of rotational and open/closed hand DoF's. However, using an actual prosthesis might result in auditory feedback being provided to the subject through prosthetic actuation sounds, eliminating the interest of solely exploring the impact of sensory feedback. Hence, it was chosen to simulate a velocity based virtual prosthesis which enabled evaluation of the developed feedback schemes. In figure 5 is a depiction of a grid system and a black




Fig. 4. Pipeline showing the stages of the experiment. The stages in the first block focused on developing and evaluating a subject specific simulated prosthetic control system. Then electrotactile sensory thresholds were determined. The second block focused on training the understanding of the feedback schemes and evaluating their use in combination with prosthetic control. The sensory feedback block was repeated for the remaining feedback scheme.

cursor symbolizing the different possible prosthetic states and the current prosthetic state, respectively.

would be provided corresponding to the stimulation pattern for each scheme. In the neutral position (current location of cursor in figure 5), no sensory feedback was provided.




Fig. 5. Image of the grid map and cursor used in the experiment. Wrist supination moved the cursor to the right, pronation moved it to the left and closing the hand moved it downwards. For left-handed subjects, the rotational movements were reversed. Opening the hand moved the cursor upwards, and was used as a correction movement if needed.

Each square corresponded to a prosthetic state of either a single DoF or combinations of two DoF's. Performing supination would make the cursor move to the right and to the left when performing pronation. Performing closed hand would make the cursor move downwards and upwards when performing open hand. Finally, performing rest (relaxing the arm), would make the cursor stand still. Furthermore, the contraction intensity was made proportional with the actuation speed, enabling the subject to have greater control of cursor movement speed. The control was sequential which only enabled the cursor to move in one DoF at a time. When the cursor entered a square a specific electrotactile stimulation

D. Data Acquisition

In order for a subject to be able to control the virtual prosthesis online, the prosthetic control system needed to be trained with previously acquired EMG signals. For EMG data acquisition the Myo Armband (MYB) from Thalmic labs was used, which contained eight dry stainless steel electrode pairs embedded on the inside of the armband. Furthermore, it could communicate wirelessly to external devices via a Bluetooth 4.0 unit, making it a highly practical recording device with minimum preparation time needed. However, it had a fixed sample rate of 200 Hz with the exclusive analogue filter being a 50 Hz notch filter, thus, making the EMG signals prone to aliasing. A study by Mendez et al. [22] showed, however, a similar mean classification accuracy of nine hand gestures in a LDA classifier, when comparing data acquired with electrodes that covered the entire EMG spectrum and the MYB acquired data. This justified the use of the MYB and only a 10 Hz cut-off second order Butterworth high-pass filter was implemented digitally to remove low frequency artefacts.

To account for the delay until steady state motions were reached, it was desired to train the prosthetic control system with both transient and steady state EMG data from each movement [23]. To achieve this, the subjects were to follow a trapezoidal trajectory, where they controlled a cursor that moved horizontally with time and vertically with EMG intensity. The recording was 11 seconds, where the trajectory had an incline/decline of three seconds and a plateau of five seconds, representing the transient and steady state, respectively. However, only data from the last second of the incline and first seconds of the decline was used to train the classifier to avoid active motion classes being misclassified with rest. The trajectory and cursor position was scaled relative to an initially recorded prolonged maximum voluntary contraction (pMVC) of 15 seconds, which was set to 1. Data was acquired from

three recordings per movement, where the plateau was 40 %, 50 % and 70 % of the pMVC's, respectively. A last recording of 15 seconds rest was also performed.

E. Feature Extraction

Features were extracted from the acquired EMG signals to expand the amount of information used to train the classifier in the prosthetic control system. Due to the risk of the EMG signals being aliased, features representing frequency content might lose fidelity. In a study by Donovan et al. [24], the classification accuracy of space-domain features exploiting the relationship between EMG signals from neighbouring electrode pairs in the MYB were compared with the commonly used Hudgins features in a LDA classifier. Here, the use of space-domain features yielded a 5 % higher accuracy than the Hudgins features with EMG data acquired from the MYB. It was therefore chosen to use the five non-redundant features derived by Donovan et al., along with the Hudgins feature Waveform Length to represent frequency content indirectly [25], resulting in a total of six features. Both offline and online features were extracted in windows of 200 ms with a 50 % overlap to obtain quick update time, while preserving robust classification accuracy [26].

F. Prosthetic Control System

The extracted features were used to train a sequential proportional control system. For sequential control, a LDA classifier was used and for proportional control multiple linear regression models were used. The following section will address the fitting of the control system.

1) Classification model: A feature set was calculated for each of the eight electrode pairs and subsequently concatenated resulting in a 48-dimensional feature matrix that was provided to the classifier. It was chosen to implement a LDA classifier due to it being quick to train, while still yielding robust control [27]. The LDA classifier determined decision boundaries by maximizing the distance between centroids of the movement class feature values. Such decision boundaries were defined as a linear combination of feature values, where the output was posterior probabilities for each movement class. The decision rule was that the movement class with the highest probability would decide the determined motor function. The classifier was trained in distinguishing between five classes: wrist supination, wrist pronation, closed hand, opened hand and rest.

2) Proportional control: The proportional control model provided the control system with an actuation velocity proportional to the contraction intensity in a direction based on which movement class that was decided on by the classifier. This was achieved by training four multiple linear regression models: one for each active movement class. The mean absolute value (MAV) was calculated in windows for all electrode pairs and provided to the regression model as independent values, where the MAV scaled relative to the pMVC was provided as dependent values. During online control, the output was limited to a maximum value of 1; corresponding to the intensity of the pMVC. The distance of 1 corresponded to

1 cm on the computer screen, thus, the maximum speed of the virtual prosthesis was 1 cm per update (100 ms). A full DoF would be completed from one extremity to another in two seconds, thus, achieving an actuation velocity similar to the commercially available Bebionic prosthesis [28]. A second restriction implemented was that a movement had to be performed with >15 % contraction intensity, for the virtual prosthesis to be actuated. This was included to get a more stable performance at rest.

G. Subject Control Training and Evaluation

The subjects were initially trained in controlling the virtual prosthesis via visual feedback. It was crucial for subjects to achieve robust control for the feedback configurations to be able to be evaluated in a closed-loop prosthetic control system. The subjects' control abilities were assessed empirically during trainings and quantitatively through a Fitts' Law inspired target reaching test. If a subject did not have a completion rate above 90 % and a mean time to reach a target at below 10 seconds the subject would be excluded.

The subject control training was divided into two runs of three minutes with a different prosthesis representation in each training. In the first training, the prosthesis was represented as a black cursor as seen in figure 5. Here, the cursor position would move continuously with each control system output. In the second training, the cursor was invisible, and the prosthesis representation was instead the square containing the cursor being highlighted. This discretized prosthesis representation was implemented to equalize the visual and sensory feedback. This prosthesis representation was used in the remaining training/test runs with visual feedback. During both training phases the subjects were instructed in practicing the ability to move the cursor in a desired direction and to transition from movement to rest.

During the target reaching test, the subjects had to reach targets (highlighted grid squares) visualized in a randomized order. The subjects had to match the discretized virtual prosthesis with the target and dwell in that position for 1.5 seconds for it to be deemed reached. The subjects had 30 seconds to reach a target. If either a target was reached or the time limit was reached, the virtual prosthesis would reset in neutral position. The test was finished when all grid squares had been highlighted, making a total of 24 targets.

H. Determination of Electrotactile Sensory Thresholds

Providing meaningful sensory feedback required determination of four distinguishable subject specific electrotactile thresholds. A multichannel electrotactile stimulation feedback device (MaxSens, TecNALIA, Spain) generating biphasic pulses was connected to a standard desktop PC for individual control of pad activation. Thresholds were made solely amplitude dependent by keeping pulse width and frequency constant at 500 μ s and 50 Hz, respectively. 1st level thresholds, termed perception thresholds, were determined for each pad by initializing the amplitude at 0 μ A and then increase it in steps of 100 μ A per second. The subject was instructed to report when stimulation could be perceived confidently.

Subsequently, the intensities were readjusted by comparing the sensation intensity in neighboring pads to achieve homogeneous sensations across all pads.

4th level thresholds termed tolerance threshold, were set using the same approach, however, the amplitude level was initialized at the perception threshold and increased in steps in steps of 200 μA per second. The thresholds were determined when the subject reported that the sensation was on the onset of getting unpleasant, the stimulation was becoming functional or a maximum of 10,000 μA was reached. Intensities were again readjusted to achieve homogeneous sensations. Throughout the process of determining thresholds, the subject was faced away from the screen to avoid bias from observing the visual increase of amplitude levels. Intermediate threshold levels 2 and 3 were calculated for the i^{th} pad based on the perception and tolerance threshold as:

$$2 \text{ lvl}(i) = \text{percep.}(i) + \left(\frac{1}{3} \cdot (\text{tol.}(i) - \text{percep.}(i)) \right) \quad (1)$$

$$3 \text{ lvl}(i) = \text{tol.}(i) - \left(\frac{1}{3} \cdot (\text{tol.}(i) - \text{percep.}(i)) \right) \quad (2)$$

I. Sensory Feedback Training

Following the subject training of prosthetic control, the subject was trained in understanding a sensory feedback scheme. The sensory feedback training was divided into two phases: familiarization and reinforced learning.

The familiarization phase provided the subjects with a short and controlled introduction to the scheme. The cursor was moved by the investigator from the neutral position to a designated state incorporating the transition from one square to the next, thus, presenting the subject with the coherence between states and stimulation patterns for a designated state. States in the top and bottom row and the middle column were presented actively, while the remaining 12 states were presented indirectly as transition states when moving to combinational states of 4th level closed hand and either 1st or 2nd level of wrist pronation and supination by first moving in the rotational DoF. Time spend in designated states was approximately four seconds and time spend in transition states was approximately two seconds. Recognition of single DoF states was assessed to be most crucial for comprehension, hence, these were favored in the familiarization phase.

For the reinforced learning phase the subject was asked to face away from the screen. The cursor was directed to a designated state and the subject then had to report what the current state was based solely on the sensory feedback. If the subject answered correctly, the cursor was reset to the neutral position and the cursor was moved to a new target. If the subject answered incorrectly, the correct state would be communicated for the subject to learn from, before continuing. Each state would be presented once and be moved to by taking the optimal path (move the cursor fully in one DoF before the other). However, which DoF the cursor would move in first was varied. Hence, the subject could utilize the transitions made when guessing the current state. The order of the

designated states was predetermined by the investigators. Time spend in transition states was approximately two seconds. When all 24 states had been trained, the subject was given a short break before repeating the reinforced learning. However, the order and paths were changed for the second run.

J. Closed-Loop Evaluation

The subject had until this point trained the prosthetic control and sensory feedback separately. The motor function and sensory feedback was combined in a closed-loop prosthetic system. During this final evaluation test the visual feedback regarding the prosthetic state was eliminated. The subject then had to rely on the sensory feedback to assess the prosthetic state.

Before undergoing the test the subject was given a three minute training period to get reacquainted with the prosthetic control and to further train the understanding of feedback scheme. The evaluation test was similar to the evaluation test presented in section II-G, besides that the virtual prosthesis was not visualized. Thus, the subject had to solely rely on sensory feedback, when reaching a highlighted state. The evaluation test was performed two times consecutively.

K. Statistical Analyses

The metrics extracted from the evaluation tests were number of reached targets, mean time spend per target and mean distance travelled per target. Paired comparisons were made between results from evaluation tests. Due to the low sample size, comparisons was made using non-parametric statistics. Wilcoxon signed rank test was applied as comparisons was made on related samples obtained from a two block study design. A significance level of 5 % was used.

III. RESULTS

A. Raw Trajectory Examples

The evaluation metrics were extracted from the performance of the target reaching evaluation test. In figure 6 two examples of trajectories, which subjects took, can be seen. The top figure shows the ideal trajectory that one subject took when moving to a two DoF state. This indicates that the subject had comprehended the feedback and knew exactly when the designated state was reached. The bottom figure in figure 6 is an example of a subject traveling to far down before moving to the right. However, this was corrected for as the subject still managed to reach the target.

B. Evaluation Metrics

The evaluation metrics from first to second evaluation test in each feedback scheme did not yield significant difference ($p > 0.05$). Therefore, it was chosen to view these evaluation tests as one, by calculating the mean between the first and second evaluation test in both blocks, resulting in a single test result for the spatial feedback and a test single result for the amplitude feedback.

Figure 7 shows box plots of the extracted metrics for the visual, spatial and amplitude feedback evaluation tests.




Fig. 6. Examples of two path taken when moving the cursor to a target in the final evaluation test. The top figure show the ideal trajectory that a subject took when moving to a two DoF state. The bottom figure gives an example of a subject traveling to far down before moving to the right. However, this was corrected for an the subject still managed to reach the target. The blue line is the path, red square is the target state and the green dot is the direct cursor output.

Worth noting was that visual feedback still outperformed solely relying on electrotactile feedback both when spatially or amplitude modulated for the completion rate and time to reach a target metrics. When comparing the spatial and amplitude feedback tests, using amplitude feedback yielded a slightly higher completion rate ($p = 0.044$). This quantitative result was also supported by the subjective opinion of the subjects as 64 % of the subjects favored the amplitude feedback. However, all subjects struggled in choosing a favored feedback scheme as they found both intuitive to understand.

C. Target State Hit Distribution

When observing the completion rate of the specific targets in figure 8, it can be derived that targets generally had a higher hit rate when using amplitude feedback compared to spatial (Mean completion rate with spatial feedback was 87 % and amplitude was 93 %). However, common for both feedback

scheme was that the centered targets were more troublesome to reach (Mean completion rate for centered targets was 76 % with spatial feedback and 88 % with amplitude feedback; mean completion rate for peripheral targets was 93 % with spatial feedback and 97 % with amplitude feedback). A possible reason for this finding is that the subjects had to achieve complete rest to dwell inside these targets.

In the peripheral targets, the subjects did not necessarily need to achieve complete rest, as they could continue performing a movement and still be on the boundary of the target. Furthermore, combined DoF targets generally had a lower completion rate for the spatial feedback scheme (Mean completion rate for single DoF targets was 90 % with spatial feedback and 93 % with amplitude feedback; mean completion rate for combined DoF targets was 85 % with spatial feedback and 93 % with amplitude feedback). This could indicate that the sensory feedback regarding combined prosthetic states in the spatial feedback scheme was harder to interpret than in the amplitude feedback scheme.

Spatial hit distribution				
100 %	96 %		100 %	96 %
93 %	79 %	86 %	86 %	100 %
89 %	75 %	82 %	75 %	89 %
79 %	57 %	64 %	82 %	82 %
96 %	89 %	93 %	93 %	96 %

100 %	100 %		93 %	96 %
100 %	86 %	89 %	93 %	100 %
93 %	86 %	93 %	86 %	96 %
93 %	89 %	82 %	86 %	93 %
100 %	96 %	93 %	96 %	100 %

Fig. 8. Hit rate for each target in the spatial and amplitude evaluation test, respectively. The more transparent a target is, the higher the hit rate was. 100 % accounts for a total of 28 hits for each test.




Fig. 7. Box plots of the metrics extracted from the visual, spatial and amplitude feedback evaluation tests. The two evaluation tests in the spatial and amplitude feedback block, respectively, were combined by calculating the mean between the two tests.

IV. DISCUSSION

Two intuitive electrotactile feedback schemes were developed for a two DoF velocity-based virtual prosthesis, one spatially modulated and one amplitude modulated. The schemes were integrated in an easy implementable 16 pad electrode array and tested in combination with sequential proportional myoelectric control. Unique sensory feedback was provided for four levels of prosthetic states in single DoF's and for 16 prosthetic states representing combined DoF's. The objective was to investigate the usability of the developed feedback schemes when removing visual dependency.

V. CONCLUSION

APPENDIX

ACKNOWLEDGEMENT

The authors would like to thank supervisors Strahinja Dosen and Jakob Lund Dideriksen for providing constructive feedback, and the School of Medicine and Health at Aalborg University for providing equipment and the facilities to complete this study. Additionally, the authors are very thankful for all the voluntary participants.

REFERENCES

- [1] J. S. Schofield, K. R. Evans, J. P. Carey, and J. S. Hebert, "Applications of sensory feedback in motorized upper extremity prostheses: A review," *Expert Review of Medical Devices*, vol. 11, no. 5, pp. 499–511, 2014.
- [2] K. Østlie, P. Magnus, O. H. Skjeldal, B. Garfelt, and K. Tambs, "Mental health and satisfaction with life among upper limb amputees: A Norwegian population-based survey comparing adult acquired major upper limb amputees with a control group," *Disability and Rehabilitation*, vol. 33, no. 17-18, pp. 1594–1607, 2011.
- [3] P. Geethanjali, "Myoelectric control of prosthetic hands : state-of-the-art review," *Medical Devices: Evidence and Research*, pp. 247–255, 2016.
- [4] E. Biddiss and T. Chau, "Upper limb prosthesis use and abandonment: A survey of the last 25 years," *Prosthetics and Orthotics International*, vol. 31, no. 3, pp. 236–257, 2007.
- [5] B. Peerdeman, H. Hermens, S. Stramigioli, H. Rietman, H. Witteveen, R. Huis in 't Veld, S. Misra, P. Veltink, and D. Boere, "Myoelectric forearm prostheses: State of the art from a user-centered perspective," *The Journal of Rehabilitation Research and Development*, vol. 48, no. 6, pp. 719–738, 2011.
- [6] V. Systems, "VINCENT EVOLUTION 2," <https://vincentsystems.de/en/prosthetics/vincent-evolution-2/>, 2005.
- [7] D. Pamungkas and K. Ward, "Electro-tactile feedback system for a prosthetic hand," *22nd Annual International Conference on Mechatronics and Machine Vision in Practice, M2VIP 2015*, pp. 27–38, 2015.
- [8] B. Stephens-Fripp, G. Alici, and R. Mutlu, "A review of non-invasive sensory feedback methods for transradial prosthetic hands," *IEEE Access*, vol. 6, pp. 6878–6899, 2018.
- [9] D. J. Atkins, D. C. Heard, and W. H. Donovan, "Epidemiologic Overview of Individuals with Upper-Limb Loss and Their Reported Research Priorities," *Journal of Prosthetics and Orthotics*, vol. 8, pp. 1–11, 1996.
- [10] H. Xu, D. Zhang, J. C. Huegel, W. Xu, and X. Zhu, "Effects of Different Tactile Feedback on Myoelectric Closed-Loop Control for Grasping Based on Electrotactile Stimulation," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 24, no. 8, pp. 827–836, 2016.
- [11] M. Štrbac, M. Belić, M. Isaković, V. Kojić, G. Bijelić, I. Popović, M. Radotić, S. Došen, M. Marković, D. Farina, and T. Keller, "Integrated and flexible multichannel interface for electrotactile stimulation," *Journal of Neural Engineering*, vol. 13, no. 4, pp. 1–16, 2016.
- [12] B. Geng, K. Yoshida, L. Petrini, and W. Jensen, "Evaluation of sensation evoked by electrocutaneous stimulation on forearm in nondisabled subjects," *The Journal of Rehabilitation Research and Development*, vol. 49, no. 2, p. 297, 2012.
- [13] S. B. Godfrey, M. Bianchi, A. Bicchi, and M. Santello, "Influence of Force Feedback on Grasp Force Modulation in Prosthetic Applications: a Preliminary Study," *Conf Proc IEEE Eng Med Biol Soc.*, pp. 1–8, 2016.
- [14] A. Ninu, S. Dosen, S. Muceli, F. Rattay, H. Dietl, and D. Farina, "Closed-loop control of grasping with a myoelectric hand prosthesis: Which are the relevant feedback variables for force control?", *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 22, no. 5, pp. 1041–1052, 2014.
- [15] M. Nabeel, K. Aqeel, M. N. Ashraf, M. I. Awan, and M. Khurram, "Vibrotactile stimulation for 3D printed prosthetic hand," *2016 2nd International Conference on Robotics and Artificial Intelligence, ICRAI 2016*, pp. 202–207, 2016.
- [16] F. Cordella, A. L. Ciancio, R. Sacchetti, A. Davalli, A. G. Cutti, E. Guglielmelli, and L. Zollo, "Literature Review on Needs of Upper Limb Prosthesis Users," *Frontiers in Neuroscience*, vol. 10, no. May, pp. 1–14, 2016.
- [17] R. E. Prior, J. Lyman, P. A. Case, and C. M. Scott, "Supplemental Sensory Feedback for the VA/NU Myoelectric Hand Background and Preliminary Designs," *Bulletin of Prosthetics Research*, vol. 101, no. 134, 1976.
- [18] A. Chatterjee, P. Chaubey, J. Martin, and N. V. Thakor, "Quantifying prosthesis control improvements using a vibrotactile representation of grip force," *2008 IEEE Region 5 Conference*, pp. 1–5, 2008.

- [19] H. J. Witteveen, F. Luft, J. S. Rietman, and P. H. Veltink, "Stiffness feedback for myoelectric forearm prostheses using vibrotactile stimulation," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 22, no. 1, pp. 53–61, 2014.
- [20] C. Hartmann, J. Linde, S. Dosen, D. Farina, L. Seminara, L. Pinna, M. Valle, and M. Capurro, "Towards prosthetic systems providing comprehensive tactile feedback for utility and embodiment," *IEEE - 2014 Biomedical Circuits and Systems Conference, BioCAS 2014 - Proceedings*, pp. 620–623, 2014.
- [21] M. Franceschi, L. Seminara, L. Pinna, S. Dosen, D. Farina, and M. Valle, "Preliminary evaluation of the tactile feedback system based on artificial skin and electrotactile stimulation," *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society, EMBS*, pp. 4554–4557, 2015.
- [22] I. Mendez, B. W. Hansen, C. M. Grabow, E. J. L. Smedegaard, N. B. Skogberg, X. J. Uth, A. Bruhn, B. Geng, and E. N. Kamavuako, "Evaluation of the Myo Armband for the Classification of hand motions," *International Conference on Rehabilitation Robotics*, pp. 1211–1214, 2017.
- [23] A. Boschmann, B. Nofen, and M. Platzner, "Improving Transient State Myoelectric Signal Recognition in Hand Movement Classification using Gyroscopes," *2013 35th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, pp. 6035–6038, 2013.
- [24] I. M. Donovan, J. Puchin, K. Okada, and X. Zhang, "Simple space-domain features for low-resolution sEMG pattern recognition," *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society, EMBS*, pp. 62–65, 2017.
- [25] B. Hudgins, P. Parker, and R. Scott, "A new strategy for multifunction myoelectric control," *IEEE Transactions on Biomedical Engineering*, vol. 40, no. 1, pp. 82–94, 1993.
- [26] R. Menon, H. Lakany, G. Di Caterina, B. A. Conway, L. Petropoulakis, and J. J. Soraghan, "Study on Interaction Between Temporal and Spatial Information in Classification of EMG Signals for Myoelectric Prostheses," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 25, no. 10, pp. 1832–1842, 2017.
- [27] K. Englehart and B. Hudgins, "A robust, real-time control scheme for multifunction myoelectric control," *IEEE transactions on bio-medical engineering*, vol. 50, no. 7, pp. 848–854, 2003.
- [28] J. T. Belter, J. L. Segil, A. M. Dollar, and R. F. Weir, "Mechanical design and performance specifications of anthropomorphic prosthetic hands: a review," *Journal of rehabilitation research and development*, vol. 50, no. 5, pp. 599–618, 2013.

Part II

Worksheets

1 | Background

The background chapter will outline the considerations that needs to be made when testing the usability of sensory feedback configurations in combination with myoelectric prosthetic control. The feedback will be given based on which motion state a pattern recognition controlled prosthesis is in.

The main idea behind myoelectric prosthetic control is to translate recorded muscle signals (EMG signals) into a motion performed by the prosthesis. Often, if possible, EMG is recorded from the muscle which were used to perform movements with the natural hand and used for prosthetic control. A pattern recognition model can be trained to differentiate between a set of movement classes. When receiving a segmented part of a EMG signal it then decides upon which movement class that most likely is being performed. In combination with the elicited muscle contraction level, this is used as input in the control system and the prosthesis should perform a corresponding motion. [1] In a closed loop prosthesis, the motion state the prosthesis is in can be coded to be equivalent to a certain sensory feedback. This should enable the user to interpret the sensory feedback and use as additional information to visual feedback about the prosthesis' state. [2] A closed loop prosthesis iteration can be seen in figure 1.1.

Regarding control the background chapter will explain the following: generation of EMG signals, data acquisition, data processing, pattern recognition and proportional control. Regarding sensory feedback the following will be explained: types of sensory feedback, prior investigations on sensory feedback and sensory feedback configurations.




Figure 1.1: The figure shows the stages of a closed loop prosthesis. First, EMG signals are recorded from the user. The signals are decoded and an output is relayed to the control system, which is used for the prosthesis to perform a motion. The motion state is then read and sensory feedback is delivered to the user regarding which motion state the prosthesis is in.

1.1 Sensory Feedback Stimulation

It has been known for some time that vision alone does not provide a sufficient amount information to achieve efficient daily life use of a prosthetic device, as the use requires full visual attention. In the effort of regaining the cutaneous sensations previously felt by the lost limb of a transradial amputee, stimulations of various sorts can be applied on the skin of the remaining stump. These stimulations mimic the information sensed by the lost limb by activating cutaneous receptors. It has been shown both in the acute phase and long term that providing the amputee with sensory feedback can reorganize neurological pathways or even recover original pathways during motor tasks, when trained amply. [3] Hence, efforts have been put in investigating methods of providing proprioceptive and exteroceptive information of i.e. grasp strength and prosthetic state through the means of artificial stimulation. [4, 5] Presently, there are multiple ways of providing the user with a variety of sensory feedback. These can be divided into three categories: Somatotopically feedback, modality matched feedback and substitution feedback. [4]

This section will present general terms in sensory feedback stimulation and give a brief overview of the types of sensory feedback in order to give insight in the possibilities and eventual disadvantages when providing the user of a prosthetic device with feedback.

1.1.1 Somatotopically Feedback

Somatotopically feedback aims to provide the user with a sensory experience which is perceived as natural as what was felt by their missing limb, both in location and sensation. To achieve such an experience, somatotopically feedback uses invasive approaches by making use of invasive neural electrodes and targeted reinnervation. The former is known as peripheral nerve stimulation and relies on the invasive neural electrodes being interfaced with the original neural pathways preserved proximally on the residual limb. Currently, two different types of electrodes have been exploited: One where a cuff is placed around a nerve fascicle and another where an electrode is implanted into the nerve fiber. But to this date, none of these methods have been comprehensively studied. Targeted reinnervation also enable the possibility of stimulating the original neural pathways from the missing limb. The corresponding sensory afferents are relocated to innervate new sites which can selectively be chosen and stimulated by non-invasive tactors. Somatotopically matched feedback is hypothesized to reduce the users cognitive burden due to its 'naturalness', facilitating increased compliance and less conscience attention. [4]

1.1.2 Modality Matched Feedback

In modality matched feedback, the type of sensory experience which would have been felt by the missing limb is communicated to the user. For instance, when pressure is felt in the palm of a prosthetic hand by pressure sensors, a proportional amount of pressure is delivered to the user somewhere on the skin. Thus, the sensation is not matched in location, but only in sensation. Mechanotactile feedback which conveys pressure information is utilized by the use of i.e. pressure cuffs or servomotors. These types of tactors are very useful for modality matched feedback, but have a disadvantage by being more power consuming compared to other stimulation types. [4, 6]

1.1.3 Substitution Feedback

Substitution feedback methods convey information about the state of the prosthesis without regarding the type of sensation and location which would have been felt by the missing limb. Thereby, the sensory information is said to be non-physiologically representative. The feedback methods are often straightforward to implement, but demands a greater amount of user adaption to interpret what the feedback information represents. Often used methods for substitution feedback are vibrotactile and electrotactile feedback. [4, 6]

Vibrotactile Stimulation

Vibrotactile stimulation utilizes small mechanical vibrators to convey information to a selected area of the skin which activates cutaneous mechanoreceptors. This method is most often used to transfer tactile information in prosthetic grasping tasks. [4] A recognizable sensation is evoked using frequencies between 10 and 500 Hz. The sensory threshold varies between users and location, resulting in the need for specific user threshold calibration.

[6]

Electrotactile Stimulation

In electrotactile feedback a sensory sensation is achieved by stimulating the primary myelinated afferent nerves with an electrical current. This sensation which the stimulation invokes has been reported to be both tingling, prickling, itching, buzzing, touch and burning pain. [4] Electrotactile stimulation rely on small and lightweight electrodes to provide the electrical stimulation. When compared to other feedback methods as vibrational and pressure stimulation, which depend on heavier actuators and moving parts to provide the feedback, these properties can be seen as a drawback as prosthetic users strongly desire lightweight systems [5, 7]. Furthermore, through the use of electrotactile stimulation, multiple factors such as amplitude, pulse width, frequency and location of the stimulation can be controlled facilitating development of agile feedback schemes. This enables the possibility of varying the perceived feedback as either vibration, tapping or touch by modulating the signal waveform. The downside of using electrodes is the requirement for recalibration of sensory thresholds, pulse width and frequency to reproduce the same perceived stimulation every time the electrodes are placed on the user. In addition, interference between electrodes used for stimulation and recording have been found to result in noise in recorded EMG-signal used for myoelectric control. Concentric electrodes are able to limit the interference by limiting the spread of current. Concentric electrodes have also been found to increase localization and perceptibility of the induced stimuli. [4, 5, 6]

1.2 State of Art in Electrotactile Feedback

Section 1.1 presented different types of sensory feedback from which the choice of stimulation in this project can be drawn upon. Somatotopically might restore the most natural sensations, but is also the most complicated. Modality matching the feedback should instead be sought, however present tactors are larger and more power consuming than electrodes uses in electrotactile feedback. Furthermore, the dimensions of electrodes facilitates easier integration with the prosthesis as these can be placed inside the socket, along with electrodes used for acquisition. However, this requires that a solution for leakage current is found. Modulating pulse width, frequency and amplitude in electrotactile feedback gives more possibilities for conveying complex tactile information. Therefore, the state of art methods using electrotactile sensory feedback in the current literature have been reviewed and will presented to ensure that the later derived feedback schemes extends recent evidence.

Multiple studies have investigated the use of electrotactile feedback regarding both how distinguishable sensations are evoked and how to convey sensory feedback in different coding schemes for improving myoelectric prosthetic control [5]. In 2015, Shi and Shen [8] investigated how subjects would perceive the effects of varying amplitude, frequency and pulse width of an electrical stimulation in various combinations. Results showed that appropriate sensations from electrical stimulation would be achieved by varying amplitude from 0.3 mA to 3 mA, pulse width from 0.1 ms to 20 ms and frequency from 40

Hz to 70 Hz. Furthermore, varying these ranges properly would make it possible to have proportionally increased stimulation grades felt by the subject. Additionally, the authors stated the importance of electrode size, as stimulation through to big or to small electrode diameters could result in sensations of pain or discomfort. [8]

Several studies [9, 10, 11, 12] using electrical stimulation have investigated its use in conveying grasping force/pressure feedback. Jorgovanovic et al.[11] investigated users' recognition of grip strength, when controlling a joystick controlled robotic hand, through varying the pulse width and keeping the frequency and intensity constant at 100 Hz and 3 mA, respectively. Results showed that providing electrotactile feedback improved the users' ability to move objects with the robotic hand. [11] Similar result were found by Isakovic et al. [12], who also showed that electrotactile feedback supported a faster learning than no feedback in grasp force control, and that electrotactile feedback might facilitate short-term learning.

A study by Xu et al. [10] tested and evaluated different types of pressure and slip information feedback through electrotactile stimulation and compared this to visual feedback and no feedback. The study recruited 12 subjects, 6 able bodied, and provided electrotactile feedback by keeping the intensity and frequency constant and then varying the pulse width between 0 μ s and 500 μ s indicating changes in grasp force. In this case, visual feedback was found to outperform electrotactile feedback. [10]

Pamumgkas et al. [9] also tested the use of electrotactile feedback to convey information from pressure sensors located in a robotic had. Their setup used six feedback channels corresponding to a pressure sensor in each of the fingers and one in the palm. Pressure information in the sensors were given in three discretized frequency levels of 100 Hz, 60 Hz and 30 Hz for the fingers and 20 Hz for the palm. Reported results stated that the subjects learned how to appropriately use the feedback when picking up objects of various sizes. Furthermore, the subjects reported that they preferred having electrotactile feedback accompanied by visual feedback opposed to only having visual feedback. [9] The purpose of restoring the sensation that would be experienced by touch of the skin has also been pursued in more elaborate efforts through artificial skin [13, 14]. In these cases, a grid of 64 pressure sensors were used to translate information of touch into 32 electrotactile electrodes placed on the arm of the subjects.

The use of electrotactile feedback has proven useful in cases of restoring the haptic feedback through pressure sensors on a prosthetic hand or by the touch on artificial skin. However, the possibilities of electrotactile feedback have also been investigated in the case of improving prosthetic control. In 2016, Strbac et al. [2] presented a novel electrotactile feedback stimulation system, which could be used to convey information about the current state of a multi-DoF prosthesis. The system comprised of four different dynamic stimulation patters communicating the states of four different DoF's through a 16 multi-pad array electrode, possibly restoring both proprioception and force. The state of the three of the DoF's were communicated by altering the electrodes activated in patterned fashion and the fourth DoF by modulating the stimulation frequency. Tests of the stimulation design showed that six amputees were able to recognize the four DoF's with an average accuracy of 86 %. [2]

In summary most studies have focused on using electrotactile feedback for exteroceptive

means while only [2] have investigated its use for proprioceptive feedback. However, their results encourage further investigation into how electrotactile feedback can be utilized for providing meaningful proprioceptive feedback.

1.2.1 Sensory Adaptation in Electrotactile Feedback

Before implementing a electrotactile feedback interface, it is important to consider the effect electric stimulation might impose on the sensory system.

Adaption is defined as a changing sensory response to a constant stimulus, and all sensory systems have shown adaptive tendencies. This could result in unreliable effects during prolonged electric stimulation. Hence, it is crucial to consider stimulation parameters which reduce adaption. Sensory adaption usually occurs within minutes, and reaches a maximum after 15 min. Furthermore, the adaption rate is related to the stimulation amplitude as adaption occurs faster when closer to the pain threshold. Low frequencies (<10 Hz) show less adaption compared to higher frequencies (>1000 Hz). The adaption response is found to be exponential in decay and recovery. [15, 16] However, sensory adaptation can be overcome by using intermittent stimulation, and preferably, stimulation interfaces should consider conveying feedback information through diversified patterns [16, 17].

Developed feedback schemes should consider using as low amplitudes as possible to reduce the rate of sensory adaption. Furthermore, continuously changing the site of stimulation should also facilitate less adaption.

1.3 Closing the Loop

The loss of a limb does not only result in loss of motor function as sensory function also gets impaired. Providing an amputee with a prosthetic device, which does not provide sensory feedback, only restores one half of the once closed limb control loop. To close the loop the prosthetic device needs to contain proprioceptive and exteroceptive sensors, whose recorded information should be conveyed to the amputee in a intuitive and meaningful way [18]. This can be achieved using methods of sensory substitution mentioned in section 1.1.3.

Closing the loop is a well recognized need amongst prosthetic users and using substitutional sensory feedback holds the possibility of lowering need for visual attention to track correct prosthetic movement. This might furthermore improve easiness of use and embodiment, which could lower rejection rates. [2] However, the advantages of closing the loop by providing sensory substitution feedback have been contradictory [11]. In 2008, Cipriani et al. [19] investigated the use of vibrotactile feedback for improving prosthesis grasp function and did not find any improvement when providing the sensory feedback. Later findings by Witteveen et al. [20] disproved this as they found that when providing information of grasp force and slip through vibrotactile feedback improved a virtual grasping task.

Even though studies like [11, 20] found closing the loop by providing grasp force sensory

feedback helpful (review by Stephens-Fripp et al.) [5], currently only one commercial feedback providing device, the VINCENT evolution 2 (Vincent Systems GmbH, DE) is available [21]. However, no devices have yet to implement means of proprioceptive feedback. Additionally, closed loop control systems bypassing human interaction have also been investigated and implemented by commercial manufacturers i.e. Otto Bock and RSL steeper. Actuators are made to autonomously adjust grip force based on sensors located in the prosthetic hand, thereby not involving the user in the final execution of the task. [10] Such an approach might improve reliability of the prosthesis, but does not provide proprioceptive and exteroceptive feedback to the user, hence not promoting embodiment.

1.4 Feedback Stimulation Setup

To elicit electrotactile stimulation in this project, the MaxSens stimulation device will be used along with a 16 multi-pad electrode. The following section will provide a short overview of the stimulation device and multi-pad electrode specifications.

1.4.1 Stimulation Electrode

The 1×16 multi-pad stimulation electrode, can be seen in figure 1.2. It is made of 16 circular cathodes, which each share a common long anode. The electrode consists of a polyester layer, an Ag/AgCl conductive layer and a insulation coating. The electrode to skin contact is improved by applying conductive hydrogel pads to the electrode pads. [2]




Figure 1.2: The 16 multi-pad electrode used for stimulation consists of 16 circular cathode pads, which each share a common anode.

1.4.2 MaxSens Stimulation Device

The stimulation device is made by MaxSens, Tecnalia, San Sebastian, Spain. Communication between PC and the stimulation device can be achieved either through Bluetooth or USB serial connection. The device can be controlled through a series of commands.

The MaxSens device allows for independent control of the 16 pads in the electrode. It generates biphasic stimulation pulses where the pulse width can be controlled within a 50 - 1000 μs range with 10 μs steps, frequency ranges from 1 - 400 Hz with 1 Hz steps and current amplitude ranges from 50 - 10000 μA with 0.1 μA steps. Whereas current amplitude and pulse width can be controlled independently for each pad, the pad frequency is set globally limiting all pads to have same frequency.

1.5 Electromyography

The control of a myoelectric prosthesis is based on recorded myoelectric signals. [22] Enabling the use of myoelectric signals for control of functional prosthetics requires a theoretical background knowledge of the signals origin and how it can be acquired. The following section will describe myoelectric signals and how they are acquired through the acquisition method of electromyography (EMG).

The process of executing a voluntary movement can be explained through electric potentials and the excitability of skeletal muscle fibers. The nerve impulse carrying excitation information of a voluntary muscle contraction will travel from the motor cortex down the spinal cord to a alpha motor neuron. The alpha motor neuron will activate and direct an nerve impulse along its axon to multiple motor endplates, which each innervate a muscle fiber. The motor neuron and the muscle fibers it innervates is in collection called a motor unit. [23]

The nerve impulse initiates the release of neurotransmitters forming an endplate potential. The muscle fibers consist of muscle cells, which each are surrounded by a semi-permeable membrane. The resting potential over the membrane is held at a equilibrium, typically at -80 mV to -90 mV, by ion pumps, which passively and actively control the flow of ions through the membrane. The release of neurotransmitters affects the flow through the ion pumps resulting in a greater influx of Na^+ . This results in a depolarization of the cell membrane. However, only if the influx of Na^+ is great enough to create a depolarization surpassing a certain threshold, an action potential is formed. The action potential is characterized by the cell membrane potential, which changes from around -80 mV to +30 mV. The created action potential will propagate in both directions on the surface of the muscle fiber. This process happens across all muscle fibers in a motor unit. The action potential is also known as a motor unit action potential (MUAP), and it is the superposition of multiple MUAPs that is recorded through surface EMG. The action potential is still measurable on the skin surface, however, some limitations in surface EMG are that the recording is restricted to superficial muscles, that the amplitude of the EMG signal is affected by the depth of subcutaneous tissue and that the distinguishability of MUAP's from adjacent muscles is unreliable. [23, 24]

Acquisition of EMG-signal can either be carried out through surface EMG or intramuscular EMG. The latter measures MUAPs through needles inserted into the muscle and can and collect MUAPs from single muscle fibers individually. Surface EMG is acquired through electrodes on the skin surface. [25] Using surface EMG requires preparation of the skin surface to minimize impedance and maximize skin contact. Hence, the skin should be clean and dry before electrode placement. To further minimize skin-electrode impedance

removal of excess body hair or flaky skin and cleansing the area using alcohol swabs should be considered. [23, 25] In this project MUAPs will be recorded through surface EMG. An example of a surface EMG recording of two different movements (pronation and supination of the wrist) can be seen in figure 1.3. Here, the surface electrodes are placed at the circumference of the forearm of the subject. It can be seen that some electrode channels are more or less active when comparing the two movements. This corresponds to different muscles being more or less contracted depending on which movement that is performed. This enables the recognition of which movement is being performed. A prerequisite for this to work is that the electrode placement must be identical throughout the recording. If not, the activation of the various channels shifts spatially, and the accuracy of the trained recognition system might be decreased.




Figure 1.3: Illustration of an eight electrode channel surface EMG of the forearm during pronation (left side) and supination (right side) of the wrist. The recording was acquired using the Myo Armband, see figure 1.4. The 4th channel was placed on the thickest part of the forearm centrally on the anterior side, when using a pronated arm as reference position, with the horizontal LED light faced laterally towards the wrist.

1.5.1 Data Acquisition

Before a user can utilize a myoelectric prosthesis the control system needs to be taught how certain movements look like represented as EMG signals. This process is called training the control system. The acquisition of training data from the user is therefore the first step in training the control system.

Chapter 1. Background

In the acquisition of EMG signals the Myo armband (MYB) from Thalmic Labs will be used. It contains eight dry stainless-steel electrode pairs placed inside the armband. The advantage of using dry electrodes is that they do not need to be disposed after use, in contrary to conventional gel electrodes. Thus, the MYB can be reused for all subjects participating in the project, which enables less time consuming experiments. An additional usability advantage is that it communicates wirelessly to external devices via Bluetooth 4.0, leaving no loose wires to possibly limit mobility or distort connection.

The MYB acquires EMG signals in an 8-bit resolution. Instead of acquiring the signal in millivolts, the output is scaled to decimal numbers between -1 and 1. However, the amplitude of the EMG signal output is still proportional to muscle contraction intensity. To avoid signal frequencies from the power grid to interfere with the EMG signal, an analogue 50 Hz notch filter is built in the MYB. This is, however, the only analogue filter implemented in the MYB, and as it has a sample rate of 200 Hz, which is inside the EMG spectrum (10-500 Hz), the acquired EMG signal will likely be aliased. The implementation of a digital anti-aliasing filter would therefore be a irrelevant task. However, a comparison study showed that using the MYB in a Linear Discriminant Analysis (LDA) control scheme can archive similar performance accuracy compared to using conventional gel electrodes with a sample rate of 1000 Hz [26]. Additionally, the MYB contains a 9 axes inertial measurement unit, but will not be utilized in this project and will therefore not be further elaborated on.

During initialization of the MYB the user has to follow two calibration steps: the warm up and the synchronization. In the warm-up step, the MYB is establishing a strong electrical connection between the skin and the armband, which reduces skin-electrode impedance and enables the electrodes to transduce properly. This happens as the user's skin becomes more moist from light sweating, which works similar to the gel in conventional EMG electrodes. During the synchronization step the MYB determines its orientation in space, its position and on which arm it is placed, based on a wrist extension movement the user must perform. The MYB works most optimally when tightly fit. To ensure a close fit, a set of clips can be used if necessary.



Figure 1.4: Image of the Myo armband from Thalmic Labs. Electrode channel 1 corresponds to the first output in the recording and electrode channel 2 as the second etc., as seen in figure 1.3.

1.6 Data Processing

In order to use the acquired data most optimally in the myoelectric prosthetic control scheme the data must be processed. In this processing, undesired frequencies are filtered out and features that represents the data are extracted from segments of the data in order to obtain more information about the movement than what is only contained in the raw EMG signal. This data processing will be covered in the following sections.

1.6.1 Filtering

To remove unwanted frequencies from the EMG signal, it should be filtered. According to the Nyquist Theorem, the rate the signal is sampled with must be at least twice the highest frequency contained in the signal to archive a non-aliased digital recording. However, as mentioned in section 1.5.1, the MYB samples with a rate lower than the highest frequency in the EMG spectrum, without having any analogue bandpass filter implemented. The rationale behind incorporating a digital anti-aliasing filter is therefore defeated. Implementing a digital high-pass filter with a corner frequency at 10 Hz to remove low frequency miscellaneous biological noise would, however, be desirable. [25]

1.6.2 Segmentation

The extraction of features are done in discretely segmented windows of data, instead of calculating the features from instantaneous values. In online control, the length of windows is a compromise between classification accuracy and delay in prosthetic control. Often an window overlap is implemented. This is a technique applied to ensure short delays, while still enabling a high classification accuracy. When applying an overlap values

from the previous window is reused in the current window. The amount of overlap chosen is significant for the performance of the control scheme. Generally, it is recommended to have window lengths of 150-250 ms and use a 50 % overlap [27]. Choosing a large overlap will result in short delays, but worse classification accuracy and vice versa. When using the MYB it is important to take the low sample rate into consideration, as a window will contain less data compared to if the sampling was appropriate to the EMG frequency properties. [27] Short windows will therefore likely result in worse classification accuracy compared to appropriately sampled data segmented when using identical window length.

1.6.3 Feature Extraction

Instead of only utilizing the raw EMG signal in a control scheme, features are extracted to exploit more representations of the EMG signal that optimally results in robust control. Various independent features can be extracted from the signal either from the time domain, frequency domain or the time-frequency domain. Most commonly features from the frequency and time domain are used. When extracting frequency domain features it is required for the EMG signal to be transformed into the frequency domain. This takes more computation time compared to extracting features directly from the time domain. For this reason features in the time domain are usually favoured. [28] Especially used are the Hudgins features: Mean Absolute Value (MAV), Zero Crossings (ZC), Slope Sign Changes (SSC) and Waveform Length (WL) [29]. However, both ZC and SSC represent the frequency content of the signal, which most likely has been distorted by the low sample rate. When using the MYB for EMG acquisition an alternative set of features has been suggested by Donovan et al. to extract from the data [30]. These features are so called space domain features, since they exploit the relationship between the output from the electrode channels. When evaluating data acquired from the MYB the space domain features increased classification accuracy by 5 % in a LDA-based control scheme compared to using the Hudgins features [30]. A final consideration to make when choosing features is to avoid redundancy as they then would not provide additional information about the signal.

1.7 Pattern Recognition

For a myoelectric prosthesis to know which movement to perform, it needs to know how to differentiate between the movements it will be trained to perform. For this purpose classification is a commonly applied model. The classification model, or classifier, is fed known data consisting of features extracted from the raw EMG signals, which were recorded while the user was performing different movements. If each of the known feature data sets related to each movement is known they can be labelled appropriately, and the classifier will then learn which data represents which movement. Each label is known as a class and the process of labelling the data is called supervised learning. The known data is also called training data, hence this process is called training the classifier. If the classifier is trained properly, it is able to categorize unknown data accurately into the

correct class. This is what happens online in each segmented data window when using a pattern recognition-based myoelectric prosthesis. The classifier is, however, only able to categorize unknown data into one of the trained classes. [31]

A frequently used supervised classifier for myoelectric prosthetic control is the LDA classifier. An advantage of using LDA is that it enables robust control, while having a low computational cost [32]. LDA will be used in this project to determine motor function and an overview of the theory behind LDA will be given in the following section.

1.7.1 Linear Discriminant Analysis

LDA determines decision boundaries between the desired number of classes, where the distance between the decision boundary and the centroid of the class feature values is maximized. Such a decision boundary is defined as a linear combination of the feature values x :

$$g_j(x) = \text{weight}_j x + \text{bias}_j \quad (1.1)$$

where weight_j decides the orientation of the decision boundary of class j , and bias_j is a bias that decides the position of the decision boundary of class j in relation to origo.

The decision rule of a LDA classifier is based on which class that has the highest probability of having produced the input feature values; also called the posterior probability. Given this decision rule, LDA can be derived from the Bayes theorem, which expresses the posterior probability as:

$$P(\omega_j|x) = P(x|\omega_j)P(\omega_j) \quad (1.2)$$

where $P(x|\omega_j)$ is the class conditional probability, the probability that a feature value from class j appears, and $P(\omega_j)$ is the prior probability, the probability that class j appears. This can be written as the function:

$$g_j(x) = P(x|\omega_j)P(\omega_j) \quad (1.3)$$

An constraint in LDA is that each class is Gaussian distributed and all classes share the same covariance matrix. The class conditional probability can therefore be written as the multivariate normal distribution, in which the class conditional covariance matrix can be written as the common covariance matrix. This leaves the following function:

$$g_j(x) = \mu_j \Sigma^{-1} x' - \frac{1}{2} \mu_j \Sigma^{-1} \mu_j' - \ln(P(\omega_j)) \quad (1.4)$$

where μ_j and Σ^{-1} are the mean vector for class j and the common covariance matrix, respectively. The function in equation (1.4) can be written in the common linear discriminant classifier form as in equation (1.1). [33]

Thus, a posterior probability is calculated for each class based on the decision boundaries, and according to the decision rule, the class with the highest probability of having produced the input feature values will be chosen as the determined motor function. However, LDA only determines the movement class, but does not enable control of the motion speed the actuator must perform. For this purpose an additional control scheme must be applied to activate the determined motor function in a proportional matter. [34]

1.8 Proportional Control

After the motor function has been determined, a mapping of a control output needs to be performed. The advantage of providing a continuous output to the actuator proportional to the contraction intensity compared to a one-speed controller is that the user has the possibility of grasping objects quickly, while still being able to perform more slow and dexterous tasks. Additionally, proportional control resembles the human neuromotor system, which makes it more intuitive. [34]

A widely used proportional control scheme is linear regression [34]. Here, a dependent output value can be calculated based on a function of an independent input value. In the case of using several electrode channels as when using the MYB, the output needs to be computed based on several independent values. For this purpose multivariate linear regression would be appropriate:

$$\hat{Y} = \alpha + \beta_1 X_1 + \beta_2 X_2 + \cdots + \beta_i X_i + \epsilon_i \quad (1.5)$$

where \hat{Y} is the control output and X_i is the independent input values, where the index i will correspond to the number of electrode channels in the MYB. α and β are the estimated value of \hat{Y} at $X = 0$ and estimated regression coefficients, respectively. The absolute values of the recorded EMG signals can be used directly as the independents input value in such a proportional control scheme. [35] However, a regression model needs to be estimated for each motor function in the control system. Then the appropriate regression model will be selected based on the classification output.

1.9 Performance Evaluation

Evaluating the performance of a derived control system can be achieved through the completion of various tasks. If available, the system can be interfaced with a myoelectric prosthesis, and based on the completion of tasks mimicking daily life functionality (e.g. grasp and movement of objects), performance can be evaluated [36]. Otherwise, virtual environments have been widely used showing movements of virtual prostheses [37] or by moving a cursor to targets resembling motor function, where performance can be

quantified through measurements based on Fitts' Law [33, 38, 39]. An obvious measure to observe is the completion rate (CR), which is the ratio of reached targets compared to the total number of targets. This describes the overall ability the user has when using the control system. Path efficiency (PE) can be used to observe how efficiently continuous movement control is achieved by comparing the distance travelled to reach a target and comparing it to the most direct route. To observe how well the user can keep the system at rest and control velocity, stopping distance (SD) and overshoot (OS) can be measured. The former measures the distance travelled at times where no movement is intended, and the latter tracks the number of times the user reaches a shown target, but leaves before completion. [33]

2 | Study Objective

In summary, there is still a need for myoelectric prosthetic devices to fully close the neural loop by providing amputees with proprioceptive feedback to lower the need for visual attention. As presented in section 1.2 most studies have focused on providing exteroceptive feedback while only very few studies have investigated how proprioceptive information could be conveyed to aid prosthetic control in cases where visual attention is less wanted. Using the modality of electrotactile stimulation as a mean of transferring information of the prosthetic state offers multiple stimulation parameters which can be modulated through several channels enabling possibilities for intuitive and meaningful sensory feedback. However, even though several opportunities present themselves in modulating the stimulation amplitude, frequency and active channels, it would be of great interest to investigate which would lead the sensory feedback to be perceived most intuitively. As stated in section 1.4 the frequency cannot be controlled individually for each pad in the electrode, thus a feedback scheme modulating frequency will no be investigated in this study.

Investigating whether spatially coded or amplitude coded information assists the most when neglecting visual attention will provide insight into which parameters future configurations should encapsulates. This leaves the following study objective:

Test and evaluate two novel stimulation schemes, one based on modulating amplitude and one based on spatial localization of activation, for conveying sensory feedback of the prosthesis state in a closed loop prosthetic control system.

3 | Methods

Applying the knowledge acquired regarding myoelectric prosthetic control and feedback stimulation, the following section will document the implementation of the two novel feedback configurations, and the remaining requirements needed to test the usability of these in a closed-loop system. The methods chapter will contain sections documenting the implementation of study design, feedback configurations, data acquisition, data processing, fitting of prosthetic control system, validation of subjects' control abilities, feedback configuration training, feedback configuration evaluation and statistical analyses.

3.1 Study Design

In order to investigate whether amplitude or spatially based electrotactile feedback aids prosthetic control the most when removing the visual dependency, an experiment was set up. A feedback coding scheme based on spatial activation and a feedback scheme based on amplitude modulation was developed and will be presented later in section 3.2. 14 able bodied subjects (13 right handed) were recruited, where the order of which feedback scheme the subjects would be trained/tested in was randomized. However, an equal number of subjects was assigned to each order. An overview of the subject population and demographics can be seen in table 3.1.

Table 3.1: Overview of subject population and demographics.

	Age, mean(std)	Gender(n)	
		Female	Male
Total (n = 14)	26.1(2.4)	2	14
Order 1 (n = 7)	25.7(2.3)	1	6
Order 2 (n = 7)	26.4(2.6)	1	6

Prior to enrollment, the subjects were assessed to meet the inclusion criteria stated in the experimental protocol, which can be found in section A.1. The subjects were handed a short experiment description prior to the experiment session, which gave an introduction to the background of the study and the different task the subjects would have to go through. Upon enrollment, the subjects were asked to sign an informed consent form, which stated that the subjects had received adequate information about the experiment and that they were always able to withdraw from the experiment. The experiment was

Chapter 3. Methods

ethically approved by the ethic committee of Region Nordjylland, Denmark (Approval number N-20150075).

The experiment was designed such that each subject was trained and tested in using both feedback schemes along with control during a single session experiment. A graphical illustration of the main stages that the subject went through can be seen in figure 3.1. For all subjects data used to build the prosthetic motor control system was acquired first. Secondly, the subjects were given time to familiarize with the control system and subsequently, the achieved control was evaluated through a target reaching test. Next, sensory thresholds used for feedback were determined for the subject. Subjects assigned to order 1 went through four steps of training and test using the spatial scheme followed by the same four steps using the amplitude scheme. Subjects assigned to order 2 went through the schemes in opposite order. The next sections will further document the implementation and execution of the experiment.




Figure 3.1: Pipeline showing the stages of the experiment. Firstly, the stages common for all subjects were carried out followed by the order dependent stages.

3.2 Feedback Configurations

The fundamental interest of this project was to develop two novel, intuitive and useful feedback schemes to convey proprioceptive information through electrotactile stimulation and evaluate which one would aid control the most. This was to be implemented in a two DoF virtual prosthesis using wrist rotation and closed hand as motions, where four levels of sensory feedback would be provided for each DoF to communicate prosthetic motion states. The following section will present the two developed spatial and amplitude feedback schemes. All anatomical directions are explained with the hand being pronated

as reference.

3.2.1 Spatial Configuration

The spatial feedback scheme was created with the interest of achieving an intuitive way to convey the feedback by focusing stimulation to localized regions on the skin. An illustration of the spatial feedback scheme can be seen in figure 3.2. The idea behind this configuration was to communicate wrist rotation movements by rotating the activated electrode pads and to convey the closed hand movement by narrowing distance between active electrodes.

To archive this, the electrode pads were broken into two groups: one upper and one lower. The upper eight pads (5-12) were used to convey information of the rotational DoF using four pads for either side. The four pads were divided into pairs of two, where the first from the center in e.g. pronation would be activated when the cursor entered the first level prosthetic state. Likewise, the second level pair would be activated when the cursor entered the second level. Hence, during a transition from neutral to level one and level one to level two, the subject should feel the stimulation moving either laterally or medially depending on the prosthetic state. Finally, this meant that when the virtual prosthesis was in a supinated state the stimulation should be felt in the anterior lateral region of the arm, while when the virtual prosthesis was in a pronated state stimulation should be felt in the anterior medial region of the arm.

The lower eight pads (1-4 and 13-16) were used to convey information about the closed hand DoF. The pads were paired in an opposite manner for each of the four levels respectively: 4 and 13; 3 and 14; 2 and 15; 1 and 16. As the hand would move to one closed hand state out of four possible, a certain pad pair would activate. Transitioning from neutral to fully closed, the subject should sense the stimulation converging on the posterior side of the arm. The virtual prosthesis was able to be in states, which was a combination of the two DoFs. In these cases the feedback would be a combination of any upper and lower pair, resulting in the activation of four pads simultaneously.




Figure 3.2: Illustration of the developed spatial scheme, which was based on different pads being activated depending on the level of the prosthetic state. The level assigned to the various electrode pairs corresponded to levels of prosthesis states in figure 3.4. The highest number of possibly activated pads was four at a time. For left handed subjects the rotational states were reversed.

3.2.2 Amplitude Configuration

In this scheme, attention to the recognition of stimulation localization should be less. Instead, the subject would have to focus on discriminating between the intensity of stimulation in the regions which were active.

Compared to the spatial configuration where the feedback was given through dynamically changing the pads activated, the amplitude configuration instead conveyed feedback in three greater regions and solely modulated the amplitude of the stimulation. The upper eight pads were again used for the rotational DoF, illustrated in figure 3.3. Pads 5-8 were activated with different amplitude levels at pronated prosthetic states 1 and 2, respectively. The same was applicable for supinated states using pads 9-12.

For conveying information about closed hand pads 1,2,15 and 16 were used. When in a combined DoF prosthetic state, eight electrodes would be active in amplitude levels relative to the level of the prosthetic state. It was chosen to use groups of four electrode pads to exploit the largest number of pads in the electrode array, while still maintaining a symmetric distribution of possible active pads.




Figure 3.3: Illustration of the developed amplitude scheme. Here, the amplitude of the active pads would increase with the increase of the prosthetic state; the higher the level the higher the current amplitude of the given electrodes. The level of amplitude strength assigned to the various electrodes corresponded to levels of prosthesis state in figure 3.4. The highest number of possibly activated pads was eight at a time. For left handed subjects the rotational states were reversed.

3.3 The Virtual Closed-Loop Prosthesis

In order to test the usability of the two sensory feedback configurations in a closed-loop control system, a prosthesis which accommodates this was simulated. As no commercial or research prostheses were available in this project, a virtual system resembling prosthetic control was made. Using a virtual prosthesis also had the benefit of providing no sounds that might indicate the prosthesis state during evaluation tests.

The aim was to develop a system which could provide control and feedback of two degrees of freedom. For this purpose, it was chosen to incorporate the rotational degree of freedom in wrist supination and pronation and a grasp degree of freedom in open and closed hand. In figure 3.4 is a depiction of a grid system and a black cursor symbolizing the different possible prosthetic states and the current prosthetic state, respectively, where each square corresponded to a prosthetic state. Performing supination would make the cursor move to the right into one of two possible states, while performing pronation would make the cursor move left into one of two states. Performing the closed hand movement would make the cursor move downwards into one of four possible states and performing open hand would make the cursor move upwards. In total, the prosthesis could achieve a total of 25 different states, which represented single DoF movements or combinations of two DoFs. However, as the control was sequential it was only possible to move the cursor in a single DoF at a time along an axis. In each of the squares, a unique electrotactile

feedback was provided in each of the two feedback configurations, as explained in section 3.2, thus ascribing four states for each DoF.




Figure 3.4: Image of the grid map and cursor used in the experiment. Performing supination moved the cursor to the right, pronation moved it to the left and closing the hand moved it downwards. For left handed subjects the rotational movements were reversed. Opening the hand moved the cursor upwards, and was used as a correction movement if needed.

3.4 Acquiring Control System Training Data

As presented in section 1.7, in order for a prosthetic control system to be able to differentiate between movements, a classifier could be trained with EMG data acquired while performing each movement. Therefore, individual subject EMG training data had to be acquired. Training data was acquired using the MYB placed around the thickest part of the dominant forearm while the subject performed wrist pronation, wrist supination, open hand, closed hand and rest. The subsequent section will document how the data used for training the classifier was acquired.

First, a baseline recording was made, where the subject was instructed in keeping the hand perfectly still. The baseline consisted of a 15 second recording and was subtracted

Chapter 3. Methods

from each of the other recordings to reduce baseline noise. If the signal was below the baseline amplitude it was set to zero.

During a muscle contraction two main states can be recognized: a transient state, described by inconsistent myoelectric activity as the muscle length is changed, and steady state, where a constant muscle activity is reached. [40] Classification is often based solely on steady state data, however, including transient state might make for a more robust classifier as the delay until steady state is reached is eliminated [41, 42].




Figure 3.5: The trapezoidal plot (left) and contraction validation plot (right) used during acquisition of the training data. The trapezoidal plot represented the contraction amplitude requested. The black cursor moved continuous horizontally with time, and the height of the cursor position indicated the currently elicited contraction intensity. The validation plot was used to assess the homogeneity of the performed movement. The mean of a segment in each electrode channel was plotted in different directions to get an overview of the activation recorded from the individual channels. The shape of the validation plot should be similar during a recording for it to be homogeneous.

To feed the classifier with training data representing muscle contractions with varying force, different fractions of maximum contraction force were recorded. In the process of obtaining training data for each movement, the same four contraction were carried out: a prolonged maximum voluntary contraction (pMVC) recording and 40 %, 50 % and 70 % of the pMVC recordings. The pMVC was recorded for 15 seconds where the subject was instructed to elicit the contraction with a maximum contraction force which could be held steady, over the course of the 15 seconds. This resulted in an pMVC for each channel in the MYB, which was calculated as the mean of the absolute values of the EMG signal for each channel. The mean for each channel was used as a maximum reference when acquiring the subsequent fraction recordings.

Acquisition of the 40 %, 50 % and 70 % fraction of pMVC were done using a graphical user interface (GUI) made in MATLAB 2018b, which can be seen in figure 3.5. The image shows the trapezoidal trajectory the subject were instructed to follow using the black cursor, where the height of the cursor was calculated as the mean absolute intensity across all channels of the elicited muscle contraction. The data were extracted over a 200 ms window of the EMG signal. The cursor would automatically move positively along the x-axis in relation to time. The high plateau of the trapezoid represented either the 40 %, 50 % or 70 % fractions of the pMVC. Data were recorded during 2.5 s rest periods in the beginning and end, a 2.5 incline transition, 5 s steady state and 2.5 second decline transition, summing to a total time of 15 seconds. However, only data recorded during the steady state and the last and first second of the incline and decline of the transition phase, respectively, were used to train the classifier.

The additional plot, seen on the right in figure 3.5, plotted the amplitude of each of the eight channels in the MYB and was used by the investigators to assess whether the performed movements were done correctly. If the amplitude of the channels responsible for the performed movement shifted rapidly, or if channels not responsible for the performed movement were active, it would indicate that the subject did not perform a pure contraction and the recording would have to be redone.

3.5 Data Processing

The following sections will cover which filtering, segmentation and feature extraction solutions that were decided to implement, based on the background information presented in section 1.6.

3.5.1 Filtering

Due to the EMG-bandwidth being 10-500 Hz and taking the MYB specifications into consideration the only interest was to remove low-frequency artefact noise. Hence, a 2nd order Butterworth high-pass filter with a cut-off at 10 Hz was implemented. The order of the filter was chosen, as fast update time was highly desired in the online prosthetic control, and a higher order might have slowed the update due to a longer computation time. The choice of implementing a Butterworth filter was due to the desire of minimizing phase shift inside the EMG bandwidth, as this could distort the fidelity of some of the extracted features.

3.5.2 Segmentation

In online myoelectric prosthetic control, quick update time is important to maintain naturalness in the prosthesis motion, while still ensuring robust classification. A windowing of 200 ms with 50 % overlap was therefore chosen. This would update the prosthesis motion state every 100 ms and segment 40 samples per window to feed the classifier. In initial tests, this proved adequate to yield smooth and reliable prosthesis motions, when

using the MYB for EMG acquisition.

3.5.3 Feature Extraction

For this project it was decided to extract the space domain features recommended by Donovan et al. [30], due to the increased classification accuracy obtained compared to using Hudgins features when applying the MYB for data acquisition. The features formulated in [30] were MAV, Mean MAV (MMAV), Scaled MAV (SMAV), Correlation Coefficient (CC), Mean Absolute Difference Normalized (MADN), MAD Raw (MADR) and Scaled MADR (SMADR). Additionally, it was decided to extract the Hudgins feature, WL, to exploit frequency related information of the signal in the classification. All these features will be explained in the following text.

MAV is a commonly used feature to represent information on muscle contraction intensity and how much force a subject needs to produce to perform a movement at a given intensity. Its changes are linearly proportional with contraction intensity; the more intense the contraction is the higher the feature value will be and vice versa. For one window in the i^{th} channel, MAV is calculated as:

$$MAV_i = \frac{\sum_{n=1}^{ws} |x_i[n]|}{ws} \quad (3.1)$$

where $x_i[n]$ denotes the n^{th} raw sample from channel i and ws denotes window size or number of samples in one window.

Scaling MAV with the mean of MAV across all channels will remove the dependency of specific movement intensity - some movements produce higher mean intensities than others at the same fraction of the MVC. The average of MAV across all channels is denoted MMAV and is calculated as:

$$MMAV = \frac{\sum_{i=1}^8 MAV_i}{8} \quad (3.2)$$

MAV scaled by MMAV is denoted SMAV and is calculated as follows for each window in the i^{th} channel:

$$SMAV_i = \frac{MAV_i}{MMAV} \quad (3.3)$$

Each EMG channel in the MYB records a mixture of sources. Some individual sources can affect multiple channels, which will increase the correlation between channels, while other more local sources might only affect a single channel, which decreases the correlation. To

represent the correlation between channel i and the neighbouring channel $i + 1$, Donovan et al. proposed the calculation of a correlation coefficient (CC), which is expressed as:

$$CC_i = \frac{\sum_{n=1}^{ws} X_i[n]X_{i+1}[n]}{ws} \quad (3.4)$$

where $X_i[n]$ is the n^{th} sample from channel i in one window after the sample has been normalized. The normalization is done by subtracting the mean of raw samples from each samples followed by dividing the resulting values with their standard deviation.

In an effort to further represent the relationship between channels, the mean of the absolute value of the difference between normalized channel values was calculated. This is referred to as mean absolute difference normalized (MADN), and is expressed as:

$$MADN_i = \frac{\sum_{n=1}^{ws} |X_i[n] - X_{i+1}[n]|}{ws} \quad (3.5)$$

To decrease the computational denseness of first calculating the normalized values as done with CC and MADN, MAD was calculated using just the raw samples. This is referred to as MADR and is expressed as:

$$MADR_i = \frac{\sum_{n=1}^{ws} |x_i[n] - x_{i+1}[n]|}{ws} \quad (3.6)$$

Similar to MAV, MAD is affected by movement intensity and is therefore scaled by MMAV, which results in the final space domain feature SMADR:

$$SMADR_i = \frac{MADR_i}{MMAV} \quad (3.7)$$

Finally, to increase the amount information the classifier based its decisions upon, the Hudgins feature WL was included. WL represents both amplitude and frequency content of the signal by measuring the summed absolute difference between neighbouring samples in the signal in channel i in one window:

$$WL_i = \sum_{n=1}^{ws-1} |x_i[n+1] - x_i[n]| \quad (3.8)$$

To avoid redundancy in signal representation only SMAV, CC, MADN SMADR and WL were used to train the classifier and for online control.

3.6 The Prosthetic Control System

Having extracted features from the three EMG datasets of one movement for each of the four movements, the control system could be build in order to achieve online recognition of movements. The implementation of the control system was divided into two parts. To achieve recognition of performed movement a classifier was trained, however, this only produces a recognition of a movement and does not reflect the intensity of which the movement is being performed with. Therefore, following the recognition of the performed movements a linear regression model was implemented to achieve proportional control.

3.6.1 Movement Classification

Online classification of movements was accomplished by implementing a LDA classifier. As presented in section 1.7 the classifier needed to be trained using data from each movement. Hence, the five features extracted for each of the 40 %, 50 % and 70 % fraction of the pMVC for one movement were assembled into one labeled matrix. The same was done for the three remaining movements. A fifth class was labeled rest and its matrix only contained the features from a single rest acquisition. These matrices were made for the data acquired from each of the eight channels in the MYB. All matrices were assembled into one training matrix with labels for each movement. Using the *fitcdiscr*-function in MATLAB a LDA classifier was trained by feeding it the training matrix. Hereby, the classifier was trained in separating the classes of pronation, supination, open hand, closed hand and rest, using linear decision boundaries. During online use, the *predict*-function was used to evaluate features from new input data in the classifier and decide which movement there was being performed.

3.6.2 Proportional Control

When a movement was decided upon by the classifier, the proportional control provided the control system with an actuation velocity which was proportional to the contraction intensity. For this purpose a multiple linear regression model was created for each of the four movements, through the MATLAB function *fitlm*. Each model were fitted with the MAV features extracted from the three recorded fractions of pMVC. These were set as the independent variable when training the regressor. The dependent variable was a vector of MAV features from the three fractions which had been normalized to the recorded pMVC. The data was scaled such that the pMVC data was set to 1 and the remaining data was scaled in relation to that. In the online control, the output of the decided regression models was written such that it would move in the direction described in section 3.3. The output from the regression models was limited to a maximum of 1, meaning that maximum level of activation detecting was the pMVC level. This was


equivalent to moving the cursor 1 cm on the computer screen per segment (100 ms). Thus, the cursor had a maximum speed of 1 cm per 100 ms. As the grid system was quadratic with 20 cm in length, each DoF had a full actuation speed of 2 seconds at maximum activation. Thereby, the virtual prosthesis achieved an actuation speed similar to the commercially available Bebionic prosthesis [43]. Another threshold implemented was a minimum activation of at least 15 % for a movement to count, otherwise no output would be provided. This idea behind this implementation was to provide a more stable resting state in addition to the classifier.

3.7 Online Control Training and Test

After the acquisition of training data and the training of the classifier, two stages where the subject could familiarize themselves with the control and test how well they were able to use the control system, were implemented. It was highly critical that the subject was able to achieve sufficient control abilities such that it would not be due to poor control that a subject was not able to perform well when combined with sensory feedback in the final evaluation tests. However, as the classifier only had five classes, representing five separable movements to distinguish between, a robust online control was achieved in an evaluation test during pilot tests after short training trials. Therefore, the need for subject training could be kept to a minimum.

3.7.1 Familiarization with Control

At first, the subject was presented with an image of the grid, which can be seen in figure 3.4, in section 3.3. The subject was able to control a cursor and navigate around inside the grid by performing supination to go right, pronation to go left, open and closed hand to go up and down, and rest to stand still. The familiarization was separated into two stages. In the first stage, the user had three minutes to get acquainted the functionality of the control moving the cursor shown in figure 3.6 (a). The cursor acted as a direct output of the control. During the three minute familiarization, the subject was instructed in training unflawed performance of each movement and the transition from a movement to rest, such that the prosthesis acted accordingly to the subject's intend. In the second stage, the visual feedback in the form of the cursor was made more discrete. This was done by making the cursor invisible and instead highlighting the outlines of the target that contained the cursor. This discretized cursor representation was included to make the visual feedback identical in resolution to the sensory feedback. The discrete visual feedback can be seen in figure 3.6 (b). Again the subject had three minutes to familiarize with the control.



(a) Illustration of the cursor used during the first familiarization stage.

(b) Illustration of the cursor used during the second familiarization stage.

Figure 3.6: Illustration of the cursor used in the two stage familiarization with control. Using the cursor in (a) the subject was informed the exact location of cursor in a square. Using the cursor in (b) the subject was provided with the information of which square the cursor was in, by highlighting it in blue, but not the exact location of the cursor inside the square.

3.7.2 Evaluation Test

After completion of the two stages of familiarization, a target reaching test was carried out. In this project, performance evaluation will be carried out in a virtual environment through a Fitts' Law based target reaching test. In this test, a square was highlighted with red and the subject would have to move the discretized cursor to the desired target and dwell inside it for 1.5 seconds in order for it to be deemed reached. The subject had 30 seconds to reach a target. The test was designed such that each of the 24 squares in the grid was to be reached. When a target was either reached or the time to reach the target ran out, the cursor would be reset to the neutral position (first row, third column in figure 3.4).

The target reaching test was made such that a measure of how well the subjects was able to use the control system could be obtained. Hereby it also acted as a reference for the later comparisons with control in combination with the two feedback schemes. Furthermore, if the investigators deemed the achieved subject control insufficient, the investigators could choose to exclude the subject. This was assessed empirically from observing the subject performance during the control training and quantitatively by evaluating the results of the target reaching test where the subject had to reach 90 % of the target while maximally using a mean time of 10 s per target. From the target reaching test, performance measures of completion rate, time to reach a target and length travelled to reach a target were extracted for performance evaluation.

3.8 Determination of Sensory Thresholds

Having completed the prosthetic control part of the experiment, the next step was to determine the subject's sensory thresholds, which would be used to convey feedback information. The center electrode pads were placed most central on the anterior side of arm, when using a pronated arm as reference. See illustration in figure 3.7. The electrode array was fixated on the contra-lateral arm such that there was a maximum gap of three cm between the outer electrodes posteriorly. It was chosen to place the electrode array on the non-dominant arm since the electrode pads used did not prohibit current leakage. Thus, placing the electrode on the same arm as the MYB would result in interference, which might corrupt recorded EMG data and impair the control system. Providing feedback to contra-lateral arm has, however, proven usable for providing meaningful artificial proprioceptive feedback [44].




Figure 3.7: Illustration of the placement of the electrode array.

In order to provide meaningful sensory feedback to the subject, a range of distinguishable sensory thresholds had to be determined for the subject. As presented in section 3.3 the virtual prosthesis had a range of one to four states. Hence, four thresholds based on amplitude values were determined to accommodate four levels of feedback in the amplitude scheme. Furthermore, since the sensory sensitivity varies across different locations of the circumference of the arm, the sensory thresholds had to be determined for each individual pad in the electrode armband. Sensory thresholds were found by slowly increasing the amplitude while fixating the pulse width and frequency at $500 \mu\text{s}$ and 50 Hz, respectively. In the first round, the lowest threshold was determined, which from now on will be referred

Chapter 3. Methods

to as the perception threshold. For each pad, the amplitude was set to start at $0 \mu A$ and then increase in steps of $100 \mu A$ per second. The subject was instructed in reporting when electrical stimulation could be sensed and that the subject was sure the activated pad was the origin of the perceived stimulation. The pad was deactivated and reactivated once more with the reported sensory threshold amplitude for a second verification. This process was carried out for each pad starting from pad 1 to 16. Subsequently, the subject was presented with the determined amplitudes in each pad, such that the sensation in the current pad was compared to the neighboring pad. This was carried out such that the determined amplitudes could be readjusted to achieve more homogeneous sensations across all pads.

In the second round, the fourth level thresholds, referred to as tolerance threshold, were determined using the same procedure as in round one. The tolerance threshold was defined at highest amplitude the subject felt pleasant. The starting amplitude was in this round, however, the perception threshold. The amplitude was set to increase in steps of $200 \mu A$ per second. The amplitude was increased until the subject reported that the threshold was reached, the stimulation was causing functional muscle activation or a maximum of $10000 \mu A$ was reached. Again the intensities were readjusted to achieve homogeneous sensations. Throughout the process of determining the sensory thresholds, the subject was placed facing away from the computer screen to avoid bias from seeing the used amplitude values.

Using the determined perception and tolerance thresholds, the second and third level thresholds were calculated for the i^{th} pad as the following:

$$2 \text{ lvl}(i) = \text{perception}(i) + \left(\frac{1}{3} \cdot (\text{tolerance}(i) - \text{perception}(i)) \right) \quad (3.9)$$

$$3 \text{ lvl}(i) = \text{tolerance}(i) - \left(\frac{1}{3} \cdot (\text{tolerance}(i) - \text{perception}(i)) \right) \quad (3.10)$$

3.9 Feedback Training

After determining the amplitude thresholds, the subject was trained in understanding the sensory feedback. Depending on which order the subject would test the configurations, the subject was either trained in understanding the spatial or amplitude scheme first. The structure of the training was, however, the same. The feedback training was divided into two phases: familiarization and reinforced learning. These phases will be presented in the following sections.

3.9.1 Familiarization

In the familiarization phase, the subject was presented with the sensation of 12 different grid squares and 12 grid squares indirectly, while observing which grid location corresponded to which sensation. This was carried out by the investigators by moving the

cursor seen in figure 3.8 with the arrow keys on the keyboard. One press with an arrow key would move the cursor in a different grid square in the direction relative to the arrow key pressed. Pressing return would place the cursor in the staring point (the third grid square in the first row). The order of which grid square the cursor would be moved to can be seen in figure 3.8. After reaching a designated square, the cursor would be reset to the starting point. When moving the cursor to a grid square not adjacent to the staring point, the feedback from the transition grid squares would be felt. This transition is transferable to practical proportional prosthetic control/feedback, where the transition from rest to an outer prosthetic position is apparent. When moving the cursor to the grid squares 9, 10, 11 and 12, representing combined DoF positions, the direct route (moving fully in one direction and then the other) was used. In the familiarization phase, the cursor would be moved fully horizontally and then vertically. This enabled stimulations relative to all grid square to be included in the familiarization, without setting all grid squares to be designated targets. This design was chosen due to the single DoF direction being assessed to be most important to get familiar with, and to save time while still exposing the subject to all possible stimulations. Time spend in designated squares was approximately four seconds and time spend in transition squares was approximately two seconds.

2	1	•	3	4
		5		
		6		
		7		
10	9	8	11	12

Figure 3.8: Illustration of the order the cursor would be moved to which grid squares in. The numbered grey squares are designated squares. After reaching a designated square, the cursor would be reset to staring position (the grid square the cursor is located in).

3.9.2 Reinforced Learning

The reinforced learning phase consisted of two blocks, in which all grid squares were designated squares and presented in a predetermined order. When a designated square was reached the investigator asked the subject about the location of the cursor. If the subject

Chapter 3. Methods

answered incorrectly, the investigator would reveal the actual location of the cursor. The stimulation related to the designated would be active until the subject answered. Afterwards, the cursor would be reset to the starting point.

The difference between the two blocks was the order and path to the designated squares, which was predetermined by the investigators. The blocks were, however, identical for all subjects. The path to a designated square was the direct route. However, which direction that would be moved in first was predetermined by the investigators. Thus, the transition stimulations would be felt by the subject before the designated was reached. This design was implemented to avoid bias in being accustomed to always receiving stimulation related to the same DoF first before the other. Time spent in transition squares was approximately two seconds.

3.10 Combining Control and Sensory Feedback

After the subject was trained in understanding one of the feedback schemes, the sensory feedback was combined with prosthetic control. Similar to the control step of the study, see section 3.7, the subject would go through a familiarization phase before undergoing evaluation tests. These stages will be explained in the following sections. An illustration of the full experiment setup used in this trial can be seen in figure 3.9.

3.10.1 Familiarization

This stage was similar to the second stage control training, see section 3.7.1, with the addition of receiving sensory feedback. The subject was instructed in familiarize with the prosthetic control again, while adapting further to the feedback related to the various prosthesis states as well the transitions felt when changing prosthesis state. These focus point were assessed most beneficial to train in order to perform well in the final evaluation test. The duration of the familiarization phase was three minutes.

3.10.2 Final Evaluation Test

This evaluation test was similar to the evaluation test from the control step, see section 3.7.2, with the addition of receiving sensory feedback. In this test, the prosthesis state would, however, not be visualized. Thus, the subject had to use the sensory feedback to determine the prosthesis state and reach the highlighted targets. The target reaching test was performed two times consecutively. Again the completion rate, time to reach a target and length travelled to reach a target were extracted for performance evaluation.




Figure 3.9: Image of the experimental setup. 1) is the stimulation device with the stimulation electrode placed under the brown armband, 2) is the electrode armband used to record EMG signals and 3) is the computer screen used to guide the subject and display tasks.

4 | Results

Bibliography

- [1] Li Guanglin, Aimee E. Schultz, and Todd A. Kuiken. “Quantifying Pattern Recognition—Based Myoelectric Control of Multifunctional Transradial Prostheses”. In: *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 18.2 (2010), pp. 185–192.
- [2] M Štrbac et al. “Integrated and flexible multichannel interface for electrotactile stimulation”. In: *Journal of Neural Engineering* 13.4 (2016), pp. 1–16.
- [3] G. Di Pino, E. Guglielmelli, and P. M. Rossini. “Neuroplasticity in amputees: Main implications on bidirectional interfacing of cybernetic hand prostheses”. In: *Progress in Neurobiology* 88.2 (2009), pp. 114–126.
- [4] Jonathon S. Schofield et al. “Applications of sensory feedback in motorized upper extremity prosthesis: A review”. In: *Expert Review of Medical Devices* 11.5 (2014), pp. 499–511.
- [5] Benjamin Stephens-Fripp, Gursel Alici, and Rahim Mutlu. “A review of non-invasive sensory feedback methods for transradial prosthetic hands”. In: *IEEE Access* 6 (2018), pp. 6878–6899.
- [6] Christian Antfolk et al. “Sensory feedback in upper limb prosthetics”. In: *Expert Reviews Ltd* 10.1 (2013), pp. 45–54.
- [7] H Benz et al. “Upper extremity prosthesis user perspectives on innovative neural interface devices”. In: *Neuromodulation* 20 (2) (2016), pp. 287–290.
- [8] Ping Shi and Xiafeng Shen. “Sensation Feedback and Muscle Response of Electrical Stimulation on the Upper Limb Skin: A Case Study”. In: *Proceedings - 2015 7th International Conference on Measuring Technology and Mechatronics Automation, ICMTMA 2015* (2015), pp. 969–972.
- [9] D Pamungkas and K Ward. “Electro-tactile feedback system for a prosthetic hand”. In: *22nd Annual International Conference on Mechatronics and Machine Vision in Practice, M2VIP 2015* (2015), pp. 27–38.
- [10] Heng Xu et al. “Effects of Different Tactile Feedback on Myoelectric Closed-Loop Control for Grasping Based on Electrotactile Stimulation”. In: *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 24.8 (2016), pp. 827–836.
- [11] Nikola Jorgovanovic et al. “Virtual grasping: Closed-loop force control using electro-tactile feedback”. In: *Computational and Mathematical Methods in Medicine* 2014 (2014).
- [12] Milica Isaković et al. “Electrotactile feedback improves performance and facilitates learning in the routine grasping task”. In: *European Journal of Translational Myology* 26.3 (2016), pp. 197–202.

Bibliography

- [13] C. Hartmann et al. "Towards prosthetic systems providing comprehensive tactile feedback for utility and embodiment". In: *IEEE 2014 Biomedical Circuits and Systems Conference, BioCAS 2014 - Proceedings* (2014), pp. 620–623.
- [14] M. Franceschi et al. "Preliminary evaluation of the tactile feedback system based on artificial skin and electrotactile stimulation". In: *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society, EMBS* (2015), pp. 4554–4557.
- [15] Dorindo G Buma et al. "Intermittent Stimulation Delays Adaptation to Electrocutoaneous Sensory Feedback". In: 15.3 (2007), pp. 435–441.
- [16] Andrew Y.J. Szeto and Frank A. Saunders. "Electrocutoaneous Stimulation for Sensory Communication in Rehabilitation Engineering". In: *IEEE Transactions on Biomedical Engineering* BME-29.4 (1982), pp. 300–308.
- [17] Strahinja Dosen et al. "Multichannel Electrotactile Feedback With Spatial and Mixed Coding for Closed-Loop Control of Grasping Force in Hand Prostheses". In: *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 25.3 (2016), pp. 183–195.
- [18] Marko Markovic et al. "The clinical relevance of advanced artificial feedback in the control of a multi-functional myoelectric prosthesis". In: *Journal of NeuroEngineering and Rehabilitation* 15.1 (2018), pp. 1–15.
- [19] Christian Cipriani et al. "On the Shared Control of an EMG-Controlled Prosthetic Hand: Analysis of User-Prosthesis Interaction". In: *IEEE Transactions on Robotics* 24.1 (2008), pp. 170–184.
- [20] Heidi J.B. Witteveen, Johan S. Rietman, and Peter H. Veltink. "Grasping force and slip feedback through vibrotactile stimulation to be used in myoelectric forearm prostheses". In: *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society, EMBS* (2012), pp. 2969–2972.
- [21] VINCENT Systems. "VINCENT EVOLUTION 2". In: <https://vincentsystems.de/en/prosthetics/vincent-evolution-2/> (2005).
- [22] Purushothaman Geethanjali. "Myoelectric control of prosthetic hands : state-of-the-art review". In: *Medical Devices: Evidence and Research* (2016), pp. 247–255.
- [23] Hande Türker and Hasen Sözen. "Surface Electromyography in Sports and Exercise". In: *Electrodiagnosis in New Frontiers of Clinical Research up 2* (2013), pp. 175–194.
- [24] Frederic H. Martini, Judi L. Nath, and Edwin F. Bartholomew. *Fundamentals of Anatomy and Physiology*. 9th. 2012.
- [25] Jeffrey R. Cram. *Cram's Introduction to Surface EMG*. Ed. by Criswell Eleanor. Second edi. 2012.
- [26] I Mendez et al. "Evaluation of the Myo Armband for the Classification of hand motions". In: *International Conference on Rehabilitation Robotics* (2017), pp. 1211–1214.

- [27] Radhika Menon et al. “Study on Interaction Between Temporal and Spatial Information in Classification of EMG Signals for Myoelectric Prostheses”. In: *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 25.10 (2017), pp. 1832–1842.
- [28] Angkoon Phinyomark, Pornchai Phukpattaranont, and Chusak Limsakul. “Feature reduction and selection for EMG signal classification”. In: *Expert Systems with Applications* 39 (2012), pp. 7420–7431.
- [29] B. Hudgins, P. Parker, and R.N. Scott. “A new strategy for multifunction myoelectric control”. In: *IEEE Transactions on Biomedical Engineering* 40.1 (1993), pp. 82–94.
- [30] Ian M. Donovan et al. “Simple space-domain features for low-resolution sEMG pattern recognition”. In: *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society, EMBS* (2017), pp. 62–65.
- [31] Richard O. Duda, Peter E. Hart, and David G. Stork. *Pattern Classification*. 2nd Editio. 2001, p. 680.
- [32] Kevin Englehart and Bernard Hudgins. “A robust, real-time control scheme for multifunction myoelectric control.” In: *IEEE transactions on bio-medical engineering* 50.7 (2003), pp. 848–854.
- [33] Erik J. Scheme, Bernard S. Hudgins, and Kevin B. Englehart. “Confidence-based rejection for improved pattern recognition myoelectric control”. In: *IEEE Transactions on Biomedical Engineering* 60.6 (2013), pp. 1563–1570.
- [34] Anders Fougnier et al. “Control of upper limb prostheses: Terminology and proportional myoelectric controla review”. In: *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 20.5 (2012), pp. 663–677.
- [35] Jerrold H. Zar. *Biostastical Analysis*. 5th ed. Pearson, 2009.
- [36] Enzo Mastinu et al. “An Alternative Myoelectric Pattern Recognition Approach for the Control of Hand Prostheses: A Case Study of Use in Daily Life by a Dysmelia Subject”. In: *IEEE Journal of Translational Engineering in Health and Medicine* 6.June 2017 (2018).
- [37] Michael A. Powell, Rahul R. Kaliki, and Nitish V. Thakor. “User training for pattern recognition-based myoelectric prostheses: Improving phantom limb movement consistency and distinguishability”. In: *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 22.3 (2014), pp. 522–532.
- [38] Sophie M. Wurth and Levi J. Hargrove. “A real-time comparison between direct control, sequential pattern recognition control and simultaneous pattern recognition control using a Fitts’ law style assessment procedure”. In: *Journal of NeuroEngineering and Rehabilitation* 11.1 (2014), pp. 1–13.
- [39] J. M. Hahne et al. “Linear and nonlinear regression techniques for simultaneous and proportional myoelectric control”. In: *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 22.2 (2014), pp. 269–279.

Bibliography

- [40] M. Pla Mobarak et al. “Hand movement classification using transient state analysis of surface multichannel EMG signal”. In: *Pan American Health Care Exchanges, PAHCE April* (2014).
- [41] Alexander Boschmann, Barbara Nofen, and Marco Platzner. “Improving Transient State Myoelectric Signal Recognition in Hand Movement Classification using Gyroscopes”. In: *2013 35th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)* (2013), pp. 6035–6038.
- [42] Michele Pla Mobarak, Juan Manuel, and Gutiérrez Salgado. “Transient State Analysis of the Multichannel EMG Signal Using Hjorth ’ s Parameters for Identification of Hand Movements”. In: *The Ninth International Multi-Conference on Computing in the Global Information Technology* (2014), pp. 24–30.
- [43] Joseph T Belter et al. “Mechanical design and performance specifications of anthropomorphic prosthetic hands: a review.” In: *Journal of rehabilitation research and development* 50.5 (2013), pp. 599–618.
- [44] Tobias Pistohl et al. “Artificial proprioceptive feedback for myoelectric control”. In: *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 23.3 (2015), pp. 498–507.

Bibliography

A | Appendices

A.1 Experiment Protocol

Project Title

Evaluation of electrotactile feedback schemes in combination with myoelectric prosthetic control - closing the loop.

Information on Investigators

The investigators are Biomedical Engineering Master students at Aalborg University.

Background

Losing an upper limb can be hugely debilitating and can result in lowered quality of life due to restrictions in function, appearance and sensation. As a mean to regain the functionality, transradial amputees can receive a functional prosthesis, where the majority are controlled by muscles signals, or myoelectric (EMG) signals. However, still 25% of myoelectric prosthesis users reject their device, where a major reason for the low satisfaction is due to lack of sensory feedback. Many advancements have been made in the academic community to improve function accuracy. However, combining function with sensory feedback, thus closing the motor/sensory loop, is still a scarcely investigated area. Therefore, this experiment will combine the control of a prosthesis with sensory feedback delivered via electrotactile stimulation electrodes placed on the forearm. During the experiment the subjects will test two different feedback configurations while controlling a virtual prosthesis, represented as a cursor on a computer screen. The subject can move the cursor in a two-dimensional coordinate system, where the axes represents a degree of freedom (DOF) each (wrist rotation and opened/closed hand).

Purpose

The purpose of the experiment is to compare how subjects' perform in an evaluation test when receiving feedback from two different electrotactile stimulation configurations, respectively, in a closed loop virtual prosthesis. This might provide information on which feedback that seems more intuitive to use in practice in a prosthesis.

Research Aim

Test and evaluate two novel stimulation schemes, one based on modulating amplitude and one based on spatial localization of activation, for conveying sensory feedback of the prosthesis state in a closed loop prosthetic control system.

Experiment Duration

Approximately 2 hours and 30 minutes.

Inclusion Criteria

The subject must be:

- able bodied.
- at least 18 years of age.
- able to understand, read and speak English and/or Danish.
- assessed by the investigators to comply with the instructions given during the experiment.

Exclusion Criteria

The subject must:

- not have any diseases/conditions that may influence sensory perception.
- be willing to receive low amplitude current stimulation.
- be assessed by the investigators to have robust prosthetic control during the experiment.
- be willing to give informed consent.

Experiment Description

The main aim of the experiment is for the subject to be able to correctly interpret the two sensory feedback schemes when combined with myoelectric prosthetic control. The grid illustrated in figure A.2 is the map the subject will be able to navigate inside. Each square in the map will deliver a different stimulus corresponding to the motion state of the virtual prosthesis, represented as the black cursor. The square with center in the origin (square with cursor inside in figure A.2) corresponds to resting state and will provide no sensory feedback. The remaining squares in the first row will deliver stimuli corresponding to only the wrist rotation degree of freedom, and the remaining squares in the third column will deliver stimuli only corresponding to the closed hand DOF. The remaining squares will deliver a stimulation based on a combination of the two DOFs. The further away from resting state a square is, corresponds to the angular degree of the prosthesis state in relation to the performed movement (see figure A.1).

Appendix A. Appendices




Figure A.1: Figure (a) shows the spatial scheme, which is based on different pads being activated depending on the level of the grid square the cursor is located in. The highest number of possibly activated pads is four at a time. Figure (b) shows the amplitude scheme. Here, the amplitude of the active pads will increase with the increase of the level of the target location. The highest number of possibly activated pads is eight at a time.

The arrows in the upper right corner of figure A.2 represent the hand movements needed to be performed to move the cursor in the corresponding direction. The control system will only respond to single DOF movements. Thus, the cursor is only able to move along one axis at a time and not diagonally. The subject will control the cursor with the dominant arm through an EMG electrode armband. The subject will receive stimulation from an electrode consisting of 16 electrode pads placed around the contra-lateral forearm (see illustration in figure A.1).




Figure A.2: Image of the grid map and cursor used in the experiment. Performing supination moves the cursor the right, pronation moves it to the left and closing the hand moves it down. Opening the hand moves the cursor up, and is used as a correction movement if a wrong movement has been made.

Hand Movements Used for Prosthetic Control




Figure A.3: Image of the hand movements used in the experiment for myoelectric prosthetic control. From top left corner: Wrist pronation, wrist supination, opened hand and closed hand.

Experiment Procedure

Before the final evaluation test is carried out the subject will be trained in controlling the cursor via EMG signals, trained in interpreting the sensory feedback and trained in interpreting the sensory feedback while controlling the cursor. The evaluation test is a target reaching test, where the subject needs to move the cursor to a highlighted target consisting of one of the grid squares. The cursor will not be visible, thus, the subject will have to only rely on the information received from the sensory feedback.

During the experiment the subject must let the dominant arm hang relaxed down the side of the torso and the contra-lateral arm placed on a table without putting pressure on the stimulation electrode, as seen in figure A.5. The subject must be seated during all procedures. The following order represent the chronology of the procedures the subject needs to undergo; the steps will be divided in solely control, solely sensory feedback and feedback with control.

Control

1. Record EMG signals needed to build the prosthetic control system. To do this the subject must first perform five movements used as reference signals: 15 seconds rest, 15 seconds prolonged maximum voluntary contraction (pMVC) of wrist supination, 15 seconds pMVC of wrist pronation, 15 seconds pMVC of opened hand and 15 seconds pMVC of closed hand. Between each contraction the subject will get a 15 seconds break to avoid fatigue. Secondly, the subject must perform movements from which the recorded signals are used to build the control system. Here, the subject controls a cursor as seen in figure A.4, and must match the cursor with trapezoidal trajectory. The cursor moves horizontally with time and the subject control the contraction intensity vertically. The subject must perform three contractions per movement: 40 %, 50 % and 70 % of the pMVC. The plateau of the trapezoidal trajectory corresponds to the designated fraction. Between each performed movement, the subject gets a 15 seconds break to avoid fatigue. Lastly, a 15 seconds rest is recorded.
2. Train the subject's ability to control the cursor via letting the subject move freely around inside the grid map for three minutes.
3. Train the subject's ability to control the cursor via letting the subject move freely around inside the grid map for three minutes. In this training, the cursor will not be visible, but the square the cursor is inside will be highlighted in blue. This cursor representation will be used in the remaining trainings and tests.
4. Perform target reaching test to evaluate the subject's ability to control the cursor. The designated target will be one of the squares in the grid highlighted in red. To reach a target the blue cursor square must match the target and dwell inside it for 1.5 seconds. Then a bell sound will occur and a new target appears. The time limit for reaching a target is 30 seconds. The starting point is always the resting state square (first square in third column), and the cursor will, thus, return to starting point when a target is reached or then the time limit is reached. A total of 24 targets will appear before this test is through.

Appendix A. Appendices




Figure A.4: Image of the trapezoidal trajectory used when recording EMG signals used to build the control system. The subject controls the black cursor in height by increasing contraction intensity.

Sensory Feedback

1. Record current amplitude thresholds needed to build the sensory feedback schemes. For each electrode the subject must first note when the stimulus is felt clearly. When all thresholds are set, the sensation of neighbouring electrodes are compared to ensure homogeneity in the sensation. Afterwards the same procedure is performed for the subject's tolerance threshold.
2. Train the subject's ability to interpret sensory feedback for one of the schemes. This is done by exposing the subject to feedback from 12 grid squares. The subject will experience the transitions from square to square until the designated square is reached. The path taken to reach the designated square is the direct route (full length in one direction and then the other), but which direction that will be travelled first is predetermined by the investigators.
3. Perform reinforcement learning on all grid squares. The path taken to reach the designated square is the direct route, but which direction that will be travelled first is predetermined by the investigators. During this step the subject must look away from the computer screen. When a designated square is reached, the subject will be asked where the cursor is located. After answering the subject will be informed on whether it was correct, and told the correct location, if the answer was incorrect.
4. Repeat reinforcement learning from step 3. The order of the squares and the route to each square will vary from step 3.

Sensory Feedback with Control

Appendix A. Appendices

1. Train the subject's ability to control the cursor while receiving sensory feedback via letting the subject move freely around inside the grid map for three minutes.
2. Perform target reaching test where the cursor is invisible to evaluate how well the subject can utilize the sensory feedback regarding the cursor location. This test has the same format as the target reaching test from the control procedure step 3.
3. Repeat target reaching test from step 2.
4. Redo sensory feedback steps and sensory feedback with control steps with sensory feedback from the remaining scheme.

Experiment Setup




Figure A.5: Image of the experimental setup. 1) is the stimulation device with the stimulation electrode placed under the brown armband, 2) is the electrode armband used to record EMG signals and 3) is the computer screen used to guide the subject and display tasks.

A.2 Short Experiment Description

Project Title

Evaluation of electrotactile feedback schemes in combination with myoelectric prosthetic control - closing the loop.

Experiment Purpose

When a person gets amputated on the lower arm, he/she can receive a functional prosthesis. This is controlled by muscle signals from the user, where the muscle signals are translated into a prosthesis movement. However, many functional prostheses do not provide sensory feedback, which results in some users to abandon their prosthesis.

The purpose of the experiment is to compare how subjects perform in an evaluation test when receiving feedback from two different electrical stimulation configurations, while controlling a virtual prosthesis. The results might provide information on which feedback that seems more intuitive to use in a real prosthesis.

Experiment Overview

The experiment will take place in the laboratory D3 – 107 at Aalborg University. The duration of the experiment is estimated to be 2 hours and 30 minutes.

During the experiment a myoelectric armband will be placed on the dominant forearm and used to record muscle activity while the subject performs four different hand gestures. Subsequently, an evaluation of the ability to reproduce the gestures will be made.

Afterwards, an electrode armband capable of delivering electrical stimulation at 16 different locations will be placed on the non-dominant arm. The subject will determine the level of sensory perception and tolerance level. The subject will then be made familiar with and trained in understanding two different feedback configurations that represents the possible states the virtual prosthesis can be in. At the end, an evaluation of the subject's ability to understand the feedback while making the trained hand gestures will be made.

At the day of the experiment, please refrain from using any types of sensory deprivation drugs (painkillers and alike). The test subject will not receive monetary compensation after the experiment.