
(a) Both rod sensors agree at 5–35 kHz
Coherence γ² (1 = channels agree)
1.0
0.8
0.6
0.4
0.2
0.0
γ² = 0.9 (usable threshold)
Calibrated band (5–35 kHz)
0
20
40
60
80
100
Frequency (kHz)
(b) Response collapses above 35 kHz (LM358 limit)
Amplitude (mV RMS)
10²
10¹
9.17 kHz
eigenmode
~30 kHz
~20 kHz
CH3 (rod A)
CH4 (rod B)
0
20
40
60
80
100
Frequency (kHz)