



MIRcat™ LASER Final Test Sheet

1.0	System Information	
	Customer Name	SETI Institute
	Model	MIRcat-QT-Z-2100
	Serial Number	10524
	Software version	1.9.0.4
	Firmware version	3.1.0
	Motion Firmware Version	0.0.0
	ATP Operator's Name	D. DeLong

1.1	Operating Parameters	QCL 1		QCL 2		QCL 3		QCL 4	
	Module	M2058-PCX		-		-		-	
	Operating Mode	Pulsed	CW	Pulsed	CW	Pulsed	CW	Pulsed	CW
	Current Range (mA)	1000	775	-	-	-	-	-	-
	Laser Temp (°C)	21	21	-	-	-	-	-	-
	Cooling Water Temp (°C)	-	17	-	-	-	-	-	-

1.2	Tuning Parameters	QCL 1		QCL 2		QCL 3		QCL 4	
	Pulsed Tuning Range (μm)	4.815-6.017		-		-		-	
	CW Tuning Range (μm)	5.155-6.102		-		-		-	
	Crossover λ (μm)	6.102		-		-		-	


1.3	Test Operating Conditions	
	Scan Mode - Pulsed	Step and Measure
	Scan Mode - CW	Tune
	Frequency	100 KHz
	Pulse Width	500ns
	Duty Cycle	5%
	QCL Temperature	21°C



2.1 Graph – Performance test 100KHz 500ns 5% & CW

Test Date

11/19/2024





2.2	High Pulse Rate Graph – Performance test 2MHz 100ns 20%	
	Test Date	11/19/2024

