MICRO PRECISION

ilac-MRA

ANAB
ANSI National Accreditation Board
ACCREDITED
ISO/IEC 17025
CALIBRATION AND DIMENSIONAL MEASUREMENT
Accreditation Number: AC-1969

Certificate of Calibration

Certificate Number: 5523631031761539

Customer:

AMERICAN INDUSTRIES DE OCCIDENTE
CARRETERA GUADALAJARA MORELIA #19200 INT.1,2,3 Y 4
COL. BUENAVISTA
TLAJOMULCO DE ZUÑIGA JALISCO 45640

Date : Apr 18, 2025
Work Order : GDL-458342

MP Control #:	30191	Serial Number:	3091
Asset ID:	30191	Department:	N/A
Description:	HYGRO THERMOMETER HUMIDITY ALERT II	Location:	MPC LAB
Manufacturer:	EXTECH INSTRUMENTS	Received Condition:	IN TOLERANCE
Model Number:	445815	Returned Condition:	IN TOLERANCE
Size:	-10 to 60 ºC/ 10 to 99 %RH	Cal. Date:	Apr 17, 2025
Resolution:	0.1 ºC / 1 %RH	Cal. Interval:	12 MONTHS
Temp./RH:	22.5°C / 44 % RH	Cal. Due Date:	Apr 17, 2026

STATEMENTS OF PASS OR FAIL CONFORMANCE: The uncertainty of measurement has been taken into account when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accept does not exceed 2% in compliance with ANSI/NCSL Z540.3-2006.

THE CALIBRATION REPORT STATUS:

PASS - Term used when compliance statement is given, and the measurement result is PASS.
PASS^z - Term used when compliance statement is given, and the measurement result is conditional passed or PASS^z.
FAIL - Term used when compliance statement is given, and the measurement result is FAIL.
FAIL^z - Term used when compliance statement is given, and the measurement result is conditional failed or FAIL^z.
REPORT OF VALUE - Term used when reported measurement is not requiring compliance statement in report.
ADJUSTED - When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value as left.
LIMITED - When an instrument fails calibration but is still functional in a limited manner.

The expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%, unless otherwise stated. This calibration report complies with ISO/IEC 17025:2017 and ANSI/NCSL Z540.3. Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. All standards are traceable to SI through the National Institute of Standards and Technology (NIST) and/or recognized national or international standards laboratories. Services rendered include proper manufacturer's service instruction and are warranted for no less than thirty (30) days. The information on this report pertains only to the instrument identified, this may not be reproduced in part or in a whole without the prior written approval of the issuing MP Calibration Laboratory.

Standards Used to Calibrate Equipment

I.D.	Manufacturer	Description	Model	Traceability Number	Cal. Due Date
EM3500	VAISALA	SATURATED SALT SOLUTION	19729HM	K008-F02155	Apr 25, 2025
EM3501	VAISALA	SATURATED SALT SOLUTION	19730HM	K008-F04325	Aug 16, 2025
EM3502	VAISALA	SATURATED SALT SOLUTION	19731HM	K008-F04988	Sep 22, 2025
EM3503	VAISALA	SATURATED SALT SOLUTION	19732HM	K008-F03731	Jul 15, 2025
DG1899	VAISALA	HUMIDITY AND TEMPERATURE METER	MI70 / HMP75	5523631030751641	Jul 16, 2025
DJ8043	EXTECH INSTRUMENTS	TEMPERATURE & HUMIDITY DATALOGG	42280	5523631031211092	Jun 1, 2025

Procedures Used in this Event

Procedure Name	Description
MPC-THD-001	Temperature, Humidity and Dew Point Devices, General, Rev.03, Jul-15-2024

Additional notes: See attached data.

Calibrating Technician: JUAN LANDERO

Approved By: MAXIMILIANO LEMOS