MICRO PRECISION

ilac-MRA

ANAB
ANSI National Accreditation Board
ACCREDITED
ISO/IEC 17025
CALIBRATION AND DIMENSIONAL MEASUREMENT
Accreditation Number: AC-1969

Certificate of Calibration

Certificate Number: 5523631031421963

Customer:

ZHONGLI TALESUN HONG KONG LIMITED
CARRETERA GUADALAJARA MORELIA #19200 INT. 3
COL. BUENAVISTA
TLAJOMULCO DE ZUÑIGA JALISCO 45640

Date : Dec 4, 2024
Work Order : GDL-456069

MP Control #:	EV8003	Serial Number:	OPM0001
Asset ID:	EV8003	Department:	N/A
Description:	OPTICAL POWER METER	Location:	MP LAB
Manufacturer:	LINKEDPRO BY EPCOM	Received Condition:	IN TOLERANCE
Model Number:	DXP-50D	Returned Condition:	IN TOLERANCE
Size:	MULTIFUNCTION	Cal. Date:	Nov 26, 2024
Resolution:	MULTIRESOLUTION	Cal. Interval:	12 MONTHS
Temp./RH:	20.2°C / 45 % RH	Cal. Due Date:	Nov 26, 2025

STATEMENTS OF PASS OR FAIL CONFORMANCE: The uncertainty of measurement has been taken into account when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accept does not exceed 2% in compliance with ANSI/NCSL Z540.3-2006.

THE CALIBRATION REPORT STATUS:

PASS - Term used when compliance statement is given, and the measurement result is PASS.
PASS^{z} - Term used when compliance statement is given, and the measurement result is conditional passed or PASS^{z}.
FAIL - Term used when compliance statement is given, and the measurement result is FAIL.
FAIL^{z} - Term used when compliance statement is given, and the measurement result is conditional failed or FAIL^{z}.
REPORT OF VALUE - Term used when reported measurement is not requiring compliance statement in report.
ADJUSTED - When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value as left.
LIMITED - When an instrument fails calibration but is still functional in a limited manner.

The expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%, unless otherwise stated. This calibration report complies with ISO/IEC 17025:2017 and ANSI/NCSL Z540.3. Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. All standards are traceable to SI through the National Institute of Standards and Technology (NIST) and/or recognized national or international standards laboratories. Services rendered include proper manufacturer's service instruction and are warranted for no less than thirty (30) days. The information on this report pertains only to the instrument identified, this may not be reproduced in part or in a whole without the prior written approval of the issuing MP Calibration Laboratory.

Standards Used to Calibrate Equipment

I.D.	Manufacturer	Description	Model	Traceability Number	Cal. Due Date
BN7748	EXFO	WDM LASER SOURCE	IQS-2400	FBL-CC-D853-2024	May 29, 2026
BN7747	EXFO	WDM LASER SOURCE	IQS-2400	FBL-CC-D852-2024	May 29, 2026
BN7749	EXFO	VARIABLE ATTENUATOR	IQS-3150	FBL-CC-D854-2024	Aug 27, 2025
BN7750	EXFO	CALIBRATION POWER METER	IQS-1500	FBL-CC-D851-2024	May 24, 2026
DJ8044	EXTECH INSTRUMENTS	TEMPERATURE & HUMIDITY DATALOGG	42280	5523631031053291	Jul 2, 2025

Procedures Used in this Event

Procedure Name	Description
MPC-OPM-001	Optical Power Meters, General, Rev.03, Jul-15-2024

Additional notes: See attached data.

Calibrating Technician: José Trigo

JOSE TRIGO

Approved By: Victor Cruz

VICTORCRUZ