



## Certificate of Calibration

Certificate Number: 5523631031075636

### Customer:

ZHONGLI TALESUN HONG KONG LIMITED  
CARRETERA GUADALAJARA MORELIA #19200 INT. 3  
COL. BUENAVISTA  
TLAJOMULCO DE ZUÑIGA JALISCO 45640

Date : Jul 15, 2024

Work Order : GDL-453405

MP Control #:	EG3412	Serial Number:	C148059438
Asset ID:	EG3412	Department:	N/A
Description:	SCALE	Location:	ON SITE CALIBRATION
Manufacturer:	OHAUS	Received Condition:	IN TOLERANCE
Model Number:	DEFENDER T32XW	Returned Condition:	IN TOLERANCE
Size:	200 kg	Cal. Date:	Jul 03, 2024
Resolution:	0.2 kg	Cal. Interval:	12 MONTHS
Temp./RH:	22.6°C / 36 % RH	Cal. Due Date:	Jul 03, 2025

**STATEMENTS OF PASS OR FAIL CONFORMANCE:** The uncertainty of measurement has been taken into account when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accept does not exceed 2% in compliance with ANSI/NCSL Z540.3-2006.

### THE CALIBRATION REPORT STATUS:

PASS - Term used when compliance statement is given, and the measurement result is PASS.

PASS<sup>z</sup> - Term used when compliance statement is given, and the measurement result is conditional passed or PASS<sup>z</sup>.

FAIL - Term used when compliance statement is given, and the measurement result is FAIL.

FAIL<sup>z</sup> - Term used when compliance statement is given, and the measurement result is conditional failed or FAIL<sup>z</sup>.

**REPORT OF VALUE** - Term used when reported measurement is not requiring compliance statement in report.

**ADJUSTED** - When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value as left.

**LIMITED** - When an instrument fails calibration but is still functional in a limited manner.

The expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%, unless otherwise stated. This calibration report complies with ISO/IEC 17025:2017 and ANSI/NCSL Z540.3. Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. All standards are traceable to SI through the National Institute of Standards and Technology (NIST) and/or recognized national or international standards laboratories. Services rendered include proper manufacturer's service instruction and are warranted for no less than thirty (30) days. The information on this report pertains only to the instrument identified, this may not be reproduced in part or in a whole without the prior written approval of the issuing MP Calibration Laboratory.

### Standards Used to Calibrate Equipment

I.D.	Manufacturer	Description	Model	Traceability Number	Cal. Due Date
BP4831	METTLER TOLEDO	WEIGHT	CLASS OIML M2	5523631030589569	Dec 29, 2024
BP4835	METTLER TOLEDO	WEIGHT	CLASS OIML M2	5523631030589613	Jul 3, 2025
BP4833	METTLER TOLEDO	WEIGHT	CLASS M2	5523631030194947	Jul 12, 2024
BP4635	METTLER TOLEDO	WEIGHT	CLASS OIML M2	5523631030563213	Dec 8, 2025
BP4631	METTLER TOLEDO	WEIGHT	CLASS M2	5523631030320064	Sep 4, 2024
BP4633	METTLER TOLEDO	WEIGHT	CLASS OIML M2	5523631030586963	Dec 29, 2024
BP4630	METTLER TOLEDO	WEIGHT	CLASS OIML M2	5523631030586945	Dec 29, 2024
BP4628	METTLER TOLEDO	WEIGHT	CLASS OIML M2	5523631030586669	Dec 29, 2024
BP4627	METTLER TOLEDO	WEIGHT	CLASS OIML M2	5523631030563139	Dec 8, 2024
BP4625	METTLER TOLEDO	WEIGHT	CLASS OIML M2	5523631030591824	Jan 2, 2025
ED6101	STEREN	TERMOMETRO DIGITAL	TER-150	5523631030977330	Jun 4, 2025

Calibrating Technician:

SAMUEL ORTIZ

Approved By:

MAGNOLIA TORRES