

## Certificate of Calibration

Certificate Number: 5523631031756919

### Customer:

AMERICAN INDUSTRIES DE OCCIDENTE  
CARRETERA GUADALAJARA MORELIA #19200 INT.1,2,3 Y 4  
COL. BUENAVISTA  
TLAJOMULCO DE ZUÑIGA JALISCO 45640

Date : Apr 16, 2025  
Work Order : GDL-458413

|               |                        |                     |                  |
|---------------|------------------------|---------------------|------------------|
| MP Control #: | ED4709                 | Serial Number:      | 800000285118G8I8 |
| Asset ID:     | ED4709                 | Department:         | N/A              |
| Description:  | FLEXOMETER             | Location:           | MPC LAB          |
| Manufacturer: | N/A                    | Received Condition: | IN TOLERANCE     |
| Model Number: | MIT-3033               | Returned Condition: | IN TOLERANCE     |
| Size:         | 0 to 4 m / 0 to 160 in | Cal. Date:          | Apr 15, 2025     |
| Resolution:   | 1 mm / 1/16 in         | Cal. Interval:      | 12 MONTHS        |
| Temp./RH:     | 19.8°C / 37.6 % RH     | Cal. Due Date:      | Apr 15, 2026     |

**STATEMENTS OF PASS OR FAIL CONFORMANCE:** The uncertainty of measurement has been taken into account when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accept does not exceed 2% in compliance with ANSI/NCSL Z540.3-2006.

### THE CALIBRATION REPORT STATUS:

**PASS** - Term used when compliance statement is given, and the measurement result is PASS.

**PASS<sup>2</sup>** - Term used when compliance statement is given, and the measurement result is conditional passed or PASS<sup>2</sup>.

**FAIL** - Term used when compliance statement is given, and the measurement result is FAIL.

**FAIL<sup>2</sup>** - Term used when compliance statement is given, and the measurement result is conditional failed or FAIL<sup>2</sup>.

**REPORT OF VALUE** - Term used when reported measurement is not requiring compliance statement in report.

**ADJUSTED** - When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value as left.

**LIMITED** - When an instrument fails calibration but is still functional in a limited manner.

The expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%, unless otherwise stated. This calibration report complies with ISO/IEC 17025:2017 and ANSI/NCSL Z540.3. Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. All standards are traceable to SI through the National Institute of Standards and Technology (NIST) and/or recognized national or international standards laboratories. Services rendered include proper manufacturer's service instruction and are warranted for no less than thirty (30) days. The information on this report pertains only to the instrument identified, this may not be reproduced in part or in a whole without the prior written approval of the issuing MP Calibration Laboratory.

### Standards Used to Calibrate Equipment

| I.D.   | Manufacturer       | Description                     | Model     | Traceability Number | Cal. Due Date |
|--------|--------------------|---------------------------------|-----------|---------------------|---------------|
| DJ8045 | EXTECH INSTRUMENTS | TEMPERATURE & HUMIDITY DATALOGG | 42280     | 5523631031053293    | Jul 2, 2025   |
| CB7423 | VISION             | OPTICAL COMPARATOR              | 665406-19 | 5523631031042906    | Feb 7, 2026   |

### Procedures Used in this Event

#### Procedure Name


MPC-RTM-001

#### Description


Rulers and Tape Measures, General, Rev.05, May-29-2024

Additional notes: See attached data.

Calibrating Technician:

  
DULCE MARTINEZ

Approved By:

  
ALEJANDRO GARCIA