



Certificate of Calibration

Certificate Number: 5523631031234271

Customer:

ZHONGLI TALESUN HONG KONG LIMITED
CARRETERA GUADALAJARA MORELIA #19200 INT. 3
COL. BUENAVISTA
TLAJOMULCO DE ZUÑIGA JALISCO 45640

Date : Sep 18, 2024

Work Order : GDL-454773

MP Control #:	107755	Serial Number:	107755
Asset ID:	107755	Department:	N/A
Description:	3-PATH DIODE POWER SENSOR	Location:	ON SITE CALIBRATION
Manufacturer:	ROHDE & SCHWARZ	Received Condition:	IN TOLERANCE
Model Number:	NRP8S	Returned Condition:	IN TOLERANCE
Size:	10 MHz - 8 GHz (-70 to 23 dBm)	Cal. Date:	Sep 09, 2024
Resolution:	MULTIRESOLUTION	Cal. Interval:	12 MONTHS
Temp./RH:	26.6°C / 40 % RH	Cal. Due Date:	Sep 09, 2025

STATEMENTS OF PASS OR FAIL CONFORMANCE: The uncertainty of measurement has been taken into account when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accept does not exceed 2% in compliance with ANSI/NCSL Z540.3-2006.

THE CALIBRATION REPORT STATUS:

PASS - Term used when compliance statement is given, and the measurement result is PASS.

PASS² - Term used when compliance statement is given, and the measurement result is conditional passed or PASS².

FAIL - Term used when compliance statement is given, and the measurement result is FAIL.

FAIL² - Term used when compliance statement is given, and the measurement result is conditional failed or FAIL².

REPORT OF VALUE - Term used when reported measurement is not requiring compliance statement in report.

ADJUSTED - When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value as left.

LIMITED - When an instrument fails calibration but is still functional in a limited manner.

The expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%, unless otherwise stated. This calibration report complies with ISO/IEC 17025:2017 and ANSI/NCSL Z540.3. Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. All standards are traceable to SI through the National Institute of Standards and Technology (NIST) and/or recognized national or international standards laboratories. Services rendered include proper manufacturer's service instruction and are warranted for no less than thirty (30) days. The information on this report pertains only to the instrument identified, this may not be reproduced in part or in a whole without the prior written approval of the issuing MP Calibration Laboratory.

Standards Used to Calibrate Equipment

I.D.	Manufacturer	Description	Model	Traceability Number	Cal. Due Date
DM0727	AGILENT	SIGNAL GENERATOR	E4432B	551220084298216	Jan 2, 2025
DW1455	ROHDE & SCHWARZ	AVG POWER SENSOR	NRP-Z91	551220084113369	Jul 3, 2025
ED6101	STEREN	TERMOMETRO DIGITAL	TER-150	5523631030977330	Jun 4, 2025

Procedures Used in this Event

Procedure Name

R AND S NRPXXSN

Description

Three-Path Power Sensors, Rohde and Schwarz NRPxxS(N), Ver.21, Aug-16-2023

Additional notes: See attached data.

Calibrating Technician:

JUAN LANDERO

Approved By:

MAGNOLIA TORRES