Combined MRI and Optical Computed Tomography: Literature Review

Ciara McErlean

23/1/13

1 Introduction

Optical computed tomography reported before, as far back as 1983 (Ray and Semer-jian, only 1 projection, sodium-seeded methane flame) [1] and again in 1990 (Kawata, algae) [2]. Brown 1992 [3] has very similar setup to Gore and Sharpe. Used matching fluid (they call it clearing fluid) and cleared a sample of stained cochlea in 3BB:5MS. 600nm illumination to reduce scatter. 1.92mm radius imaged with resolution 16μm. But Brown is not referenced by either Gore or Sharpe. :(

2 Theory

2.1 Radon Transform

2.2 Filtered Backprojection

Imaging with high resolution using this principle requires that the projection data are a direct result of the line integral of the parallel rays passing through the sample. [4]

The optical images captured by the CCD are approximated as parallel projections throughout the specimen. A series of images at different views are then obtained

1

by rotating the specimen. In such a way, the cross sections of the specimen can be independently reconstructed with the data on the corresponding rows of the CCD captured at sufcient different views. [4]

In fact, two assumptions are made to enable the conventional OPT to use the term of projection and the reconstruction algorithm. First, the narrow cones of the light ray projected onto a CCD are assumed to be the strip integrals through the sample. Second, the axes of the light cones are taken to be parallel to each other. The two assumptions place severe constraints on the resolution of OPT imaging. In this paper, we present improved image-forming optics for OPT, with which the axes of the light cones are parallel to the optical axis of the optics. As a result, the parallel integral throughout a sample can be satisfied, and the second assumption mentioned above is relaxed. This method results in an improved spatial resolution, especially for the cross sections far from the optical axis. [4]

Fourier slice theorem requirements, see Wang 2007 [5]

2.3 Refractive index matching

2.4 Common artefacts

Axis of rotation problems. Corrections suggested by many groups. Recently by Dong in 2012 [6] discuss method.

3 Dosimetry

3.1 Laser scanning configuration

One of the first reported optical computed tomography (OptCT) systems was developed in the area of gel dosimetry. Accurate 3-D measurement of dose delivery in radiotherapy is extremely important in developing safe treatment plans. Specialist polymer gels, such as BANG® [7], respond to irradiation with changes in optical attenuation and scattering properties. This makes them ideal for measuring 3-D dose distributions. Previously the irradiated gels were measured by MRI and x-ray CT however, these are expensive imaging modalities. In 1996, Gore and Maryanski published the first system for scanning polymer gels using optical computed tomography. [8] In later comparisons, OptCT has been found to be more precise, have reduced noise and smoother line profiles than MRI for gel dosimetry. [9]

Gore's system consisted of a He-Ne laser source and large area photodiode detector (see Figure 1). Translate-rotate acquisition was employed whereby the sample was rotated and projection data acquired by the photodiode over 360°. The smaller the angular steps between projections, the more accurate the reconstruction. [10] For a 2-D reconstruction, projections are acquired for multiple spots across a slice of the sample by translating the laser beam using mirrors. For 3-D information, the sample height had to be manually adjusted and many 2-D slices acquired. This meant scanning an entire sample took hours and lengthy scanning times are the chief disadvantage of the laser scanning method. Accuracy of 5% is reported and spatial resolution of 2mm, which is roughly the same as the laser beam width. [8]

The idea of OptCT scanning in dosimetry was quickly developed by other groups. Laser scanning set-ups were published in 1996 by Tarte *et al.*, [11] and Kelly *et al.* [REF] *Can't find the paper 1996 Kelly references in 1998 [12] Med Phys says it doesn't exist.* Kelly *et al.* claim to have independently developed their scanner which is very similar to that of Gore's. In in both Kelly's and Tarte's scanners, the sample is rotated and translated using a stage whereas Gore used mirrors to translate the laser spot across the sample.

A commercial laser scanning OptCT system, OCTOPUS™by MGS Research, Inc. (Madison, CT), is an extension of Gore's original set-up with the addition of a platform capable of vertical movement for automated slice-selection. [13] For a number of years it was the only commercially available system and has been characterised

by several groups. [13–16] According to Oldham, characterisation of OptCT systems should include checks on geometric distortion, accuracy of reconstruction, scatter artefacts and reflection and refraction artefacts. [17]


Figure 1: A first generation, Laser Scanning OptCT system developed by Gore. The sample is rotated and projections recorded at a number of angles. The mirrors scan the laser beam across the sample but movement in the vertical direction is by manual adjustment only (figure from [8]).

Laser scanning systems include a beam splitter before the sample to create a reference beam. Dividing projections by the reference intensity corrects for laser beam intensity fluctuations. [8]

Refraction and reflection at container walls are significant concerns for all configurations of dosimetry with OptCT. Generally, laser beams are incident on the gel container at a small angle, such as 5°, to avoid large reflection at the interface. In addition, the gel container is usually placed in a tank containing 'matching fluid' with a refractive index close to that of the gel. This prevents significant refraction as the light passes into the gel. Doran found through ray tracing simulations that the refractive index of the walls of the matching tank and gel container are not important compared to the gel and matching fluid. The optimum difference in refractive index between these two was calculated to be 0.0025 and not zero as originally thought. [18]

To maximise the dynamic range of the system, food dye is commonly added to the

4

matching fluid so both the refractive index and optical density of the matching fluid and gel are very similar. [19]

3.2 Pixelated detector based systems

In 1997 the first charge coupled device (CCD) camera based OptCT system was published by Tarte *et al.* which employed an incoherent white light source and CCD camera detection. [20] The advantage of a pixelated detector based system is that an entire 2-D projection can be imaged at once, potentially increasing the scanning speed by several orders of magnitude depending on the data through-put rate. Tarte's system used a divergent light source and diffusing screen to measure optical density in a thin gel section (see Figure 2).


Figure 2: Diagram of the first CCD-based OptCT system, developed by Tarte *et al.* It uses divergent illumination from a white light source and CCD camera detection to record an entire 2D projection at once (figure from [20]).

The accuracy of Tarte's system was checked by comparison with the standard measure of dosimetry, the parallel plate ionisation chamber. It was found to be on average within 3% of the value from the ionisation chamber. [20] A comparison between Tarte's laser scanning and CCD set-ups found that they had similar spatial resolutions. The CCD method had improved speed of acquistion but suffered from consistently worse SNR as a photodiode detector can collect many more photons per 'pixel' than a CCD camera. [20]

Advances in technology have meant that high quality detectors are much more affordable. A cheaper alternative to very high quality CCD cameras is the CMOS

(Complementary Metal-Oxide-Semiconductor) detector which has the potential for higher resolution and dynamic range. [21] Using a higher quality detector would improve many OptCT systems in terms of scanning speed and reduced artefacts. [18,20]

Parallel beam configuration: One method to reconstruct 3-D images with a CCD or CMOS detector is to create a broad parallel beam. This allows the use of parallel reconstruction algorithms, very similar to those used for x-ray CT. Each 2-D projection image recorded corresponds to one row for every slice in the 3-D reconstruction sinogram. [21] Telecentric optics, in which the chief rays are parallel to the optical axis, are key in the design of this configuration. [22] Telecentric optics can be achieved either through a careful arrangement of a large converging lens before the sample and standard camera lens [18] (see Figure 3) or through an expensive telecentric lens [23]. The process of forming a parallel beam results in non-uniformities in the lightfield. This is compensated for by subtracting a 'correction' or 'open lightfield', image which is a projection taken with no sample in the tank. [18]


Figure 3: Diagram of a parallel beam OptCT system, developed by Doran *et al.* Telecentric optics create a parallel beam (figure from [18]).

Initial systems suffered from 'graininess' due to the unstable gain of cheap CCD cameras and granularity of the diffusing screen. [18] Doran *et al.* proposed some methods of correcting these problems. Oscillating the diffuser screen at high frequency "'smears' out the granularity" while randomly horizontally displacing the

CCD camera by a few pixels between acquisitions can reduce the effect of 'bad' pixels. [18] The parallel configuration appears to be more susceptible to schlieren artefacts caused by refractive index inhomogeneities in the sample. [24]

Cone beam configuration: Wolodzko *et al.* published the first cone beam OptCT system with CCD detection for gel dosimetry. [25] One advantage of this configuration is the optics for producing a cone beam are much simpler than those for producing accurate parallel beams. [21] However, the reconstruction is computationally more complex. [26] A commercial cone-beam system, Vista^{TM}by Modus Medical Devices Inc. (London, ON, Canada), is available and reviewed recently by Olding *et al.* [27]


Figure 4: Cone-beam CCD configuration (figure from [25]).

When pixelated detectors are used, there appears to be more literature based on the parallel beam configuration than cone-beam. Although there has not been experimental comparison of the two Doran suggests that while cone-beam is usually somewhat cheaper due to simplified optics, modern parallel-beam systems have better scatter-rejection and may have fewer stray light problems. [21, 27, 28]

7

4 Tissue imaging

4.1 Optical Projection Tomography

Another version of OptCT was developed by Sharpe *et al.* in the area of 3-D microscopy for gene expression studies. [29] Although this set-up in 2002 came after Gore's they are apparently independent and Sharpe named his technique Optical Projection Tomography (OPT).

Other techniques for 3-D microscopy include confocal microscopy which can image to a depth of about 1mm. [30] However, this is limited to fluorescent signals meaning many optical stains used routinely in histology would not work. Optical coherence tomography is another technique capable of micrometer-scale resolution. [31] However, its depth is limited to 2-3mm in tisse. Sharpe's OPT set-up could image much larger specimen, up to 15mm thick. [29] OPT is a projection based tomography technique, as the name implies, which means mathematical reconstruction is required. The other techniques mentioned give tomographic images by sectioning. [32] **Discuss**

Sharpe's system includes a microscope to focus projections of a mouse embryo onto a camera imaging chip (CIC). Image-focusing optics are one difference between OPT and x-ray CT, which records shadows of the sample. Sharpe reports some impressive images (see Figure 5). Use of the microscope gives resolution of about 5-$10\mu m$ meaning single-cell membranes, around $10\mu m$ thick, can be seen. [29] The axis of rotation is chosen so only half of the specimen is in focus at once which compensates for poor depth of focus.

To reduce scattering and refraction within the specimen it was immersed in Murray's Clear, also known as BABBs (1:2 mixture of benzyl alcohol and benzyl benzoate). BABBs works as an optical clearing agent (OCA) which reduces refraction by replacing fluid in a specimen. The OCA is chosen with a refractive index close to that of the solid structures in the specimen. See Section 5 for more detail on the clearing mechanism and possible agents. The result of clearing means that the light paths through the specimen can be approximated as parallel line integrals making high resolution reconstruction through back projection possible.

(a) OPT setup

(b) Mouse images

Figure 5: Part (a) shows the optical setup for the first OPT system by Sharpe. CIC indicates the camera imaging chip. A microscope is used to focus. The specimen is set in agarose gel for stability. Part (b) shows some sample images from OPT scanning of a mouse embryo. The iso-surface shows contours linking all regions above a certain intensity. Both figures are adapted from [29].

In 2005 Fauver reported a modified version of OPT capable of imaging single cell nuclei with 0.9μm resolution (see Figure 6). [33] The OPT microscope includes a rotation stage and piezoelectrically driven objective lens. In a technique similar to Hausler [34] the objective lens is scanned axially to create an extended depth of field (DOF) image which is also known as a pseudoprojection. The extended DOF means features have the same focus from all angles, allowing high resolution

reconstruction. However, this is not a truly quantitative technique, hence pseudo and not true projections are recorded. A high numerical aperture (NA) lens gives high resolution at the expense of low depth of field. If such high resolution is not required, low NA optics such as those used by Sharpe would be a more accurate way to generate projections than scanning a high NA lens.**CHECK**


Figure 6: OPT microscope for imaging single cell nuclei. A microcapillary tube injected with cells is rotated to sub-micron precision with refractive index matching to 0.02. The piezoelectric objective lens is scanned axially to create extended depth of field images (figure from [33]).

Wang and Wang report an improvement to OPT giving higher axial and lateral resolution even for slices far from the optical axis. [4,5] As previously mentioned, to obtain high quality reconstructions the projections should closely approximate a line integral of parallel rays passing through the sample. [4] This is not exactly the case for OPT which limits the resolution possible. Wang proposed placing an iris at the back focus of the objective lens. This reduces divergence of the projection rays from paths parallel to the optical axis. Resolution 40μm in 2006. For 2007 system with NA of 0.014 resolution of 25μm is achieved. (Also covers technique for placing iris in eOPT, mention and discuss in Section 4.2.)**Finish**

Live specimen OPT not a topic taken up by many other groups due to difficulties

10

in keeping specimen alive, problems with OCA, can only image very small specimen anyway. [35] Boot and Sharpe 2008. Colas and Sharpe do live, time-lapse OPT in 2009, not particularly useful. Poor resolution due to lack of clearing [36]

Computational methods for improving reconstructions in OPT by Birk in 2011 [37]

4.2   Fluorescent OPT/emission OptCT

There are many optical stains which are fluorescent which meant the development of fluorescent or emission OptCT/OPT very useful for the study of biological specimen. REF

Walls 2005 - first mention of eOPT? Correction of artefacts. [22]

Oldham has dual set-up. However, it was not originally quantitative. Just pretty pictures. Steps by various groups to get quantitative information out of emission. What is the change to the set-up? [38] quantitative: [39]

Imperial group investigating Opt combined with FRET and FLIM. Define these and some uses for the combined modality. What physical/software changes needed for this imaging. FLIM OPT: [40], In vivo FLIM OPT [41].

Resolution improvement to eOPT in 2007 by Walls [42] is this quantitative?

Lorbeer - SLOT [43]

5 Optical Clearing

6 Optical Staining

Sharpe 2002 stains for gene expression and limb bud growth.

Discuss which stains are relevant for cancer biomarkers. [44]

Being able to use optical stains is extremely useful for computer recognition of organs, can pick better thresholds. [32]

Soufan 2003 [45] gene expression during cardiac development. Highlights problem of whole-mount staining not working for all stains. Depends on size of specimen. Therefore, they took slices of embryo heart.

7 Recent Research

Time gated OPT Bassi 2010 [46]

CLAHE Hornblad 2011 [47]

References

[1] S R Ray and H G Semerjian. Laser tomography for simultaneous concentration and temperature measurement in reacting flows. In *American Institute of Aeronautics and Astronautics, Thermophysics Conference, 18 th, Montreal, Canada*, 1983.

[2] S Kawata, O Nakamura, T Noda, H Ooki, K Ogino, Y Kuroiwa, and S Minami. Laser computed-tomography microscope. *Applied optics*, 29(26):3805–3809, 1990.

[3] C S Brown, D H Burns, F A Spelman, and A C Nelson. Computed tomography from optical projections for three-dimensional reconstruction of thick objects. *Applied optics*, 31(29):6247–6254, 1992.

[4] Y Wang and R Wang. Imaging using parallel integrals in optical projection tomography. *Physics in Medicine and Biology*, 51(23):6023–6032, October 2006.

[5] Y Wang and R K Wang. Optimization of image-forming optics for transmission optical projection tomography. *Applied optics*, 46(27):6815–6820, 2007.

[6] D Dong, S Zhu, C Qin, V Kumar, J V Stein, S Oehler, C Savakis, J Tian, and J Ripoll. Automated Recovery of the Center of Rotation in Optical Projection Tomography in the Presence of Scattering. 2012.

[7] M J Maryanski, Y Z Zastavker, and J C Gore. Radiation dose distributions in three dimensions from tomographic optical density scanning of polymer gels: Ii. optical properties of the bang polymer gel. *Physics in medicine and biology*, 41(12):2705, 1999.

[8] J C Gore, M Ranade, and M J Maryanski. Radiation dose distributions in three dimensions from tomographic optical density scanning of polymer gels: I. Development of an optical scanner - Abstract - Physics in Medicine and Biology - IOPscience. *Physics in medicine . . .* , 1999.

[9] M Oldham, J H Siewerdsen, A Shetty, and D A Jaffray. High resolution gel-dosimetry by optical-CT and MR scanning. *Medical Physics*, 28(7):1436, 2001.

[10] J C Russ. *The Image Processing Handbook*. Image Processing Handbook. Crc Press, 2002.

[11] B J Tarte, P A Jardine, and T van Doorn. Laser-scanned agarose gel sections for radiation field mapping. *International Journal of Radiation Oncology* Biology* Physics*, 36(1):175–179, 1996.

[12] R G Kelly, K J Jordan, and J J Battista. Optical CT reconstruction of 3D dose distributions using the ferrous–benzoic–xylenol (FBX) gel dosimeter. *Medical Physics*, 25(9):1741–50, September 1998.

[13] K T S Islam, J F Dempsey, Manisha K Ranade, M J Maryanski, and D A Low. Initial evaluation of commercial optical CT-based 3D gel dosimeter. *Medical Physics*, 30(8):2159, 2003.

[14] Y Xu, C Wuu, and Marek J Maryanski. Determining optimal gel sensitivity in optical CT scanning of gel dosimeters. *Medical Physics*, 30(8):2257, 2003.

[15] Y Xu, C S Wuu, and M J Maryanski. Performance of a commercial optical CT scanner and polymer gel dosimeters for 3-D dose verification. *Medical Physics*, 31(11):3024, 2004.

[16] H S Sakhalkar, J Adamovics, G Ibbott, and M Oldham. A comprehensive evaluation of the PRESAGE/optical-CT 3D dosimetry system. *Medical Physics*, 36(1):71, 2009.

[17] M Oldham. Optical-CT scanning of polymer gels. *Journal of Physics: Conference Series*, 3:122–135, November 2004.

[18] S J Doran, K K Koerkamp, M A Bero, P Jenneson, E J Morton, and W B Gilboy. A CCD-based optical CT scanner for high-resolution 3D imaging of radiation dose distributions: equipment specifications, optical simulations and preliminary results. *Physics in Medicine and Biology*, 46(12):3191–3213, November 2001.

[19] N Krstajić and S J Doran. Focusing optics of a parallel beam CCD optical tomography apparatus for 3D radiation gel dosimetry. *Physics in Medicine and Biology*, 51(8):2055–2075, April 2006.

[20] B J Tarte, P A Jardine, T van Doorn, K N Nitschke, and M G Poulsen. Development of a CCD array imaging system for measurement of dose distributions in doped agarose gels. *Medical Physics*, 24:1521, 1997.

[21] S J Doran. The history and principles of optical computed tomography for scanning 3-D radiation dosimeters: 2008 update. *Journal of Physics: Conference Series*, 164:012020, June 2009.

[22] J R Walls, J G Sled, J Sharpe, and R M Henkelman. Correction of artefacts in optical projection tomography. *Physics in Medicine and Biology*, 50(19):4645–4665, September 2005.

[23] H S Sakhalkar and M Oldham. Fast, high-resolution 3D dosimetry utilizing a novel optical-CT scanner incorporating tertiary telecentric collimation. *Medical Physics*, 35(1):101, 2008.

[24] N Krstajić and S J Doran. Fast laser scanning optical-CT apparatus for 3D radiation dosimetry. *Physics in Medicine and Biology*, 52(11):N257–N263, May 2007.

[25] J G Wolodzko, C Marsden, and A Appleby. CCD imaging for optical tomography of gel radiation dosimeters. *Medical Physics*, 26:2508, 1999.

[26] J Hsieh. *Computed Tomography: Principles, Design, Artifacts, and Recent Advances.* SPIE Press monograph. SPIE Press, 2003.

[27] T Olding and L J Schreiner. Cone-beam optical computed tomography for gel dosimetry II: imaging protocols. *Physics in Medicine and Biology*, 56(5):1259–1279, February 2011.

[28] A Thomas, J Newton, and M Oldham. A method to correct for stray light in telecentric optical-CT imaging of radiochromic dosimeters. *Physics in Medicine and Biology*, 56(14):4433–4451, June 2011.

[29] J Sharpe. Optical Projection Tomography as a Tool for 3D Microscopy and Gene Expression Studies. *Science*, 296(5567):541–545, April 2002.

[30] R H Webb. Confocal optical microscopy. *Reports on Progress in Physics*, 59:427–471, 1996.

[31] D Huang, E A Swanson, C P Lin, J S Schuman, W G Stinson, W Chang, M R Hee, T Flotte, K Gregory, C A Puliafito, et al. *Optical coherence tomography.* PhD thesis, Massachusetts Institute of Technology, Whitaker College of Health Sciences and Technology, 1993.

[32] J Sharpe. Optical projection tomography as a new tool for studying embryo anatomy. *Journal of Anatomy*, 202(2):175–181, February 2003.

[33] M Fauver, E J Seibel, J R Rahn, M G Meyer, F W Patten, T Neumann, and A C Nelson. Three-dimensional imaging of single isolated cell nuclei using optical projection tomography. *Opt. Express*, 13(11):4210–4223, 2005.

[34] G Häusler. A method to increase the depth of focus by two step image processing. *Optics Communications*, 6(1):38–42, 1972.

[35] M J Boot, C H Westerberg, J Sanz-Ezquerro, J Cotterell, R Schweitzer, M Torres, and J Sharpe. In vitro whole-organ imaging: 4D quantification of growing mouse limb buds. *Nature Methods*, 5(7):609–612, May 2008.

[36] J F Colas and J Sharpe. Live optical projection tomography. *Organogenesis*, 5(4):211–216, 2009.

[37] U Jochen Birk, A Darrell, N Konstantinides, A Sarasa-Renedo, and J Ripoll. Improved reconstructions and generalized filtered back projection for optical projection tomography. *Applied optics*, 50(4):392–398, 2011.

[38] M Oldham, H Sakhalkar, T Oliver, G Allan Johnson, and M Dewhirst. Optical clearing of unsectioned specimens for three-dimensional imaging via optical transmission and emission tomography. *Journal of Biomedical Optics*, 13(2):021113, 2006.

[39] E Kim, J Bowsher, A S Thomas, H Sakhalkar, M Dewhirst, and M Oldham. Improving the quantitative accuracy of optical-emission computed tomography by incorporating an attenuation correction: application to HIF1 imaging. *Physics in Medicine and Biology*, 53(19):5371–5383, September 2008.

[40] J McGinty, K B Tahir, R Laine, C B Talbot, C Dunsby, M A A Neil, L Quintana, J Swoger, J Sharpe, and P M W French. Fluorescence lifetime optical projection tomography. *Journal of Biophotonics*, 1(5):390–394, October 2008.

[41] J McGinty, H B Taylor, L Chen, L Bugeon, J R Lamb, M J Dallman, and P M W French. In vivo fluorescence lifetime optical projection tomography. *Biomedical optics express*, 2(5):1340–1350, 2011.

[42] J R Walls, J G Sled, J Sharpe, and R M Henkelman. Resolution improvement in emission optical projection tomography. *Physics in Medicine and Biology*, 52(10):2775–2790, April 2007.

[43] R A Lorbeer, M Heidrich, C Lorbeer, D F Ramírez Ojeda, G Bicker, H Meyer, and A Heisterkamp. Highly efficient 3D fluorescence microscopy with a scanning laser optical tomograph. *Optics Express*, 19(6):5419–5430, 2011.

[44] D Hanahan and R A Weinberg. Hallmarks of Cancer: The Next Generation. *Cell*, 144(5):646–674, March 2011.

[45] A T Soufan, J M Ruijter, M J B van den Hoff, P A J de Boer, J Hagoort, and A F M Moorman. Three-dimensional reconstruction of gene expression patterns during cardiac development. *Physiological genomics*, 13(3):187–195, 2003.

[46] A Bassi, D Brida, C D'Andrea, G Valentini, R Cubeddu, S De Silvestri, and G Cerullo. Time-gated optical projection tomography. *Optics letters*, 35(16):2732–2734, 2010.

[47] A Hörnblad, A Cheddad, and U Ahlgren. An improved protocol for optical projection tomography imaging reveals lobular heterogeneities in pancreatic islet and -cell mass distribution. *Islets*, 3(4):204–208, July 2011.