

Customer	York
Cust Ref #	
Instrument	POA200
Inst. SN.	AP10007513

Date:	3/5/2026
-------	----------

Test Flow	2SCFH
Ambient Temperature:	23.5°C
Relative Humidity	51.6%RH

Comments & Observations:

Test Point Number	NIST Traceable Standard gas ppm	Analyzer pv ppm
1	9.96(N ₂ balance)	9.93

We hereby certify that the analyzer detailed above has been tested and calibrated by the Undersigned, and its performance was found to meet our specifications unless noted otherwise in the comments section.

For and on behalf of PhyMetrix Ltd.



For Quality Assurance Manager